

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF110618\
 Data File : BF110531.D
 Acq On : 6 Nov 2018 22:34
 Operator : JU/SJ
 Sample : J5764-07 2X
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 JC-03-110218-D

Quant Time: Nov 07 12:47:47 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF102318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 01 16:06:12 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.96	152	92831	20.00	ng	0.00
21) Naphthalene-d8	8.24	136	351831	20.00	ng	0.00
39) Acenaphthene-d10	10.00	164	134565	20.00	ng	0.00
64) Phenanthrene-d10	11.49	188	214583	20.00	ng	0.00
76) Chrysene-d12	14.14	240	160489	20.00	ng	0.00
87) Perylene-d12	15.67	264	114885	20.00	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.57	112	222634	39.05	ng	0.00
7) Phenol-d6	6.57	99	281543	38.61	ng	0.00
23) Nitrobenzene-d5	7.52	82	155170	26.16	ng	0.00
42) 2,4,6-Tribromophenol	10.79	330	43721	32.80	ng	0.00
45) 2-Fluorobiphenyl	9.31	172	295117	34.44	ng	0.00
79) Terphenyl-d14	13.07	244	229209	29.70	ng	0.00
Target Compounds						
50) Dimethylphthalate	9.70	163	40620	4.089	ng	98
75) Fluoranthene	12.71	202	28543	2.505	ng	97
78) Pyrene	12.94	202	29679	2.580	ng	97
84) Bis(2-ethylhexyl)phthalate	14.09	149	16540	2.450	ng	# 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF110618\
Data File : BF110531.D
Acq On : 6 Nov 2018 22:34
Operator : JU/SJ
Sample : J5764-07 2X
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
JC-03-110218-D

Quant Time: Nov 07 12:47:47 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF102318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 01 16:06:12 2018
Response via : Initial Calibration

