

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF110723\
 Data File : BF136177.D
 Acq On : 07 Nov 2023 15:47
 Operator : CG\JU
 Sample : PB156921BL
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB156921BL

Quant Time: Nov 08 02:57:05 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF110723.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 08 02:12:01 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.816	152	92028	20.000	ng	0.00
21) Naphthalene-d8	8.098	136	367949	20.000	ng	0.00
39) Acenaphthene-d10	9.857	164	188969	20.000	ng	0.00
64) Phenanthrene-d10	11.351	188	341560	20.000	ng	0.00
76) Chrysene-d12	14.004	240	196372	20.000	ng	0.00
86) Perylene-d12	15.492	264	181071	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.457	112	725672	125.554	ng	0.01
7) Phenol-d6	6.451	99	891315	125.322	ng	0.00
23) Nitrobenzene-d5	7.381	82	571281	86.641	ng	0.00
42) 2,4,6-Tribromophenol	10.651	330	259719	131.228	ng	0.00
45) 2-Fluorobiphenyl	9.181	172	1087912	85.101	ng	0.00
79) Terphenyl-d14	12.951	244	1174267	84.882	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF110723\
Data File : BF136177.D
Acq On : 07 Nov 2023 15:47
Operator : CG\JU
Sample : PB156921BL
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB156921BL

Quant Time: Nov 08 02:57:05 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF110723.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 08 02:12:01 2023
Response via : Initial Calibration

