

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF110818\
 Data File : BF110602.D
 Acq On : 8 Nov 2018 23:08
 Operator : JU/SJ
 Sample : J5838-11
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 HC-SOIL-2

Manual Integrations
 APPROVED

Sohil
 11/9/2018 4:26:11 PM

Quant Time: Nov 09 12:43:44 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF110818.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 08 14:14:47 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.95	152	87370	20.00	ng	0.00
21) Naphthalene-d8	8.22	136	350824	20.00	ng	0.00
39) Acenaphthene-d10	9.98	164	167576	20.00	ng	-0.01
64) Phenanthrene-d10	11.47	188	322506	20.00	ng	-0.01
76) Chrysene-d12	14.13	240	207987	20.00	ng	-0.01
87) Perylene-d12	15.65	264	173091	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.57	112	777079	144.47	ng	0.00
7) Phenol-d6	6.57	99	997057	144.96	ng	0.00
23) Nitrobenzene-d5	7.51	82	640019	105.06	ng	0.00
42) 2,4,6-Tribromophenol	10.78	330	239733	141.98	ng	0.00
45) 2-Fluorobiphenyl	9.30	172	1123861	95.78	ng	0.00
79) Terphenyl-d14	13.06	244	1074165	96.72	ng	0.00
Target Compounds						
10) Phenol	6.59	94	47954	6.013	ng	Qvalue # 63
50) Dimethylphthalate	9.69	163	89259	7.138	ng	100
75) Fluoranthene	12.70	202	53621	3.095	ng	98
78) Pyrene	12.93	202	42658	2.664	ng	98
88) Benzo(b)fluoranthene	15.20	252	25545m	2.467	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF110818\
Data File : BF110602.D
Acq On : 8 Nov 2018 23:08
Operator : JU/SJ
Sample : J5838-11
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
HC-SOIL-2

Manual Integrations
APPROVED

Sohil
11/9/2018 4:26:11 PM

Quant Time: Nov 09 12:43:44 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF110818.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 08 14:14:47 2018
Response via : Initial Calibration

