

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF110818\
 Data File : BF110605.D
 Acq On : 9 Nov 2018 00:29
 Operator : JU/SJ
 Sample : J5857-02 2X
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PS-CON-2

Quant Time: Nov 09 13:27:45 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF110818.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 08 14:14:47 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.95	152	88718	20.00	ng	0.00
21) Naphthalene-d8	8.23	136	364557	20.00	ng	0.00
39) Acenaphthene-d10	9.99	164	157532	20.00	ng	0.00
64) Phenanthrene-d10	11.48	188	250671	20.00	ng	0.00
76) Chrysene-d12	14.13	240	208046	20.00	ng	-0.01
87) Perylene-d12	15.65	264	168887	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.56	112	311908	57.11	ng	0.00
7) Phenol-d6	6.56	99	440422	63.06	ng	-0.01
23) Nitrobenzene-d5	7.50	82	282857	44.68	ng	-0.01
42) 2,4,6-Tribromophenol	10.77	330	71127	44.81	ng	-0.01
45) 2-Fluorobiphenyl	9.30	172	540383	48.99	ng	0.00
79) Terphenyl-d14	13.06	244	415870	37.44	ng	-0.01

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF110818\
Data File : BF110605.D
Acq On : 9 Nov 2018 00:29
Operator : JU/SJ
Sample : J5857-02 2X
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PS-CON-2

Quant Time: Nov 09 13:27:45 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF110818.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 08 14:14:47 2018
Response via : Initial Calibration

