

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF110819\
 Data File : BF117729.D
 Acq On : 8 Nov 2019 13:51
 Operator : JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 09 01:21:05 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF110619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 06 17:27:10 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	96	0.00
2	1,4-Dioxane	0.551	0.543	1.5	98	-0.01
3	Pyridine	1.514	1.493	1.4	99	-0.01
4	n-Nitrosodimethylamine	0.667	0.564	15.4	84	-0.02
5 S	2-Fluorophenol	1.105	1.117	-1.1	100	0.00
6	Aniline	1.912	1.881	1.6	97	0.00
7 S	Phenol-d6	1.362	1.359	0.2	99	0.00
8	2-Chlorophenol	1.229	1.241	-1.0	99	0.00
9	Benzaldehyde	0.948	0.927	2.2	94	0.00
10 C	Phenol	1.468	1.460	0.5	99	0.00
11	bis(2-Chloroethyl)ether	1.212	1.183	2.4	96	0.00
12	1,3-Dichlorobenzene	1.425	1.431	-0.4	99	0.00
13 C	1,4-Dichlorobenzene	1.426	1.439	-0.9	99	0.00
14	1,2-Dichlorobenzene	1.313	1.342	-2.2	100	0.00
15	Benzyl Alcohol	1.086	1.082	0.4	97	0.00
16	2,2'-oxybis(1-Chloropropane	1.917	1.850	3.5	95	0.00
17	2-Methylphenol	1.056	1.045	1.0	99	0.00
18	Hexachloroethane	0.527	0.519	1.5	97	0.00
19 P	n-Nitroso-di-n-propylamine	0.896	0.854	4.7	95	0.00
20	3+4-Methylphenols	1.257	1.278	-1.7	98	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	97	0.00
22	Acetophenone	0.457	0.449	1.8	99	0.00
23 S	Nitrobenzene-d5	0.339	0.307	9.4	92	0.00
24	Nitrobenzene	0.354	0.328	7.3	93	0.00
25	Isophorone	0.648	0.612	5.6	98	0.00
26 C	2-Nitrophenol	0.164	0.146	11.0	89	0.00
27	2,4-Dimethylphenol	0.277	0.265	4.3	98	0.00
28	bis(2-Chloroethoxy)methane	0.416	0.399	4.1	98	0.00
29 C	2,4-Dichlorophenol	0.285	0.282	1.1	99	0.00
30	1,2,4-Trichlorobenzene	0.327	0.320	2.1	98	0.00
31	Naphthalene	0.914	0.915	-0.1	100	0.00
32	Benzoic acid	0.219	0.203	7.3	99	0.00
33	4-Chloroaniline	0.425	0.419	1.4	100	0.00
34 C	Hexachlorobutadiene	0.187	0.183	2.1	98	0.00
35	Caprolactam	0.095	0.093	2.1	103	0.00
36 C	4-Chloro-3-methylphenol	0.288	0.284	1.4	100	0.00
37	2-Methylnaphthalene	0.622	0.624	-0.3	100	0.00
38	1-Methylnaphthalene	0.591	0.594	-0.5	100	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	100	0.00
40	1,2,4,5-Tetrachlorobenzene	0.547	0.537	1.8	100	0.00
41 P	Hexachlorocyclopentadiene	0.295	0.296	-0.3	102	0.00
42 S	2,4,6-Tribromophenol	0.196	0.197	-0.5	104	0.00
43 C	2,4,6-Trichlorophenol	0.398	0.383	3.8	100	0.00
44	2,4,5-Trichlorophenol	0.404	0.400	1.0	101	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF110819\
 Data File : BF117729.D
 Acq On : 8 Nov 2019 13:51
 Operator : JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 09 01:21:05 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF110619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 06 17:27:10 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.014	1.067	-5.2	100	0.00
46	1,1'-Biphenyl	1.401	1.400	0.1	102	0.00
47	2-Chloronaphthalene	1.162	1.140	1.9	100	0.00
48	2-Nitroaniline	0.364	0.332	8.8	94	0.00
49	Acenaphthylene	1.693	1.684	0.5	101	0.00
50	Dimethylphthalate	1.352	1.320	2.4	101	0.00
51	2,6-Dinitrotoluene	0.303	0.275	9.2	94	0.00
52 C	Acenaphthene	1.092	1.055	3.4	99	0.00
53	3-Nitroaniline	0.365	0.343	6.0	98	0.00
54 P	2,4-Dinitrophenol	0.104	0.098	5.8	101	0.00
55	Dibenzofuran	1.529	1.511	1.2	100	0.00
56 P	4-Nitrophenol	0.275	0.263	4.4	99	0.00
57	2,4-Dinitrotoluene	0.388	0.344	11.3	90	0.00
58	Fluorene	1.127	1.119	0.7	101	0.00
59	2,3,4,6-Tetrachlorophenol	0.336	0.340	-1.2	104	0.00
60	Diethylphthalate	1.330	1.324	0.5	102	0.00
61	4-Chlorophenyl-phenylether	0.586	0.584	0.3	101	0.00
62	4-Nitroaniline	0.360	0.347	3.6	100	0.00
63	Azobenzene	1.259	1.239	1.6	101	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	101	0.00
65	4,6-Dinitro-2-methylphenol	0.080	0.075	6.3	101	0.00
66 c	n-Nitrosodiphenylamine	0.591	0.570	3.6	101	0.00
67	4-Bromophenyl-phenylether	0.218	0.214	1.8	103	0.00
68	Hexachlorobenzene	0.252	0.246	2.4	103	0.00
69	Atrazine	0.195	0.194	0.5	105	0.00
70 C	Pentachlorophenol	0.141	0.144	-2.1	107	0.00
71	Phenanthrene	0.959	0.930	3.0	102	0.00
72	Anthracene	0.954	0.930	2.5	102	0.00
73	Carbazole	0.872	0.865	0.8	103	0.00
74	Di-n-butylphthalate	1.060	1.057	0.3	104	0.00
75 C	Fluoranthene	0.979	0.977	0.2	104	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	106	0.00
77	Benzidine	0.695	0.656	5.6	98	0.00
78	Pyrene	1.460	1.371	6.1	104	0.00
79 S	Terphenyl-d14	0.972	0.902	7.2	103	0.00
80	Butylbenzylphthalate	0.674	0.663	1.6	109	0.00
81	Benzo(a)anthracene	1.245	1.173	5.8	103	0.00
82	3,3'-Dichlorobenzidine	0.465	0.470	-1.1	110	0.00
83	Chrysene	1.229	1.162	5.5	104	0.00
84	Bis(2-ethylhexyl)phthalate	0.870	0.846	2.8	107	0.00
85 c	Di-n-octyl phthalate	1.369	1.331	2.8	106	0.00
86	Indeno(1,2,3-cd)pyrene	1.144	0.917	19.8	95	0.00
87 I	Perylene-d12	1.000	1.000	0.0	103	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF110819\
Data File : BF117729.D
Acq On : 8 Nov 2019 13:51
Operator : JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Nov 09 01:21:05 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF110619.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 06 17:27:10 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.260	1.216	3.5	105	0.00
89	Benzo(k)fluoranthene	1.123	1.101	2.0	99	0.00
90 C	Benzo(a)pyrene	1.092	1.046	4.2	103	0.00
91	Dibenzo(a,h)anthracene	0.979	0.837	14.5	94	0.00
92	Benzo(q,h,i)perylene	0.949	0.798	15.9	94	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0