

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF110821\
 Data File : BF126057.D
 Acq On : 8 Nov 2021 10:31
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 08 11:37:48 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF110321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 03 17:21:25 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	112	0.00
2	1,4-Dioxane	0.484	0.444	8.3	106	-0.02
3	Pyridine	1.303	1.188	8.8	104	-0.01
4	n-Nitrosodimethylamine	0.914	0.816	10.7	102	-0.02
5 S	2-Fluorophenol	1.165	1.116	4.2	109	0.00
6	Aniline	1.810	1.728	4.5	110	0.00
7 S	Phenol-d6	1.642	1.554	5.4	108	0.00
8	2-Chlorophenol	1.244	1.182	5.0	109	0.00
9	Benzaldehyde	1.081	0.907	16.1	105	0.00
10 C	Phenol	1.679	1.603	4.5	109	0.00
11	bis(2-Chloroethyl)ether	1.223	1.169	4.4	111	0.00
12	1,3-Dichlorobenzene	1.417	1.367	3.5	112	0.00
13 C	1,4-Dichlorobenzene	1.445	1.389	3.9	112	0.00
14	1,2-Dichlorobenzene	1.392	1.343	3.5	113	0.00
15	Benzyl Alcohol	1.402	1.356	3.3	108	0.00
16	2,2'-oxybis(1-Chloropropane	2.248	2.041	9.2	104	0.00
17	2-Methylphenol	1.059	1.015	4.2	110	0.00
18	Hexachloroethane	0.543	0.539	0.7	113	0.00
19 P	n-Nitroso-di-n-propylamine	1.082	1.047	3.2	110	0.00
20	3+4-Methylphenols	1.432	1.360	5.0	110	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	112	0.00
22	Acetophenone	0.578	0.547	5.4	109	0.00
23 S	Nitrobenzene-d5	0.424	0.449	-5.9	112	0.00
24	Nitrobenzene	0.420	0.426	-1.4	110	0.00
25	Isophorone	0.764	0.732	4.2	110	0.00
26 C	2-Nitrophenol	0.125	0.154	-23.2#	125	0.00
27	2,4-Dimethylphenol	0.278	0.266	4.3	108	0.00
28	bis(2-Chloroethoxy)methane	0.450	0.437	2.9	111	0.00
29 C	2,4-Dichlorophenol	0.314	0.319	-1.6	113	0.00
30	1,2,4-Trichlorobenzene	0.383	0.379	1.0	114	0.00
31	Naphthalene	1.036	1.009	2.6	111	0.00
32	Benzoic acid	0.144	0.154	-6.9	112	0.00
33	4-Chloroaniline	0.414	0.405	2.2	111	0.00
34 C	Hexachlorobutadiene	0.264	0.263	0.4	113	0.00
35	Caprolactam	0.088	0.086	2.3	113	0.00
36 C	4-Chloro-3-methylphenol	0.362	0.362	0.0	111	0.00
37	2-Methylnaphthalene	0.715	0.704	1.5	112	0.00
38	1-Methylnaphthalene	0.693	0.680	1.9	112	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	113	0.00
40	1,2,4,5-Tetrachlorobenzene	0.670	0.652	2.7	112	0.00
41 P	Hexachlorocyclopentadiene	0.348	0.333	4.3	100	0.00
42 S	2,4,6-Tribromophenol	0.229	0.248	-8.3	117	0.00
43 C	2,4,6-Trichlorophenol	0.406	0.417	-2.7	113	0.00
44	2,4,5-Trichlorophenol	0.438	0.441	-0.7	112	0.00
45 S	2-Fluorobiphenyl	1.480	1.432	3.2	112	0.00
46	1,1'-Biphenyl	1.546	1.505	2.7	112	0.00
47	2-Chloronaphthalene	1.217	1.196	1.7	114	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF110821\
 Data File : BF126057.D
 Acq On : 8 Nov 2021 10:31
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 08 11:37:48 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF110321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 03 17:21:25 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.355	0.396	-11.5	114	0.00
49	Acenaphthylene	1.840	1.778	3.4	110	0.00
50	Dimethylphthalate	1.456	1.436	1.4	114	0.00
51	2,6-Dinitrotoluene	0.250	0.280	-12.0	117	0.00
52 C	Acenaphthene	1.161	1.144	1.5	112	0.00
53	3-Nitroaniline	0.256	0.280	-9.4	115	0.00
54 P	2,4-Dinitrophenol	0.093	0.109	-17.2	131	0.00
55	Dibenzofuran	1.747	1.691	3.2	111	0.00
56 P	4-Nitrophenol	0.202	0.206	-2.0	109	0.00
57	2,4-Dinitrotoluene	0.317	0.375	-18.3	120	0.00
58	Fluorene	1.403	1.356	3.3	110	0.00
59	2,3,4,6-Tetrachlorophenol	0.371	0.392	-5.7	112	0.00
60	Diethylphthalate	1.460	1.414	3.2	110	0.00
61	4-Chlorophenyl-phenylether	0.750	0.738	1.6	112	0.00
62	4-Nitroaniline	0.267	0.301	-12.7	119	0.00
63	Azobenzene	1.586	1.483	6.5	106	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	112	0.00
65	4,6-Dinitro-2-methylphenol	0.083	0.100	-20.5	129	0.00
66 c	n-Nitrosodiphenylamine	0.637	0.623	2.2	110	0.00
67	4-Bromophenyl-phenylether	0.242	0.247	-2.1	111	0.00
68	Hexachlorobenzene	0.253	0.257	-1.6	114	0.00
69	Atrazine	0.225	0.230	-2.2	111	0.00
70 C	Pentachlorophenol	0.136	0.140	-2.9	108	0.00
71	Phenanthrene	1.085	1.055	2.8	109	0.00
72	Anthracene	1.083	1.062	1.9	110	0.00
73	Carbazole	1.004	0.970	3.4	109	0.00
74	Di-n-butylphthalate	1.170	1.173	-0.3	109	0.00
75 C	Fluoranthene	1.208	1.182	2.2	108	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	101	0.00
77	Benzydine	0.669	0.605	9.6	100	0.00
78	Pyrene	1.604	1.684	-5.0	108	0.00
79 S	Terphenyl-d14	1.240	1.300	-4.8	106	0.00
80	Butylbenzylphthalate	0.584	0.652	-11.6	106	0.00
81	Benzo(a)anthracene	1.374	1.354	1.5	100	0.00
82	3,3'-Dichlorobenzidine	0.401	0.406	-1.2	97	0.00
83	Chrysene	1.278	1.265	1.0	101	0.00
84	Bis(2-ethylhexyl)phthalate	0.810	0.851	-5.1	99	0.00
85 c	Di-n-octyl phthalate	1.150	1.212	-5.4	97	0.00
86 I	Perylene-d12	1.000	1.000	0.0	113	0.01
87	Indeno(1,2,3-cd)pyrene	1.184	1.254	-5.9	122	0.02
88	Benzo(b)fluoranthene	1.301	1.286	1.2	103	0.00
89	Benzo(k)fluoranthene	1.262	1.200	4.9	108	0.01
90 C	Benzo(a)pyrene	1.025	0.998	2.6	109	0.00
91	Dibenzo(a,h)anthracene	0.972	1.017	-4.6	120	0.02
92	Benzo(g,h,i)perylene	0.932	1.007	-8.0	124	0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF110821\
Data File : BF126057.D
Acq On : 8 Nov 2021 10:31
Operator : JU/CG
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Nov 08 11:37:48 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF110321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 03 17:21:25 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 1