

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF110918\
 Data File : BF110627.D
 Acq On : 9 Nov 2018 15:02
 Operator : JU/SJ
 Sample : J5813-08
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 TS-H

Manual Integrations
 APPROVED

Sohil
 11/12/2018 10:30:09 AM

Quant Time: Nov 09 17:23:25 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF110818.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 09 15:13:13 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.95	152	78708	20.00	ng	0.00
21) Naphthalene-d8	8.23	136	292007	20.00	ng	0.00
39) Acenaphthene-d10	9.99	164	114947	20.00	ng	0.00
64) Phenanthrene-d10	11.48	188	185876	20.00	ng	0.00
76) Chrysene-d12	14.13	240	156127	20.00	ng	0.00
87) Perylene-d12	15.66	264	107646	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.57	112	464110	95.78	ng	0.00
7) Phenol-d6	6.57	99	589823	95.19	ng	0.00
23) Nitrobenzene-d5	7.51	82	358179	70.64	ng	0.00
42) 2,4,6-Tribromophenol	10.78	330	88236	76.18	ng	0.00
45) 2-Fluorobiphenyl	9.30	172	562635	69.91	ng	0.00
79) Terphenyl-d14	13.06	244	441931	53.01	ng	0.00
Target Compounds						
10) Phenol	6.58	94	38726	5.390	ng	Qvalue # 72
50) Dimethylphthalate	9.69	163	81797	9.537	ng	100
71) Phenanthrene	11.50	178	42375	4.255	ng	99
75) Fluoranthene	12.70	202	88607	8.875	ng	98
78) Pyrene	12.93	202	69372	5.771	ng	98
80) Butylbenzylphthalate	13.53	149	22726	4.392	ng	96
81) Benzo(a)anthracene	14.12	228	26434	2.833	ng	98
83) Chrysene	14.16	228	32282	3.631	ng	99
88) Benzo(b)fluoranthene	15.20	252	31083m	4.827	ng	
90) Benzo(a)pyrene	15.59	252	18316	3.130	ng	# 94
92) Benzo(g,h,i)perylene	17.70	276	11480	2.337	ng	# 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF110918\
Data File : BF110627.D
Acq On : 9 Nov 2018 15:02
Operator : JU/SJ
Sample : J5813-08
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
TS-H

Manual Integrations
APPROVED

Sohil
11/12/2018 10:30:09 AM

Quant Time: Nov 09 17:23:25 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF110818.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 09 15:13:13 2018
Response via : Initial Calibration

