

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF110918\  
 Data File : BF110636.D  
 Acq On : 9 Nov 2018 19:20  
 Operator : JU/SJ  
 Sample : J5864-01  
 Misc :  
 ALS Vial : 7 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 OLDMANS-1

Quant Time: Nov 11 03:40:31 2018  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF110818.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Fri Nov 09 15:13:13 2018  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units | Dev(Min)    |
|-----------------------------|-------|------|----------|--------|-------|-------------|
| 1) 1,4-Dichlorobenzene-d4   | 6.95  | 152  | 74371    | 20.00  | ng    | 0.00        |
| 21) Naphthalene-d8          | 8.23  | 136  | 304553   | 20.00  | ng    | 0.00        |
| 39) Acenaphthene-d10        | 9.99  | 164  | 151027   | 20.00  | ng    | 0.00        |
| 64) Phenanthrene-d10        | 11.48 | 188  | 305586   | 20.00  | ng    | 0.00        |
| 76) Chrysene-d12            | 14.13 | 240  | 242609   | 20.00  | ng    | -0.01       |
| 87) Perylene-d12            | 15.66 | 264  | 148782   | 20.00  | ng    | 0.00        |
| System Monitoring Compounds |       |      |          |        |       |             |
| 5) 2-Fluorophenol           | 5.57  | 112  | 606763   | 132.52 | ng    | 0.00        |
| 7) Phenol-d6                | 6.57  | 99   | 768084   | 131.19 | ng    | 0.00        |
| 23) Nitrobenzene-d5         | 7.51  | 82   | 483525   | 91.43  | ng    | 0.00        |
| 42) 2,4,6-Tribromophenol    | 10.78 | 330  | 187810   | 123.41 | ng    | 0.00        |
| 45) 2-Fluorobiphenyl        | 9.30  | 172  | 885788   | 83.76  | ng    | 0.00        |
| 79) Terphenyl-d14           | 13.06 | 244  | 948344   | 73.21  | ng    | 0.00        |
| Target Compounds            |       |      |          |        |       |             |
| 10) Phenol                  | 6.58  | 94   | 50192    | 7.393  | ng    | Qvalue # 69 |
| 50) Dimethylphthalate       | 9.69  | 163  | 56200    | 4.987  | ng    | 100         |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF110918\  
Data File : BF110636.D  
Acq On : 9 Nov 2018 19:20  
Operator : JU/SJ  
Sample : J5864-01  
Misc :  
ALS Vial : 7 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
OLDMANS-1

Quant Time: Nov 11 03:40:31 2018  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF110818.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Fri Nov 09 15:13:13 2018  
Response via : Initial Calibration

