

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF110918\  
 Data File : BF110652.D  
 Acq On : 10 Nov 2018 2:37  
 Operator : JU/SJ  
 Sample : J5887-15 5X  
 Misc :  
 ALS Vial : 23 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 W-3-S

Quant Time: Nov 11 04:12:25 2018  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF110818.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Fri Nov 09 15:13:13 2018  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc  | Units | Dev(Min) |
|-----------------------------|-------|------|----------|-------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4   | 6.95  | 152  | 79878    | 20.00 | ng    | 0.00     |
| 21) Naphthalene-d8          | 8.23  | 136  | 323323   | 20.00 | ng    | 0.00     |
| 39) Acenaphthene-d10        | 9.99  | 164  | 133407   | 20.00 | ng    | 0.00     |
| 64) Phenanthrene-d10        | 11.48 | 188  | 199572   | 20.00 | ng    | 0.00     |
| 76) Chrysene-d12            | 14.13 | 240  | 145904   | 20.00 | ng    | 0.00     |
| 87) Perylene-d12            | 15.66 | 264  | 103712   | 20.00 | ng    | 0.00     |
| System Monitoring Compounds |       |      |          |       |       |          |
| 5) 2-Fluorophenol           | 5.56  | 112  | 57556    | 11.70 | ng    | 0.00     |
| 7) Phenol-d6                | 6.56  | 99   | 76630    | 12.19 | ng    | -0.01    |
| 23) Nitrobenzene-d5         | 7.51  | 82   | 44423    | 7.91  | ng    | 0.00     |
| 42) 2,4,6-Tribromophenol    | 10.78 | 330  | 11728    | 8.72  | ng    | 0.00     |
| 45) 2-Fluorobiphenyl        | 9.30  | 172  | 82416    | 8.82  | ng    | 0.00     |
| 79) Terphenyl-d14           | 13.06 | 244  | 55230    | 7.09  | ng    | -0.01    |

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF110918\  
Data File : BF110652.D  
Acq On : 10 Nov 2018 2:37  
Operator : JU/SJ  
Sample : J5887-15 5X  
Misc :  
ALS Vial : 23 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
W-3-S

Quant Time: Nov 11 04:12:25 2018  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF110818.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Fri Nov 09 15:13:13 2018  
Response via : Initial Calibration

