

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF111120\
 Data File : BF122303.D
 Acq On : 11 Nov 2020 9:07
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 11 10:55:31 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF110920.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 10 12:25:42 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	110	0.00
2	1,4-Dioxane	40.000	39.832	0.4	112	0.00
3	Pyridine	40.000	39.897	0.3	111	0.00
4	n-Nitrosodimethylamine	40.000	39.867	0.3	111	0.00
5 S	2-Fluorophenol	80.000	79.958	0.1	111	0.00
6	Aniline	40.000	39.605	1.0	110	0.00
7 S	Phenol-d6	80.000	79.338	0.8	110	0.00
8	2-Chlorophenol	40.000	40.168	-0.4	110	0.00
9	Benzaldehyde	40.000	38.137	4.7	108	0.00
10 C	Phenol	40.000	39.731	0.7	110	0.00
11	bis(2-Chloroethyl)ether	40.000	39.160	2.1	109	0.00
12	1,3-Dichlorobenzene	40.000	39.269	1.8	109	0.00
13 C	1,4-Dichlorobenzene	40.000	39.280	1.8	110	0.00
14	1,2-Dichlorobenzene	40.000	39.586	1.0	110	0.00
15	Benzyl Alcohol	40.000	40.003	-0.0	109	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	38.803	3.0	108	0.00
17	2-Methylphenol	40.000	39.678	0.8	111	0.00
18	Hexachloroethane	40.000	40.336	-0.8	111	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.552	3.6	108	0.00
20	3+4-Methylphenols	40.000	39.971	0.1	110	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	111	0.00
22	Acetophenone	40.000	38.676	3.3	109	0.00
23 S	Nitrobenzene-d5	80.000	88.701	-10.9	116	0.00
24	Nitrobenzene	40.000	41.972	-4.9	112	0.00
25	Isophorone	40.000	38.376	4.1	109	0.00
26 C	2-Nitrophenol	40.000	40.913	-2.3	125	0.00
27	2,4-Dimethylphenol	40.000	40.131	-0.3	112	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.314	4.2	108	0.00
29 C	2,4-Dichlorophenol	40.000	40.996	-2.5	113	0.00
30	1,2,4-Trichlorobenzene	40.000	39.534	1.2	109	0.00
31	Naphthalene	40.000	38.875	2.8	109	0.00
32	Benzoic acid	40.000	43.103	-7.8	130	0.00
33	4-Chloroaniline	40.000	39.355	1.6	110	0.00
34 C	Hexachlorobutadiene	40.000	39.062	2.3	109	0.00
35	Caprolactam	40.000	40.023	-0.1	111	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.894	0.3	110	0.00
37	2-Methylnaphthalene	40.000	38.788	3.0	110	0.00
38	1-Methylnaphthalene	40.000	38.672	3.3	108	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	109	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.977	0.1	111	0.00
41 P	Hexachlorocyclopentadiene	40.000	42.397	-6.0	111	0.00
42 S	2,4,6-Tribromophenol	80.000	84.978	-6.2	112	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.570	-3.9	111	0.00
44	2,4,5-Trichlorophenol	40.000	40.701	-1.8	109	0.00
45 S	2-Fluorobiphenyl	80.000	76.222	4.7	109	0.00
46	1,1'-Biphenyl	40.000	39.223	1.9	110	0.00
47	2-Chloronaphthalene	40.000	39.593	1.0	110	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF111120\
 Data File : BF122303.D
 Acq On : 11 Nov 2020 9:07
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 11 10:55:31 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF110920.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 10 12:25:42 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.004	-2.5	120	0.00
49	Acenaphthylene	40.000	39.399	1.5	109	0.00
50	Dimethylphthalate	40.000	39.242	1.9	111	0.00
51	2,6-Dinitrotoluene	40.000	40.662	-1.7	117	0.00
52 C	Acenaphthene	40.000	39.626	0.9	110	0.00
53	3-Nitroaniline	40.000	40.625	-1.6	117	0.00
54 P	2,4-Dinitrophenol	40.000	46.666	-16.7	149	0.00
55	Dibenzofuran	40.000	39.434	1.4	110	0.00
56 P	4-Nitrophenol	40.000	40.388	-1.0	116	0.00
57	2,4-Dinitrotoluene	40.000	41.681	-4.2	120	0.00
58	Fluorene	40.000	38.854	2.9	109	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	42.719	-6.8	112	0.00
60	Diethylphthalate	40.000	39.270	1.8	110	0.00
61	4-Chlorophenyl-phenylether	40.000	39.049	2.4	109	0.00
62	4-Nitroaniline	40.000	40.710	-1.8	117	0.00
63	Azobenzene	40.000	38.779	3.1	109	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	111	0.00
65	4,6-Dinitro-2-methylphenol	40.000	44.614	-11.5	135	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.048	2.4	110	0.00
67	4-Bromophenyl-phenylether	40.000	39.638	0.9	110	0.00
68	Hexachlorobenzene	40.000	39.890	0.3	112	0.00
69	Atrazine	40.000	40.093	-0.2	111	0.00
70 C	Pentachlorophenol	40.000	41.926	-4.8	114	0.00
71	Phenanthrene	40.000	38.791	3.0	109	0.00
72	Anthracene	40.000	39.182	2.0	110	0.00
73	Carbazole	40.000	38.952	2.6	110	0.00
74	Di-n-butylphthalate	40.000	38.723	3.2	108	0.00
75 C	Fluoranthene	40.000	39.470	1.3	110	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	109	0.00
77	Benzydine	40.000	38.916	2.7	106	0.00
78	Pyrene	40.000	38.905	2.7	109	0.00
79 S	Terphenyl-d14	80.000	75.530	5.6	109	0.00
80	Butylbenzylphthalate	40.000	38.463	3.8	110	0.00
81	Benzo(a)anthracene	40.000	39.259	1.9	110	0.00
82	3,3'-Dichlorobenzidine	40.000	41.294	-3.2	112	0.00
83	Chrysene	40.000	39.700	0.7	112	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	39.908	0.2	111	0.00
85 c	Di-n-octyl phthalate	40.000	42.275	-5.7	110	0.00
86 I	Perylene-d12	20.000	20.000	0.0	111	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	38.804	3.0	115	0.00
88	Benzo(b)fluoranthene	40.000	38.893	2.8	108	0.00
89	Benzo(k)fluoranthene	40.000	39.655	0.9	114	0.00
90 C	Benzo(a)pyrene	40.000	39.686	0.8	112	0.00
91	Dibenzo(a,h)anthracene	40.000	38.747	3.1	114	0.00
92	Benzo(g,h,i)perylene	40.000	38.599	3.5	116	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF111120\
Data File : BF122303.D
Acq On : 11 Nov 2020 9:07
Operator : JU/CG
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Nov 11 10:55:31 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF110920.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 10 12:25:42 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0