

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF111120\
 Data File : BF122310.D
 Acq On : 11 Nov 2020 13:17
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 11 13:54:40 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF110920.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 10 12:25:42 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	116	0.00
2	1,4-Dioxane	40.000	37.954	5.1	112	-0.01
3	Pyridine	40.000	38.364	4.1	112	-0.01
4	n-Nitrosodimethylamine	40.000	38.371	4.1	112	-0.01
5 S	2-Fluorophenol	80.000	77.482	3.1	113	0.00
6	Aniline	40.000	38.126	4.7	111	0.00
7 S	Phenol-d6	80.000	76.430	4.5	112	0.00
8	2-Chlorophenol	40.000	39.359	1.6	113	0.00
9	Benzaldehyde	40.000	37.193	7.0	111	0.00
10 C	Phenol	40.000	38.362	4.1	111	0.00
11	bis(2-Chloroethyl)ether	40.000	39.035	2.4	114	0.00
12	1,3-Dichlorobenzene	40.000	38.935	2.7	113	0.00
13 C	1,4-Dichlorobenzene	40.000	38.568	3.6	113	0.00
14	1,2-Dichlorobenzene	40.000	38.967	2.6	113	0.00
15	Benzyl Alcohol	40.000	38.889	2.8	111	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	38.933	2.7	114	0.00
17	2-Methylphenol	40.000	38.909	2.7	114	0.00
18	Hexachloroethane	40.000	40.236	-0.6	116	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.666	3.3	113	0.00
20	3+4-Methylphenols	40.000	38.788	3.0	112	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	114	0.00
22	Acetophenone	40.000	38.464	3.8	112	0.00
23 S	Nitrobenzene-d5	80.000	89.771	-12.2	121	0.00
24	Nitrobenzene	40.000	42.438	-6.1	117	0.00
25	Isophorone	40.000	39.159	2.1	115	0.00
26 C	2-Nitrophenol	40.000	43.281	-8.2	138	0.00
27	2,4-Dimethylphenol	40.000	40.283	-0.7	116	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.926	0.2	116	0.00
29 C	2,4-Dichlorophenol	40.000	41.250	-3.1	117	0.00
30	1,2,4-Trichlorobenzene	40.000	39.787	0.5	114	0.00
31	Naphthalene	40.000	38.673	3.3	112	0.00
32	Benzoic acid	40.000	44.340	-10.9	139	0.00
33	4-Chloroaniline	40.000	38.514	3.7	111	0.00
34 C	Hexachlorobutadiene	40.000	40.794	-2.0	118	0.00
35	Caprolactam	40.000	40.426	-1.1	115	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.032	-0.1	114	0.00
37	2-Methylnaphthalene	40.000	38.898	2.8	114	0.00
38	1-Methylnaphthalene	40.000	39.050	2.4	113	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	116	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.544	1.1	117	0.00
41 P	Hexachlorocyclopentadiene	40.000	42.860	-7.1	119	0.00
42 S	2,4,6-Tribromophenol	80.000	84.712	-5.9	119	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.008	-2.5	116	0.00
44	2,4,5-Trichlorophenol	40.000	40.314	-0.8	115	0.00
45 S	2-Fluorobiphenyl	80.000	73.782	7.8	112	0.00
46	1,1'-Biphenyl	40.000	38.324	4.2	114	0.00
47	2-Chloronaphthalene	40.000	38.941	2.6	115	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF111120\
 Data File : BF122310.D
 Acq On : 11 Nov 2020 13:17
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 11 13:54:40 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF110920.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 10 12:25:42 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	39.770	0.6	123	0.00
49	Acenaphthylene	40.000	38.641	3.4	114	0.00
50	Dimethylphthalate	40.000	38.750	3.1	116	0.00
51	2,6-Dinitrotoluene	40.000	41.197	-3.0	126	0.00
52 C	Acenaphthene	40.000	38.968	2.6	115	0.00
53	3-Nitroaniline	40.000	39.400	1.5	120	0.00
54 P	2,4-Dinitrophenol	40.000	48.506	-21.3	169	0.00
55	Dibenzofuran	40.000	38.443	3.9	114	0.00
56 P	4-Nitrophenol	40.000	39.259	1.9	120	0.00
57	2,4-Dinitrotoluene	40.000	41.903	-4.8	129	0.00
58	Fluorene	40.000	37.930	5.2	113	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	42.109	-5.3	118	0.00
60	Diethylphthalate	40.000	39.533	1.2	118	0.00
61	4-Chlorophenyl-phenylether	40.000	39.064	2.3	116	0.00
62	4-Nitroaniline	40.000	39.492	1.3	120	0.00
63	Azobenzene	40.000	38.513	3.7	115	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	114	0.00
65	4,6-Dinitro-2-methylphenol	40.000	46.726	-16.8	149	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.703	0.7	115	0.00
67	4-Bromophenyl-phenylether	40.000	41.085	-2.7	118	0.00
68	Hexachlorobenzene	40.000	40.386	-1.0	117	0.00
69	Atrazine	40.000	41.160	-2.9	117	0.00
70 C	Pentachlorophenol	40.000	43.090	-7.7	122	0.00
71	Phenanthrene	40.000	39.274	1.8	114	0.00
72	Anthracene	40.000	38.916	2.7	113	0.00
73	Carbazole	40.000	38.546	3.6	112	0.00
74	Di-n-butylphthalate	40.000	40.488	-1.2	117	0.00
75 C	Fluoranthene	40.000	38.762	3.1	112	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	108	0.00
77	Benzydine	40.000	39.193	2.0	106	0.00
78	Pyrene	40.000	39.842	0.4	111	0.00
79 S	Terphenyl-d14	80.000	79.209	1.0	114	0.00
80	Butylbenzylphthalate	40.000	40.891	-2.2	117	0.00
81	Benzo(a)anthracene	40.000	38.925	2.7	108	0.00
82	3,3'-Dichlorobenzidine	40.000	41.109	-2.8	110	0.00
83	Chrysene	40.000	39.276	1.8	110	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	42.313	-5.8	116	0.00
85 c	Di-n-octyl phthalate	40.000	43.769	-9.4	113	0.00
86 I	Perylene-d12	20.000	20.000	0.0	102	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	37.356	6.6	101	0.00
88	Benzo(b)fluoranthene	40.000	40.459	-1.1	103	0.00
89	Benzo(k)fluoranthene	40.000	40.551	-1.4	107	0.00
90 C	Benzo(a)pyrene	40.000	39.812	0.5	103	0.00
91	Dibenzo(a,h)anthracene	40.000	37.551	6.1	101	0.00
92	Benzo(g,h,i)perylene	40.000	36.804	8.0	102	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF111120\
Data File : BF122310.D
Acq On : 11 Nov 2020 13:17
Operator : JU/CG
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Nov 11 13:54:40 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF110920.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 10 12:25:42 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0