

Data Path : Z:\svoasrv\HPCHEM1\BNA F\Data\BF111218\
 Data File : BF110671.D
 Acq On : 12 Nov 2018 13:43
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 12 16:23:45 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF110818.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 09 15:13:13 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
2	1,4-Dioxane	0.613	0.626	-2.1	102	-0.01
3	Pyridine	1.324	1.374	-3.8	98	-0.01
4	n-Nitrosodimethylamine	0.568	0.612	-7.7	103	0.00
5 S	2-Fluorophenol	1.231	1.272	-3.3	101	0.00
6	Aniline	2.053	2.099	-2.2	104	0.00
7 S	Phenol-d6	1.575	1.625	-3.2	104	0.00
8	2-Chlorophenol	1.443	1.480	-2.6	103	0.00
9	Benzaldehyde	0.893	0.910	-1.9	106	0.00
10 C	Phenol	1.826	1.899	-4.0	104	0.00
11	bis(2-Chloroethyl)ether	1.368	1.397	-2.1	103	0.00
12	1,3-Dichlorobenzene	1.579	1.610	-2.0	102	0.00
13 C	1,4-Dichlorobenzene	1.604	1.636	-2.0	104	0.00
14	1,2-Dichlorobenzene	1.492	1.522	-2.0	104	0.00
15	Benzyl Alcohol	1.232	1.291	-4.8	104	0.00
16	2,2'-oxybis(1-Chloropropane	2.149	2.212	-2.9	103	0.00
17	2-Methylphenol	1.149	1.179	-2.6	104	0.00
18	Hexachloroethane	0.592	0.609	-2.9	103	0.00
19 P	n-Nitroso-di-n-propylamine	1.005	1.051	-4.6	106	0.00
20	3+4-Methylphenols	1.475	1.522	-3.2	104	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	104	0.00
22	Acetophenone	0.466	0.469	-0.6	104	0.00
23 S	Nitrobenzene-d5	0.347	0.354	-2.0	105	0.00
24	Nitrobenzene	0.356	0.362	-1.7	105	0.00
25	Isophorone	0.569	0.571	-0.4	105	0.00
26 C	2-Nitrophenol	0.178	0.183	-2.8	103	0.00
27	2,4-Dimethylphenol	0.270	0.275	-1.9	105	0.00
28	bis(2-Chloroethoxy)methane	0.385	0.387	-0.5	104	0.00
29 C	2,4-Dichlorophenol	0.278	0.283	-1.8	103	0.00
30	1,2,4-Trichlorobenzene	0.314	0.310	1.3	101	0.00
31	Naphthalene	0.939	0.935	0.4	104	0.00
32	Benzoic acid	0.191	0.172	9.9	86	0.00
33	4-Chloroaniline	0.425	0.424	0.2	107	0.00
34 C	Hexachlorobutadiene	0.179	0.178	0.6	103	0.00
35	Caprolactam	0.093	0.098	-5.4	108	0.00
36 C	4-Chloro-3-methylphenol	0.300	0.307	-2.3	105	0.00
37	2-Methylnaphthalene	0.615	0.623	-1.3	105	0.00
38	1-Methylnaphthalene	0.592	0.594	-0.3	105	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	104	0.00
40	1,2,4,5-Tetrachlorobenzene	0.633	0.640	-1.1	105	0.00
41 P	Hexachlorocyclopentadiene	0.203	0.179	11.8	86	0.00
42 S	2,4,6-Tribromophenol	0.202	0.200	1.0	102	0.00
43 C	2,4,6-Trichlorophenol	0.398	0.403	-1.3	103	0.00
44	2,4,5-Trichlorophenol	0.424	0.422	0.5	101	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA F\Data\BF111218\
 Data File : BF110671.D
 Acq On : 12 Nov 2018 13:43
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 12 16:23:45 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF110818.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 09 15:13:13 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.400	1.368	2.3	103	0.00
46	1,1'-Biphenyl	1.866	1.874	-0.4	105	0.00
47	2-Chloronaphthalene	1.344	1.355	-0.8	106	0.00
48	2-Nitroaniline	0.414	0.427	-3.1	105	0.00
49	Acenaphthylene	1.936	1.967	-1.6	105	0.00
50	Dimethylphthalate	1.492	1.491	0.1	105	0.00
51	2,6-Dinitrotoluene	0.328	0.337	-2.7	106	0.00
52 C	Acenaphthene	1.223	1.232	-0.7	106	0.00
53	3-Nitroaniline	0.380	0.390	-2.6	107	0.00
54 P	2,4-Dinitrophenol	0.116	0.114	1.7	96	0.00
55	Dibenzofuran	1.790	1.782	0.4	104	0.00
56 P	4-Nitrophenol	0.252	0.255	-1.2	100	0.00
57	2,4-Dinitrotoluene	0.427	0.438	-2.6	105	0.00
58	Fluorene	1.375	1.380	-0.4	106	0.00
59	2,3,4,6-Tetrachlorophenol	0.334	0.344	-3.0	106	0.00
60	Diethylphthalate	1.476	1.470	0.4	106	0.00
61	4-Chlorophenyl-phenylether	0.712	0.714	-0.3	105	0.00
62	4-Nitroaniline	0.383	0.396	-3.4	107	0.00
63	Azobenzene	1.440	1.465	-1.7	106	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	104	0.00
65	4,6-Dinitro-2-methylphenol	0.101	0.100	1.0	100	0.00
66 c	n-Nitrosodiphenylamine	0.674	0.680	-0.9	106	0.00
67	4-Bromophenyl-phenylether	0.221	0.221	0.0	105	0.00
68	Hexachlorobenzene	0.233	0.236	-1.3	106	0.00
69	Atrazine	0.206	0.207	-0.5	105	0.00
70 C	Pentachlorophenol	0.124	0.126	-1.6	101	0.00
71	Phenanthrene	1.071	1.078	-0.7	107	0.00
72	Anthracene	1.081	1.084	-0.3	106	0.00
73	Carbazole	1.038	1.047	-0.9	105	0.00
74	Di-n-butylphthalate	1.217	1.232	-1.2	105	0.00
75 C	Fluoranthene	1.074	1.082	-0.7	106	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	110	0.00
77	Benzidine	0.550	0.507	7.8	105	0.00
78	Pyrene	1.540	1.510	1.9	108	0.00
79 S	Terphenyl-d14	1.068	1.024	4.1	108	0.00
80	Butylbenzylphthalate	0.663	0.673	-1.5	109	0.00
81	Benzo(a)anthracene	1.195	1.208	-1.1	111	0.00
82	3,3'-Dichlorobenzidine	0.400	0.390	2.5	109	0.00
83	Chrysene	1.139	1.150	-1.0	115	0.00
84	Bis(2-ethylhexyl)phthalate	0.822	0.857	-4.3	115	0.00
85 c	Di-n-octyl phthalate	1.209	1.343	-11.1	127	0.00
86	Indeno(1,2,3-cd)pyrene	0.817	0.782	4.3	108	0.00
87 I	Perylene-d12	1.000	1.000	0.0	112	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA F\Data\BF111218\
Data File : BF110671.D
Acq On : 12 Nov 2018 13:43
Operator : JU/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Nov 12 16:23:45 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF110818.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 09 15:13:13 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.196	1.252	-4.7	114	0.00
89	Benzo(k)fluoranthene	1.182	1.171	0.9	115	0.00
90 C	Benzo(a)pyrene	1.087	1.084	0.3	112	0.00
91	Dibenzo(a,h)anthracene	0.920	0.915	0.5	108	0.00
92	Benzo(q,h,i)perylene	0.913	0.905	0.9	107	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0