

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF111218\
 Data File : BF110687.D
 Acq On : 12 Nov 2018 21:05
 Operator : JU/SJ
 Sample : J5892-02
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PN-STONE

Quant Time: Nov 13 06:04:41 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF110818.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 09 15:13:13 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.95	152	86439	20.00	ng	0.00
21) Naphthalene-d8	8.22	136	359405	20.00	ng	0.00
39) Acenaphthene-d10	9.98	164	173612	20.00	ng	-0.01
64) Phenanthrene-d10	11.48	188	337176	20.00	ng	0.00
76) Chrysene-d12	14.13	240	220667	20.00	ng	-0.01
87) Perylene-d12	15.65	264	162057	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.57	112	522001	98.09	ng	0.01
7) Phenol-d6	6.57	99	590222	86.73	ng	0.00
23) Nitrobenzene-d5	7.51	82	477289	76.48	ng	0.00
42) 2,4,6-Tribromophenol	10.78	330	194955	111.44	ng	0.00
45) 2-Fluorobiphenyl	9.30	172	886411	72.92	ng	0.00
79) Terphenyl-d14	13.06	244	871516	73.96	ng	0.00


Target Compounds Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF111218\
Data File : BF110687.D
Acq On : 12 Nov 2018 21:05
Operator : JU/SJ
Sample : J5892-02
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PN-STONE

Quant Time: Nov 13 06:04:41 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF110818.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 09 15:13:13 2018
Response via : Initial Calibration






#1




1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.95 min Scan# 826
Delta R.T. 0.00 min
Lab File: BF110687.D
Acq: 12 Nov 2018 21:05

Instrument :
BNA_F
ClientSampleId :
PN-STONE

Tot Ion:152 Resp: 86439
Ion Ratio Lower Upper
152 100
150 157.0 122.7 184.1
115 61.1 49.1 73.7




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#5


2-Fluorophenol
Concen: 98.092 ng
RT: 5.57 min Scan# 593
Delta R.T. 0.01 min
Lab File: BF110687.D
Acq: 12 Nov 2018 21:05

Tgt Ion:112 Resp: 522001
Ion Ratio Lower Upper
112 100
64 61.5 44.1 66.1
63 32.3 23.7 35.5



Abundance Ion 112.00 (111.70 to 112.70): E
Ion 64.00 (63.70 to 64.70): BF1
Ion 63.00 (62.70 to 63.70): BF1





#7

Phenol-d6

Concen: 86.734 ng

RT: 6.57 min Scan# 762

Delta R.T. -0.01 min

Lab File: BF110687.D

Acq: 12 Nov 2018 21:05

Instrument :

BNA_F

ClientSampleId :

PN-STONE


Tot Ion: 99 Resp: 590222

Ion Ratio Lower Upper




99 100

42 20.0 15.8 23.6

71 34.7 28.0 42.0



Abundance Ion 99.00 (98.70 to 99.70): BF1
Ion 42.00 (41.70 to 42.70): BF1
Ion 71.00 (70.70 to 71.70): BF1



#21

Naphthalene-d8

Concen: 20.000 ng

RT: 8.22 min Scan# 1043

Delta R.T. -0.01 min

Lab File: BF110687.D

Acq: 12 Nov 2018 21:05

Tgt Ion: 136 Resp: 359405


Ion Ratio Lower Upper

136 100



137 10.9 8.8 13.2


54 8.4 5.4 8.2#

68 5.3 4.2 6.2



Abundance Ion 136.00 (135.70 to 136.70): E
Ion 137.00 (136.70 to 137.70): E
Ion 54.00 (53.70 to 54.70): BF1
Ion 68.00 (67.70 to 68.70): BF1









#23
 Nitrobenzene-d5
 Concen: 76.479 ng
 RT: 7.51 min Scan# 922
 Delta R.T. -0.01 min
 Lab File: BF110687.D
 Acq: 12 Nov 2018 21:05




Instrument :
 BNA_F
 ClientSampleId :
 PN-STONE


Tot Ion	82	Resp	477289
Ion Ratio	Lower	Upper	
82	100		
128	45.5	34.2	51.2
54	52.6	38.6	58.0



#39
 Acenaphthene-d10
 Concen: 20.000 ng
 RT: 9.98 min Scan# 1342
 Delta R.T. -0.01 min
 Lab File: BF110687.D
 Acq: 12 Nov 2018 21:05

Tgt Ion	164	Resp	173612
Ion Ratio	Lower	Upper	
164	100		
162	105.5	83.0	124.6
160	47.0	37.0	55.6









#42
 2,4,6-Tribromophenol
 Concen: 111.444 ng
 RT: 10.78 min Scan# 1478
 Delta R.T. -0.01 min
 Lab File: BF110687.D
 Acq: 12 Nov 2018 21:05

Instrument :
 BNA_F
 ClientSampled :
 PN-STONE

Tot Ion	Ratio	Lower	Upper
330	100		
332	95.9	77.1	115.7
141	37.8	27.0	40.4




Abundance Ion 329.80 (329.50 to 330.50): E
 Ion 331.80 (331.50 to 332.50): E
 Ion 141.00 (140.70 to 141.70): E






#45
 2-Fluorobiphenyl
 Concen: 72.919 ng
 RT: 9.30 min Scan# 1226
 Delta R.T. -0.01 min
 Lab File: BF110687.D
 Acq: 12 Nov 2018 21:05

Tgt Ion	Ratio	Lower	Upper
172	100		
171	36.1	28.6	43.0
170	23.6	19.7	29.5



Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.48 min Scan# 1597
Delta R.T. -0.01 min
Lab File: BF110687.D
Acq: 12 Nov 2018 21:05

Instrument :
BNA_F
ClientSampleId :
PN-STONE

Tot Ion:188 Resp: 337176
Ion Ratio Lower Upper
188 100
94 7.9 7.2 10.8
80 8.6 7.9 11.9






Abundance Ion 188.00 (187.70 to 188.70): E
Ion 94.00 (93.70 to 94.70): BF1
Ion 80.00 (79.70 to 80.70): BF1




#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.13 min Scan# 2047
Delta R.T. -0.01 min
Lab File: BF110687.D
Acq: 12 Nov 2018 21:05

Tgt Ion:240 Resp: 220667
Ion Ratio Lower Upper
240 100
120 11.5 8.3 12.5
236 25.2 21.2 31.8

Abundance Ion 240.00 (239.70 to 240.70): E
Ion 120.00 (119.70 to 120.70): E
Ion 236.00 (235.70 to 236.70): E





#79

Terphenyl-d14

Concen: 73.964 ng

RT: 13.06 min Scan# 1866

Delta R.T. -0.01 min

Lab File: BF110687.D

Acq: 12 Nov 2018 21:05

Instrument :

BNA_F

ClientSampleId :

PN-STONE


Tot Ion:244 Resp: 871516

Ion Ratio Lower Upper

244 100

212 7.0 5.8 8.8




122 12.7 8.7 13.1



Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E



#87

Perylene-d12

Concen: 20.000 ng

RT: 15.65 min Scan# 2306

Delta R.T. -0.01 min

Lab File: BF110687.D

Acq: 12 Nov 2018 21:05


Tgt Ion:264 Resp: 162057

Ion Ratio Lower Upper

264 100

260 22.8 18.7 28.1

265 20.7 16.7 25.1



Abundance Ion 264.00 (263.70 to 264.70): E

Ion 260.00 (259.70 to 260.70): E

Ion 265.00 (264.70 to 265.70): E

