



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF111225\
 Data File : BF144229.D
 Acq On : 12 Nov 2025 15:22
 Operator : RC/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF110525.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Nov 05 18:37:33 2025






AutoFind: Scans 1606, 1607, 1608; Background Corrected with Scan 1600

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
68	69	0.00	2	1.6	260	PASS
69	69	100	100	100.0	16235	PASS
70	69	0.00	2	0.0	0	PASS
197	198	0.00	2	0.3	118	PASS
198	198	100	100	100.0	47075	PASS
199	198	5	9	6.8	3220	PASS
365	198	1	100	4.1	1919	PASS
441	443	0.01	150	84.3	9336	PASS
442	442	100	100	100.0	57352	PASS
443	442	15	24	19.3	11069	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF111225\
 Data File : BF144229.D
 Acq On : 12 Nov 2025 15:22
 Operator : RC/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
ClientSampleId :
 DFTPP

Quant Time: Nov 12 18:08:05 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF110525.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 05 18:37:33 2025
 Response via : Initial Calibration



TIC: BF144229.D\data.ms

(70) Pentachlorophenol (C)




11.204min (-0.012) 67347.02 ng

response	58991
Ion	Exp% Act%
265.70	100.00 100.00
268.00	61.90 60.96
264.00	61.80 60.11
0.00	0.00 0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF111225\
 Data File : BF144229.D
 Acq On : 12 Nov 2025 15:22
 Operator : RC/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Quant Time: Nov 12 18:08:05 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF110525.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 05 18:37:33 2025
 Response via : Initial Calibration



TIC: BF144229.D\data.ms

(77) Benzidine

12.739min (-0.012) 0.00 ng

response 264673

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.60	13.85
183.00	10.40	11.10
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
11/12/2025	BNA_F	<u>BF144229.D</u>
Compound Name	Response	Retention Time
DDT	104834	13.563
DDD		13.274
DDE	114	12.933
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
114	104948	0.11

Instrument :
 BNA_F
 ClientSampled :
 DFTPP