

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF111318\
 Data File : BF110748.D
 Acq On : 14 Nov 2018 5:05
 Operator : JU/SJ
 Sample : J5916-04
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 110918-SIDI-F-U-B

Quant Time: Nov 14 07:21:43 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF110818.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 13 12:43:10 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.95	152	62780	20.00	ng	0.00
21) Naphthalene-d8	8.23	136	240045	20.00	ng	0.00
39) Acenaphthene-d10	9.99	164	92854	20.00	ng	0.00
64) Phenanthrene-d10	11.48	188	150473	20.00	ng	0.00
76) Chrysene-d12	14.13	240	134151	20.00	ng	-0.01
87) Perylene-d12	15.66	264	84098	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.56	112	157498	40.75	ng	0.00
7) Phenol-d6	6.56	99	115528	23.37	ng	-0.01
23) Nitrobenzene-d5	7.51	82	432938	103.87	ng	0.00
42) 2,4,6-Tribromophenol	10.78	330	72421	77.40	ng	0.00
45) 2-Fluorobiphenyl	9.30	172	730808	112.41	ng	0.00
79) Terphenyl-d14	13.06	244	554403	77.40	ng	0.00


Target Compounds Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF111318\
Data File : BF110748.D
Acq On : 14 Nov 2018 5:05
Operator : JU/SJ
Sample : J5916-04
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
110918-SIDI-F-U-B

Quant Time: Nov 14 07:21:43 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF110818.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 13 12:43:10 2018
Response via : Initial Calibration






#1




1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.95 min Scan# 826
Delta R.T. 0.00 min
Lab File: BF110748.D
Acq: 14 Nov 2018 5:05

Instrument :
BNA_F
ClientSampleId :
110918-SIDI-F-U-B

Tot Ion:152 Resp: 62780
Ion Ratio Lower Upper
152 100
150 155.9 122.7 184.1
115 61.5 49.1 73.7




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#5


2-Fluorophenol
Concen: 40.750 ng
RT: 5.56 min Scan# 591
Delta R.T. 0.00 min
Lab File: BF110748.D
Acq: 14 Nov 2018 5:05

Tgt Ion:112 Resp: 157498
Ion Ratio Lower Upper
112 100
64 58.1 44.1 66.1
63 31.5 23.7 35.5



Abundance Ion 112.00 (111.70 to 112.70): E
Ion 64.00 (63.70 to 64.70): BF1
Ion 63.00 (62.70 to 63.70): BF1





#7

Phenol-d6

Concen: 23.375 ng

RT: 6.56 min Scan# 761

Delta R.T. -0.01 min

Lab File: BF110748.D

Acq: 14 Nov 2018 5:05

Instrument :

BNA_F

ClientSampleId :

110918-SIDI-F-U-B


Tot Ion: 99 Resp: 115528

Ion Ratio Lower Upper

99 100

42 19.6 15.8 23.6




71 34.5 28.0 42.0



Abundance Ion 99.00 (98.70 to 99.70): BF1

Ion 42.00 (41.70 to 42.70): BF1

Ion 71.00 (70.70 to 71.70): BF1



#21

Naphthalene-d8

Concen: 20.000 ng

RT: 8.23 min Scan# 1044

Delta R.T. 0.00 min

Lab File: BF110748.D

Acq: 14 Nov 2018 5:05

Tgt Ion: 136 Resp: 240045


Ion Ratio Lower Upper

136 100

137 11.2 8.8 13.2

54 8.0 5.4 8.2

68 4.9 4.2 6.2






Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BF1

Ion 68.00 (67.70 to 68.70): BF1









#23
 Nitrobenzene-d5
 Concen: 103.867 ng
 RT: 7.51 min Scan# 922
 Delta R.T. -0.01 min
 Lab File: BF110748.D
 Acq: 14 Nov 2018 5:05

Instrument :
 BNA_F
 ClientSampleId :
 110918-SIDI-F-U-B

Tot Ion	82	Resp	432938
Ion Ratio	Lower	Upper	
82	100		
128	44.2	34.2	51.2
54	52.6	38.6	58.0




Abundance Ion 82.00 (81.70 to 82.70): BF1
 Ion 128.00 (127.70 to 128.70): BF1
 Ion 54.00 (53.70 to 54.70): BF1






#39
 Acenaphthene-d10
 Concen: 20.000 ng
 RT: 9.99 min Scan# 1343
 Delta R.T. -0.01 min
 Lab File: BF110748.D
 Acq: 14 Nov 2018 5:05

Tgt Ion	164	Resp	92854
Ion Ratio	Lower	Upper	
164	100		
162	104.9	83.0	124.6
160	47.3	37.0	55.6



Abundance Ion 164.00 (163.70 to 164.70): E
 Ion 162.00 (161.70 to 162.70): E
 Ion 160.00 (159.70 to 160.70): E









#42
 2,4,6-Tribromophenol
 Concen: 77.404 ng
 RT: 10.78 min Scan# 1478
 Delta R.T. -0.01 min
 Lab File: BF110748.D
 Acq: 14 Nov 2018 5:05

Instrument :
 BNA_F
 ClientSampleId :
 110918-SIDI-F-U-B

Tot Ion	Ratio	Lower	Upper
330	100		
332	95.9	77.1	115.7
141	38.4	27.0	40.4




Abundance Ion 329.80 (329.50 to 330.50): E
 Ion 331.80 (331.50 to 332.50): E
 Ion 141.00 (140.70 to 141.70): E






#45
 2-Fluorobiphenyl
 Concen: 112.405 ng
 RT: 9.30 min Scan# 1226
 Delta R.T. -0.01 min
 Lab File: BF110748.D
 Acq: 14 Nov 2018 5:05

Tgt Ion	Ratio	Lower	Upper
172	100		
171	34.9	28.6	43.0
170	23.0	19.7	29.5



Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.48 min Scan# 1597
Delta R.T. -0.01 min
Lab File: BF110748.D
Acq: 14 Nov 2018 5:05

Instrument :
BNA_F
ClientSampleId :
110918-SIDI-F-U-B

Tot Ion:188 Resp: 150473
Ion Ratio Lower Upper
188 100
94 7.9 7.2 10.8
80 8.8 7.9 11.9




Abundance Ion 188.00 (187.70 to 188.70): E
Ion 94.00 (93.70 to 94.70): BF1
Ion 80.00 (79.70 to 80.70): BF1






#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.13 min Scan# 2047
Delta R.T. -0.01 min
Lab File: BF110748.D
Acq: 14 Nov 2018 5:05

Tgt Ion:240 Resp: 134151
Ion Ratio Lower Upper
240 100
120 11.6 8.3 12.5
236 24.7 21.2 31.8



Abundance Ion 240.00 (239.70 to 240.70): E
Ion 120.00 (119.70 to 120.70): E
Ion 236.00 (235.70 to 236.70): E





#79

Terphenyl-d14

Concen: 77.395 ng

RT: 13.06 min Scan# 1866

Delta R.T. -0.01 min

Lab File: BF110748.D

Acq: 14 Nov 2018 5:05

Instrument :

BNA_F

ClientSampleId :

110918-SIDI-F-U-B


Tot Ion:244 Resp: 554403

Ion Ratio Lower Upper

244 100

212 6.3 5.8 8.8




122 9.8 8.7 13.1



Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E



#87

Perylene-d12

Concen: 20.000 ng

RT: 15.66 min Scan# 2307

Delta R.T. 0.00 min

Lab File: BF110748.D

Acq: 14 Nov 2018 5:05


Tgt Ion:264 Resp: 84098

Ion Ratio Lower Upper

264 100

260 22.7 18.7 28.1

265 21.6 16.7 25.1



Abundance Ion 264.00 (263.70 to 264.70): E

Ion 260.00 (259.70 to 260.70): E

Ion 265.00 (264.70 to 265.70): E

