

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF111419\
 Data File : BF117784.D
 Acq On : 14 Nov 2019 9:16
 Operator : JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 14 10:40:38 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF111319.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 13 18:36:07 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	106	0.00
2	1,4-Dioxane	0.528	0.519	1.7	106	0.00
3	Pyridine	1.385	1.389	-0.3	108	0.00
4	n-Nitrosodimethylamine	0.605	0.622	-2.8	112	0.00
5 S	2-Fluorophenol	1.130	1.122	0.7	110	0.00
6	Aniline	1.801	1.894	-5.2	113	0.00
7 S	Phenol-d6	1.363	1.378	-1.1	112	0.00
8	2-Chlorophenol	1.226	1.289	-5.1	113	0.00
9	Benzaldehyde	0.851	0.906	-6.5	108	0.00
10 C	Phenol	1.416	1.470	-3.8	113	0.00
11	bis(2-Chloroethyl)ether	1.137	1.184	-4.1	112	0.00
12	1,3-Dichlorobenzene	1.443	1.493	-3.5	111	0.00
13 C	1,4-Dichlorobenzene	1.459	1.513	-3.7	111	0.00
14	1,2-Dichlorobenzene	1.395	1.405	-0.7	110	0.00
15	Benzyl Alcohol	1.052	1.130	-7.4	115	0.00
16	2,2'-oxybis(1-Chloropropane	1.751	1.850	-5.7	110	0.00
17	2-Methylphenol	1.038	1.095	-5.5	113	0.00
18	Hexachloroethane	0.536	0.566	-5.6	113	0.00
19 P	n-Nitroso-di-n-propylamine	0.841	0.883	-5.0	114	0.00
20	3+4-Methylphenols	1.289	1.327	-2.9	111	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	105	0.00
22	Acetophenone	0.449	0.464	-3.3	113	0.00
23 S	Nitrobenzene-d5	0.336	0.353	-5.1	113	0.00
24	Nitrobenzene	0.347	0.365	-5.2	113	0.00
25	Isophorone	0.617	0.632	-2.4	113	0.00
26 C	2-Nitrophenol	0.169	0.179	-5.9	114	0.00
27	2,4-Dimethylphenol	0.263	0.274	-4.2	113	0.00
28	bis(2-Chloroethoxy)methane	0.391	0.400	-2.3	112	0.00
29 C	2,4-Dichlorophenol	0.285	0.296	-3.9	113	0.00
30	1,2,4-Trichlorobenzene	0.330	0.341	-3.3	112	0.00
31	Naphthalene	0.932	0.963	-3.3	111	0.00
32	Benzoic acid	0.220	0.237	-7.7	118	0.01
33	4-Chloroaniline	0.417	0.438	-5.0	114	0.00
34 C	Hexachlorobutadiene	0.193	0.199	-3.1	111	0.00
35	Caprolactam	0.090	0.097	-7.8	119	0.00
36 C	4-Chloro-3-methylphenol	0.289	0.299	-3.5	114	0.00
37	2-Methylnaphthalene	0.630	0.656	-4.1	113	0.00
38	1-Methylnaphthalene	0.596	0.615	-3.2	112	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	106	0.00
40	1,2,4,5-Tetrachlorobenzene	0.581	0.601	-3.4	114	0.00
41 P	Hexachlorocyclopentadiene	0.325	0.329	-1.2	111	0.00
42 S	2,4,6-Tribromophenol	0.212	0.220	-3.8	115	0.00
43 C	2,4,6-Trichlorophenol	0.416	0.431	-3.6	113	0.00
44	2,4,5-Trichlorophenol	0.432	0.446	-3.2	112	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF111419\
 Data File : BF117784.D
 Acq On : 14 Nov 2019 9:16
 Operator : JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 14 10:40:38 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF111319.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 13 18:36:07 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.115	1.202	-7.8	111	0.00
46	1,1'-Biphenyl	1.484	1.532	-3.2	111	0.00
47	2-Chloronaphthalene	1.221	1.260	-3.2	113	0.00
48	2-Nitroaniline	0.369	0.397	-7.6	117	0.00
49	Acenaphthylene	1.780	1.827	-2.6	111	0.00
50	Dimethylphthalate	1.402	1.447	-3.2	114	0.00
51	2,6-Dinitrotoluene	0.307	0.329	-7.2	116	0.00
52 C	Acenaphthene	1.140	1.174	-3.0	113	0.00
53	3-Nitroaniline	0.367	0.390	-6.3	116	0.00
54 P	2,4-Dinitrophenol	0.136	0.148	-8.8	123	0.00
55	Dibenzofuran	1.579	1.634	-3.5	113	0.00
56 P	4-Nitrophenol	0.274	0.296	-8.0	117	0.00
57	2,4-Dinitrotoluene	0.390	0.426	-9.2	120	0.00
58	Fluorene	1.223	1.239	-1.3	115	0.00
59	2,3,4,6-Tetrachlorophenol	0.355	0.370	-4.2	114	0.00
60	Diethylphthalate	1.367	1.440	-5.3	115	0.00
61	4-Chlorophenyl-phenylether	0.621	0.637	-2.6	112	0.00
62	4-Nitroaniline	0.367	0.396	-7.9	123	0.00
63	Azobenzene	1.244	1.292	-3.9	114	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	109	0.00
65	4,6-Dinitro-2-methylphenol	0.093	0.101	-8.6	122	0.00
66 c	n-Nitrosodiphenylamine	0.582	0.592	-1.7	114	0.00
67	4-Bromophenyl-phenylether	0.216	0.218	-0.9	113	0.00
68	Hexachlorobenzene	0.252	0.261	-3.6	115	0.00
69	Atrazine	0.191	0.187	2.1	109	0.00
70 C	Pentachlorophenol	0.147	0.152	-3.4	115	0.00
71	Phenanthrene	1.006	0.995	1.1	114	0.00
72	Anthracene	0.997	0.991	0.6	115	0.00
73	Carbazole	0.871	0.898	-3.1	116	0.00
74	Di-n-butylphthalate	1.085	1.084	0.1	116	0.00
75 C	Fluoranthene	0.973	1.014	-4.2	117	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	118	0.00
77	Benzidine	0.655	0.708	-8.1	127	0.00
78	Pyrene	1.616	1.574	2.6	118	0.00
79 S	Terphenyl-d14	1.094	1.061	3.0	118	0.00
80	Butylbenzylphthalate	0.694	0.713	-2.7	127	0.00
81	Benzo(a)anthracene	1.322	1.349	-2.0	125	0.00
82	3,3'-Dichlorobenzidine	0.462	0.497	-7.6	130	0.00
83	Chrysene	1.281	1.297	-1.2	127	0.00
84	Bis(2-ethylhexyl)phthalate	0.842	0.884	-5.0	132	0.00
85 c	Di-n-octyl phthalate	1.236	1.341	-8.5	138	0.00
86	Indeno(1,2,3-cd)pyrene	1.090	1.066	2.2	133	0.00
87 I	Perylene-d12	1.000	1.000	0.0	127	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF111419\
Data File : BF117784.D
Acq On : 14 Nov 2019 9:16
Operator : JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Nov 14 10:40:38 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF111319.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 13 18:36:07 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.299	1.322	-1.8	131	0.00
89	Benzo(k)fluoranthene	1.224	1.255	-2.5	138	0.00
90 C	Benzo(a)pyrene	1.143	1.164	-1.8	134	0.00
91	Dibenzo(a,h)anthracene	0.983	0.976	0.7	135	0.00
92	Benzo(q,h,i)perylene	0.980	0.961	1.9	135	0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0