

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF111524\
Data File : BF140392.D
Acq On : 15 Nov 2024 10:08
Operator : RC/JU
Sample : PB164789BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB164789BL

Quant Time: Nov 15 10:41:08 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF111324.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 13 14:40:06 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.875	152	139459	20.000	ng	0.00
21) Naphthalene-d8	8.157	136	521031	20.000	ng	0.00
39) Acenaphthene-d10	9.910	164	292158	20.000	ng	0.00
64) Phenanthrene-d10	11.404	188	556998	20.000	ng	0.00
76) Chrysene-d12	14.057	240	390196	20.000	ng	0.00
86) Perylene-d12	15.580	264	333601	20.000	ng	0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.522	112	1002140	122.691	ng	0.03
7) Phenol-d6	6.522	99	1305353	117.957	ng	0.02
23) Nitrobenzene-d5	7.440	82	840905	84.090	ng	0.00
42) 2,4,6-Tribromophenol	10.704	330	351266	120.906	ng	0.00
45) 2-Fluorobiphenyl	9.228	172	1519879	83.629	ng	0.00
79) Terphenyl-d14	12.992	244	1756649	78.145	ng	0.00


Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF111524\
Data File : BF140392.D
Acq On : 15 Nov 2024 10:08
Operator : RC/JU
Sample : PB164789BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB164789BL


Quant Time: Nov 15 10:41:08 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF111324.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 13 14:40:06 2024
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.875 min Scan# 813
Delta R.T. -0.006 min
Lab File: BF140392.D
Acq: 15 Nov 2024 10:08

Instrument :
BNA_F
ClientSampleId :
PB164789BL






Tgt Ion:152 Resp: 139459
Ion Ratio Lower Upper
152 100
150 157.1 127.4 191.0
115 60.8 47.4 71.2



#5
2-Fluorophenol
Concen: 122.691 ng
RT: 5.522 min Scan# 583
Delta R.T. 0.029 min
Lab File: BF140392.D
Acq: 15 Nov 2024 10:08

Tgt Ion:112 Resp: 1002140
Ion Ratio Lower Upper
112 100
64 60.9 49.2 73.8
63 31.9 25.6 38.4









#7
Phenol-d6
Concen: 117.957 ng
RT: 6.522 min Scan# 71
Delta R.T. 0.018 min
Lab File: BF140392.D
Acq: 15 Nov 2024 10:08




Instrument :
BNA_F
ClientSampleId :
PB164789BL

Tgt Ion: 99 Resp: 1305353
Ion Ratio Lower Upper
99 100
42 18.6 15.1 22.7
71 35.1 27.9 41.9



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.157 min Scan# 1031
Delta R.T. -0.000 min
Lab File: BF140392.D
Acq: 15 Nov 2024 10:08

Tgt Ion: 136 Resp: 521031
Ion Ratio Lower Upper
136 100
137 10.7 8.7 13.1
54 7.5 6.5 9.7
68 6.2 5.0 7.6





#23

Nitrobenzene-d5

Concen: 84.090 ng

RT: 7.440 min Scan# 909

Delta R.T. -0.000 min

Lab File: BF140392.D

Acq: 15 Nov 2024 10:08

Instrument :

BNA_F

ClientSampleId :

PB164789BL





Tgt Ion: 82 Resp: 840905

Ion Ratio Lower Upper

82 100

128 42.9 33.3 49.9

54 48.4 39.8 59.8



#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.910 min Scan# 1329

Delta R.T. -0.006 min

Lab File: BF140392.D

Acq: 15 Nov 2024 10:08




Tgt Ion:164 Resp: 292158

Ion Ratio Lower Upper

164 100

162 100.4 79.8 119.8


160 44.3 35.4 53.0








#42
 2,4,6-Tribromophenol
 Concen: 120.906 ng
 RT: 10.704 min Scan# 1464
 Delta R.T. -0.000 min
 Lab File: BF140392.D
 Acq: 15 Nov 2024 10:08

Instrument :
 BNA_F
 ClientSampleId :
 PB164789BL







Tgt Ion:330 Resp: 351266
 Ion Ratio Lower Upper
 330 100
 332 96.7 75.9 113.9
 141 33.6 26.9 40.3



#45
 2-Fluorobiphenyl
 Concen: 83.629 ng
 RT: 9.228 min Scan# 1213
 Delta R.T. -0.006 min
 Lab File: BF140392.D
 Acq: 15 Nov 2024 10:08

Tgt Ion:172 Resp: 1519879
 Ion Ratio Lower Upper
 172 100
 171 35.4 28.5 42.7
 170 23.9 19.1 28.7









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.404 min Scan# 11
Delta R.T. -0.000 min
Lab File: BF140392.D
Acq: 15 Nov 2024 10:08




Instrument :
BNA_F
ClientSampleId :
PB164789BL


Tgt Ion:188 Resp: 556998
Ion Ratio Lower Upper
188 100
94 9.1 7.6 11.4
80 9.8 8.0 12.0



#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.057 min Scan# 2034
Delta R.T. 0.006 min
Lab File: BF140392.D
Acq: 15 Nov 2024 10:08

Tgt Ion:240 Resp: 390196
Ion Ratio Lower Upper
240 100
120 11.1 8.8 13.2
236 26.7 20.9 31.3









#79
 Terphenyl-d14
 Concen: 78.145 ng
 RT: 12.992 min Scan# 1853
 Delta R.T. 0.006 min
 Lab File: BF140392.D
 Acq: 15 Nov 2024 10:08

Instrument :
 BNA_F
 ClientSampleId :
 PB164789BL

Tgt Ion:244 Resp: 1756649
 Ion Ratio Lower Upper
 244 100
 212 7.3 5.8 8.8
 122 10.3 8.8 13.2



#86
 Perylene-d12
 Concen: 20.000 ng
 RT: 15.580 min Scan# 2293
 Delta R.T. 0.024 min
 Lab File: BF140392.D
 Acq: 15 Nov 2024 10:08

Tgt Ion:264 Resp: 333601
 Ion Ratio Lower Upper
 264 100
 260 23.6 19.2 28.8
 265 21.6 17.1 25.7

