

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF111925\
 Data File : BF144287.D
 Acq On : 19 Nov 2025 19:47
 Operator : RC/JU
 Sample : Q3659-01
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 TANK-1-COMP-1

Quant Time: Nov 20 00:03:07 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF110525.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 05 18:37:33 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.869	152	51554	20.000	ng	0.00
21) Naphthalene-d8	8.145	136	194233	20.000	ng	-0.01
39) Acenaphthene-d10	9.904	164	99697	20.000	ng	-0.01
64) Phenanthrene-d10	11.392	188	157810	20.000	ng	-0.01
76) Chrysene-d12	14.045	240	140217	20.000	ng	0.00
86) Perylene-d12	15.521	264	159921	20.000	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.504	112	395053	119.915	ng	0.01
7) Phenol-d6	6.498	99	418171	112.024	ng	0.00
23) Nitrobenzene-d5	7.428	82	297217	88.377	ng	0.00
42) 2,4,6-Tribromophenol	10.698	330	149546	119.533	ng	0.00
45) 2-Fluorobiphenyl	9.222	172	628703	93.137	ng	0.00
79) Terphenyl-d14	12.980	244	613446	69.492	ng	-0.01
Target Compounds						
						Qvalue
9) Benzaldehyde	6.422	77	53954	25.596	ng	99
15) Benzyl Alcohol	7.034	79	1087941	368.068	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF111925\
Data File : BF144287.D
Acq On : 19 Nov 2025 19:47
Operator : RC/JU
Sample : Q3659-01
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
TANK-1-COMP-1

Quant Time: Nov 20 00:03:07 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF110525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 05 18:37:33 2025
Response via : Initial Calibration

