




Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF112024\
Data File : BF140500.D
Acq On : 20 Nov 2024 15:33
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Nov 20 17:01:13 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF111324.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 13 14:40:06 2024
Response via : Initial Calibration



TIC: BF140500.D\data.ms

(70) Pentachlorophenol (C)




11.204min (-0.000) 29007.72 ng

response 185231

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	66.10	68.93
264.00	62.90	64.29
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF112024\
Data File : BF140500.D
Acq On : 20 Nov 2024 15:33
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Nov 20 17:01:13 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF111324.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 13 14:40:06 2024
Response via : Initial Calibration



TIC: BF140500.D\data.ms

(77) Benzidine

12.739min (+ 0.006) 19345.82 ng

response 723942

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.90	14.37
183.00	10.80	11.38
0.00	0.00	0.00