

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF112120\
 Data File : BF122427.D
 Acq On : 21 Nov 2020 18:39
 Operator : JU/CG
 Sample : L4829-18
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 MW-3

Quant Time: Nov 23 01:00:27 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF110920.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 20 13:27:02 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.863	152	198423	20.00	ng	0.00
21) Naphthalene-d8	8.151	136	690864	20.00	ng	0.00
39) Acenaphthene-d10	9.904	164	392658	20.00	ng	-0.01
64) Phenanthrene-d10	11.398	188	723993	20.00	ng	0.00
76) Chrysene-d12	14.039	240	671803	20.00	ng	-0.01
86) Perylene-d12	15.515	264	598954	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.487	112	735291	56.90	ng	0.00
7) Phenol-d6	6.498	99	680057	40.23	ng	-0.01
23) Nitrobenzene-d5	7.428	82	1222353	108.32	ng	0.00
42) 2,4,6-Tribromophenol	10.698	330	622506	147.71	ng	0.00
45) 2-Fluorobiphenyl	9.228	172	2243447	89.27	ng	0.00
79) Terphenyl-d14	12.992	244	2256431	71.16	ng	0.00
Target Compounds					Qvalue	
9) Benzaldehyde	6.422	77	56139	3.67	ng	# 1
15) Benzyl Alcohol	6.928	79	196665	16.36	ng	# 9
18) Hexachloroethane	7.357	117	45374	8.53	ng	# 1
19) n-Nitroso-di-n-propyla...	7.428	70	169110	16.74	ng	# 77
22) Acetophenone	7.263	105	432237	24.02	ng	# 49
24) Nitrobenzene	7.439	77	61782	4.78	ng	# 1
25) Isophorone	7.428	82	1214501	48.27	ng	# 67
26) 2-Nitrophenol	7.769	139	405	6.88	ng	# 1
31) Naphthalene	8.181	128	4518117	123.01	ng	# 95
32) Benzoic acid	7.916	122	17348	10.16	ng	# 1
33) 4-Chloroaniline	8.181	127	700396	43.78	ng	# 36
37) 2-Methylnaphthalene	8.863	142	145668	6.00	ng	97
38) 1-Methylnaphthalene	8.963	142	660976	29.10	ng	100
48) 2-Nitroaniline	9.439	65	5536	5.84	ng	# 45
50) Dimethylphthalate	9.622	163	205896	7.38	ng	98
51) 2,6-Dinitrotoluene	9.728	165	2931	4.25	ng	# 70
53) 3-Nitroaniline	9.863	138	490	3.62	ng	# 7
56) 4-Nitrophenol	10.004	139	738	5.52	ng	# 1
57) 2,4-Dinitrotoluene	10.080	165	3163	5.08	ng	# 73
62) 4-Nitroaniline	10.457	138	1376	3.57	ng	# 73
65) 4,6-Dinitro-2-methylph...	10.698	198	1282	6.66	ng	# 1
67) 4-Bromophenyl-phenylether	10.698	248	40301	4.74	ng	# 1
71) Phenanthrene	11.416	178	153250	3.73	ng	98
73) Carbazole	11.622	167	185441	4.81	ng	98
80) Butylbenzylphthalate	13.457	149	1662	3.37	ng	# 4
84) Bis(2-ethylhexyl)phtha...	14.027	149	101158	4.45	ng	# 97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF112120\
Data File : BF122427.D
Acq On : 21 Nov 2020 18:39
Operator : JU/CG
Sample : L4829-18
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
MW-3

Quant Time: Nov 23 01:00:27 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF110920.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 20 13:27:02 2020
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 6.863 min Scan# 812
Delta R.T. -0.006 min
Lab File: BF122427.D
Acq: 21 Nov 2020 18:39




Instrument :
BNA_F
ClientSampleId :
MW-3


Tgt Ion:152 Resp: 198423
Ion Ratio Lower Upper
152 100
150 155.8 125.3 187.9
115 60.0 51.9 77.9



#5
2-Fluorophenol
Concen: 56.90 ng
RT: 5.487 min Scan# 578
Delta R.T. -0.006 min
Lab File: BF122427.D
Acq: 21 Nov 2020 18:39

Tgt Ion:112 Resp: 735291
Ion Ratio Lower Upper
112 100
64 56.5 50.4 75.6
63 31.1 25.8 38.6









#7
Phenol-d6
Concen: 40.23 ng
RT: 6.498 min Scan# 750
Delta R.T. -0.012 min
Lab File: BF122427.D
Acq: 21 Nov 2020 18:39




Instrument :
BNA_F
ClientSampleId :
MW-3


Tgt Ion: 99 Resp: 680057
Ion Ratio Lower Upper
99 100
42 17.7 15.9 23.9
71 36.0 27.1 40.7



#9
Benzaldehyde
Concen: 3.67 ng
RT: 6.422 min Scan# 737
Delta R.T. 0.006 min
Lab File: BF122427.D
Acq: 21 Nov 2020 18:39

Tgt Ion: 77 Resp: 56139
Ion Ratio Lower Upper
77 100
105 859.6 76.7 116.7#
106 74.4 71.3 111.3









#15
Benzyl Alcohol
Concen: 16.36 ng
RT: 6.928 min Scan# 823
Delta R.T. -0.082 min
Lab File: BF122427.D
Acq: 21 Nov 2020 18:39




Instrument :
BNA_F
ClientSampled :
MW-3


Tgt Ion: 79 Resp: 196665
Ion Ratio Lower Upper
79 100
108 0.2 62.3 93.5#
77 134.9 50.6 75.8#



#18
Hexachloroethane
Concen: 8.53 ng
RT: 7.357 min Scan# 896
Delta R.T. -0.024 min
Lab File: BF122427.D
Acq: 21 Nov 2020 18:39

Tgt Ion: 117 Resp: 45374
Ion Ratio Lower Upper
117 100
119 1228.9 76.3 114.5#
201 0.0 71.8 107.6#









#19
n-Nitroso-di-n-propylamine
Concen: 16.74 ng
RT: 7.428 min Scan# 908
Delta R.T. 0.141 min
Lab File: BF122427.D
Acq: 21 Nov 2020 18:39




Instrument :
BNA_F
ClientSampleId :
MW-3


Tgt Ion: 70 Resp: 169110
Ion Ratio Lower Upper
70 100
42 47.3 50.2 75.4#
101 1.5 7.5 11.3#
130 3.9 15.5 23.3#



#21
Naphthalene-d8
Concen: 20.00 ng
RT: 8.151 min Scan# 1031
Delta R.T. -0.006 min
Lab File: BF122427.D
Acq: 21 Nov 2020 18:39


Tgt Ion: 136 Resp: 690864
Ion Ratio Lower Upper
136 100
137 11.0 8.6 13.0
54 7.9 7.1 10.7
68 6.3 5.0 7.4








#22
Acetophenone
Concen: 24.02 ng
RT: 7.263 min Scan# 880
Delta R.T. -0.018 min
Lab File: BF122427.D
Acq: 21 Nov 2020 18:39

Instrument :
BNA_F
ClientSampled :
MW-3







Tgt Ion:105 Resp: 432237
Ion Ratio Lower Upper
105 100
71 1.0 5.3 7.9#
51 4.2 33.2 49.8#
120 0.7 17.4 26.0#



#23
Nitrobenzene-d5
Concen: 108.32 ng
RT: 7.428 min Scan# 908
Delta R.T. -0.006 min
Lab File: BF122427.D
Acq: 21 Nov 2020 18:39

Tgt Ion: 82 Resp: 1222353
Ion Ratio Lower Upper
82 100
128 44.9 35.0 52.4
54 49.7 43.9 65.9









#24
Nitrobenzene
Concen: 4.78 ng
RT: 7.439 min Scan# 910
Delta R.T. -0.012 min
Lab File: BF122427.D
Acq: 21 Nov 2020 18:39




Instrument :
BNA_F
ClientSampled :
MW-3


Tgt Ion: 77 Resp: 61782
Ion Ratio Lower Upper
77 100
123 0.5 35.6 53.4#
65 96.3 11.2 16.8#



#25
Isophorone
Concen: 48.27 ng
RT: 7.428 min Scan# 908
Delta R.T. -0.265 min
Lab File: BF122427.D
Acq: 21 Nov 2020 18:39

Tgt Ion: 82 Resp: 1214501
Ion Ratio Lower Upper
82 100
95 0.2 5.4 8.0#
138 0.0 13.4 20.0#









#26
2-Nitrophenol
Concen: 6.88 ng
RT: 7.769 min Scan# 966
Delta R.T. -0.000 min
Lab File: BF122427.D
Acq: 21 Nov 2020 18:39




Instrument :
BNA_F
ClientSampleId :
MW-3


Tgt Ion:139 Resp: 405
Ion Ratio Lower Upper
139 100
109 140.9 23.0 34.4#
65 490.3 39.9 59.9#



#31
Naphthalene
Concen: 123.01 ng
RT: 8.181 min Scan# 1036
Delta R.T. -0.000 min
Lab File: BF122427.D
Acq: 21 Nov 2020 18:39

Tgt Ion:128 Resp: 4518117
Ion Ratio Lower Upper
128 100
129 13.1 8.7 13.1#
127 15.6 10.8 16.2









#32
Benzoic acid
Concen: 10.16 ng
RT: 7.916 min Scan# 991
Delta R.T. -0.024 min
Lab File: BF122427.D
Acq: 21 Nov 2020 18:39




Instrument :
BNA_F
ClientSampleId :
MW-3


Tgt Ion:122 Resp: 17348
Ion Ratio Lower Upper
122 100
105 985.8 103.3 143.3#
77 318.6 73.9 113.9#



#33
4-Chloroaniline
Concen: 43.78 ng
RT: 8.181 min Scan# 1036
Delta R.T. -0.047 min
Lab File: BF122427.D
Acq: 21 Nov 2020 18:39

Tgt Ion:127 Resp: 700396
Ion Ratio Lower Upper
127 100
129 84.4 26.3 39.5#
65 4.3 26.2 39.2#
92 0.9 15.9 23.9#









#37
2-Methylnaphthalene
Concen: 6.00 ng
RT: 8.863 min Scan# 1152
Delta R.T. -0.006 min
Lab File: BF122427.D
Acq: 21 Nov 2020 18:39




Instrument :
BNA_F
ClientSampleId :
MW-3


Tgt Ion:142 Resp: 145668
Ion Ratio Lower Upper
142 100
141 91.6 70.2 105.4
115 35.2 28.6 43.0



#38
1-Methylnaphthalene
Concen: 29.10 ng
RT: 8.963 min Scan# 1169
Delta R.T. -0.006 min
Lab File: BF122427.D
Acq: 21 Nov 2020 18:39

Tgt Ion:142 Resp: 660976
Ion Ratio Lower Upper
142 100
141 93.1 74.5 111.7
116 4.3 3.2 4.8





#39

Acenaphthene-d10

Concen: 20.00 ng

RT: 9.904 min Scan# 1329

Delta R.T. -0.012 min

Lab File: BF122427.D

Acq: 21 Nov 2020 18:39

Instrument :

BNA_F

ClientSampleId :

MW-3





Tgt Ion:164 Resp: 392658

Ion Ratio Lower Upper

164 100

162 100.6 83.2 124.8

160 46.9 37.5 56.3



#42

2,4,6-Tribromophenol

Concen: 147.71 ng

RT: 10.698 min Scan# 1464

Delta R.T. -0.006 min

Lab File: BF122427.D

Acq: 21 Nov 2020 18:39




Tgt Ion:330 Resp: 622506


Ion Ratio Lower Upper

330 100

332 96.7 77.3 115.9

141 30.6 22.3 33.5









#45
2-Fluorobiphenyl
Concen: 89.27 ng
RT: 9.228 min Scan# 1214
Delta R.T. -0.006 min
Lab File: BF122427.D
Acq: 21 Nov 2020 18:39




Instrument :
BNA_F
ClientSampleId :
MW-3


Tgt Ion:172 Resp: 2243447
Ion Ratio Lower Upper
172 100
171 36.8 29.1 43.7
170 25.1 19.8 29.6



#48
2-Nitroaniline
Concen: 5.84 ng
RT: 9.439 min Scan# 1250
Delta R.T. -0.018 min
Lab File: BF122427.D
Acq: 21 Nov 2020 18:39

Tgt Ion: 65 Resp: 5536
Ion Ratio Lower Upper
65 100
92 65.2 55.1 82.7
138 11.1 80.5 120.7#









#50
Dimethylphthalate
Concen: 7.38 ng
RT: 9.622 min Scan# 1281
Delta R.T. -0.018 min
Lab File: BF122427.D
Acq: 21 Nov 2020 18:39




Instrument :
BNA_F
ClientSampled :
MW-3


Tgt Ion:163 Resp: 205896
Ion Ratio Lower Upper
163 100
194 4.0 3.3 4.9
164 11.3 8.2 12.4



#51
2,6-Dinitrotoluene
Concen: 4.25 ng
RT: 9.728 min Scan# 1299
Delta R.T. 0.029 min
Lab File: BF122427.D
Acq: 21 Nov 2020 18:39

Tgt Ion:165 Resp: 2931
Ion Ratio Lower Upper
165 100
63 109.8 51.8 77.6#
89 62.6 49.0 73.6









#53
3-Nitroaniline
Concen: 3.62 ng
RT: 9.863 min Scan# 1322
Delta R.T. -0.006 min
Lab File: BF122427.D
Acq: 21 Nov 2020 18:39




Tgt Ion:138 Resp: 490
Ion Ratio Lower Upper
138 100
108 209.2 8.6 13.0#
92 181.1 96.6 144.8#


Instrument :
BNA_F
ClientSampled :
MW-3



#56
4-Nitrophenol
Concen: 5.52 ng
RT: 10.004 min Scan# 1346
Delta R.T. -0.024 min
Lab File: BF122427.D
Acq: 21 Nov 2020 18:39

Tgt Ion:139 Resp: 738
Ion Ratio Lower Upper
139 100
109 159.5 45.4 85.4#
65 575.6 68.4 108.4#









#57
2,4-Dinitrotoluene
Concen: 5.08 ng
RT: 10.080 min Scan# 1359
Delta R.T. -0.018 min
Lab File: BF122427.D
Acq: 21 Nov 2020 18:39




Instrument :
BNA_F
ClientSampled :
MW-3


Tgt Ion:	165	Resp:	3163
Ion	Ratio	Lower	Upper
165	100		
63	51.9	41.3	61.9
89	46.6	61.8	92.8#
182	99.9	2.1	3.1#



#62
4-Nitroaniline
Concen: 3.57 ng
RT: 10.457 min Scan# 1423
Delta R.T. -0.029 min
Lab File: BF122427.D
Acq: 21 Nov 2020 18:39

Tgt Ion:	138	Resp:	1376
Ion	Ratio	Lower	Upper
138	100		
92	58.5	31.0	71.0
108	103.6	51.5	91.5#









#64
Phenanthrene-d10
Concen: 20.00 ng
RT: 11.398 min Scan# 1583
Delta R.T. -0.006 min
Lab File: BF122427.D
Acq: 21 Nov 2020 18:39




Instrument :
BNA_F
ClientSampled :
MW-3


Tgt Ion:188 Resp: 723993
Ion Ratio Lower Upper
188 100
94 8.9 8.3 12.5
80 9.5 9.1 13.7



#65
4,6-Dinitro-2-methylphenol
Concen: 6.66 ng
RT: 10.698 min Scan# 1464
Delta R.T. 0.188 min
Lab File: BF122427.D
Acq: 21 Nov 2020 18:39

Tgt Ion:198 Resp: 1282
Ion Ratio Lower Upper
198 100
51 234.1 29.6 69.6#
105 273.1 24.4 64.4#









#67
4-Bromophenyl-phenylether
Concen: 4.74 ng
RT: 10.698 min Scan# 1464
Delta R.T. -0.247 min
Lab File: BF122427.D
Acq: 21 Nov 2020 18:39




Instrument :
BNA_F
ClientSampled :
MW-3


Tgt Ion:248 Resp: 40301
Ion Ratio Lower Upper
248 100
250 202.6 77.0 115.6#
141 486.5 53.7 80.5#



#71
Phenanthrene
Concen: 3.73 ng
RT: 11.416 min Scan# 1586
Delta R.T. -0.012 min
Lab File: BF122427.D
Acq: 21 Nov 2020 18:39

Tgt Ion:178 Resp: 153250
Ion Ratio Lower Upper
178 100
176 18.9 15.9 23.9
179 15.6 12.2 18.4









#73
 Carbazole
 Concen: 4.81 ng
 RT: 11.622 min Scan# 1621
 Delta R.T. -0.012 min
 Lab File: BF122427.D
 Acq: 21 Nov 2020 18:39

Instrument :
 BNA_F
 ClientSampled :
 MW-3

Tgt Ion:167 Resp: 185441




Ion	Ratio	Lower	Upper
167	100		
166	21.5	17.8	26.6
139	12.3	10.8	16.2




#76
 Chrysene-d12
 Concen: 20.00 ng
 RT: 14.039 min Scan# 2032
 Delta R.T. -0.012 min
 Lab File: BF122427.D
 Acq: 21 Nov 2020 18:39

Tgt Ion:240 Resp: 671803

Ion	Ratio	Lower	Upper
240	100		
120	11.0	9.2	13.8
236	25.6	20.9	31.3









#79
Terphenyl-d14
Concen: 71.16 ng
RT: 12.992 min Scan# 1854
Delta R.T. -0.000 min
Lab File: BF122427.D
Acq: 21 Nov 2020 18:39




Instrument :
BNA_F
ClientSampled :
MW-3


Tgt Ion:244 Resp: 2256431
Ion Ratio Lower Upper
244 100
212 7.3 5.4 8.2
122 10.5 8.1 12.1



#80
Butylbenzylphthalate
Concen: 3.37 ng
RT: 13.457 min Scan# 1933
Delta R.T. -0.012 min
Lab File: BF122427.D
Acq: 21 Nov 2020 18:39

Tgt Ion:149 Resp: 1662
Ion Ratio Lower Upper
149 100
91 181.1 61.0 91.4#
206 20.6 16.7 25.1









#84
Bis(2-ethylhexyl)phthalate
Concen: 4.45 ng
RT: 14.027 min Scan# 2030
Delta R.T. -0.006 min
Lab File: BF122427.D
Acq: 21 Nov 2020 18:39

Instrument :
BNA_F
ClientSampleId :
MW-3

Tgt Ion:149 Resp: 101158
Ion Ratio Lower Upper
149 100
167 27.3 20.9 31.3
279 5.1 3.2 4.8#



#86
Perylene-d12
Concen: 20.00 ng
RT: 15.515 min Scan# 2283
Delta R.T. -0.018 min
Lab File: BF122427.D
Acq: 21 Nov 2020 18:39

Tgt Ion:264 Resp: 598954
Ion Ratio Lower Upper
264 100
260 23.7 18.7 28.1
265 21.2 17.3 25.9

