

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF112125\
 Data File : BF144317.D
 Acq On : 21 Nov 2025 19:49
 Operator : RC/JU
 Sample : Q3678-02
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 COMPOSITE-STONE

Quant Time: Nov 21 22:05:37 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF110525.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 05 18:37:33 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.863	152	50249	20.000	ng	-0.01
21) Naphthalene-d8	8.145	136	183689	20.000	ng	-0.01
39) Acenaphthene-d10	9.904	164	92301	20.000	ng	-0.01
64) Phenanthrene-d10	11.392	188	161310	20.000	ng	-0.01
76) Chrysene-d12	14.045	240	151042	20.000	ng	0.00
86) Perylene-d12	15.521	264	114429	20.000	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.498	112	346683	107.965	ng	0.00
7) Phenol-d6	6.498	99	351509	96.611	ng	0.00
23) Nitrobenzene-d5	7.428	82	256918	80.779	ng	0.00
42) 2,4,6-Tribromophenol	10.698	330	139140	120.127	ng	0.00
45) 2-Fluorobiphenyl	9.222	172	513715	82.201	ng	0.00
79) Terphenyl-d14	12.980	244	649470	68.300	ng	-0.01

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF112125\
Data File : BF144317.D
Acq On : 21 Nov 2025 19:49
Operator : RC/JU
Sample : Q3678-02
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
COMPOSITE-STONE

Quant Time: Nov 21 22:05:37 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF110525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 05 18:37:33 2025
Response via : Initial Calibration

