

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF112320\
 Data File : BF122460.D
 Acq On : 23 Nov 2020 22:01
 Operator : JU/CG
 Sample : L4778-07
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 358

Manual Integrations
 APPROVED

mohammad
 11/25/2020 12:51:40 PM

Quant Time: Nov 24 00:28:34 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF110920.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 20 13:27:02 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.863	152	162330	20.00	ng	0.00
21) Naphthalene-d8	8.151	136	584049	20.00	ng	0.00
39) Acenaphthene-d10	9.910	164	263354	20.00	ng	0.00
64) Phenanthrene-d10	11.404	188	409862	20.00	ng	0.00
76) Chrysene-d12	14.051	240	441068	20.00	ng	0.00
86) Perylene-d12	15.533	264	464182	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.481	112	526403	49.79	ng	-0.01
7) Phenol-d6	6.492	99	631826	45.69	ng	-0.02
23) Nitrobenzene-d5	7.422	82	399984	41.93	ng	-0.01
42) 2,4,6-Tribromophenol	10.704	330	139279	49.28	ng	0.00
45) 2-Fluorobiphenyl	9.227	172	700704	41.57	ng	0.00
79) Terphenyl-d14	12.992	244	624269	29.98	ng	0.00
Target Compounds						Qvalue
50) Dimethylphthalate	9.622	163	95320	5.09	ng	# 97
75) Fluoranthene	12.621	202	69227	2.85	ng	90
78) Pyrene	12.851	202	66326	2.14	ng	# 92
84) Bis(2-ethylhexyl)phtha...	14.033	149	68169	4.56	ng	# 95
88) Benzo(b)fluoranthene	15.098	252	80135m	2.67	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF112320\
Data File : BF122460.D
Acq On : 23 Nov 2020 22:01
Operator : JU/CG
Sample : L4778-07
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
358

Manual Integrations
APPROVED

mohammad
11/25/2020 12:51:40 PM

Quant Time: Nov 24 00:28:34 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF110920.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 20 13:27:02 2020
Response via : Initial Calibration

