

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF112322\
 Data File : BF131358.D
 Acq On : 23 Nov 2022 12:24
 Operator : CG\JU
 Sample : N5641-12
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SOIL-PILE-12

Quant Time: Nov 24 06:30:38 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF112122.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 24 05:53:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.963	152	86609	20.000	ng	0.00
21) Naphthalene-d8	8.251	136	319903	20.000	ng	0.00
39) Acenaphthene-d10	10.016	164	184406	20.000	ng	0.00
64) Phenanthrene-d10	11.516	188	302903	20.000	ng	0.01
76) Chrysene-d12	14.168	240	164315	20.000	ng	0.00
86) Perylene-d12	15.709	264	179460	20.000	ng	0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.575	112	499460	109.755	ng	0.02
7) Phenol-d6	6.575	99	628149	108.187	ng	0.00
23) Nitrobenzene-d5	7.528	82	409973	64.601	ng	0.00
42) 2,4,6-Tribromophenol	10.810	330	175062	112.840	ng	0.00
45) 2-Fluorobiphenyl	9.333	172	772465	60.942	ng	0.00
79) Terphenyl-d14	13.110	244	646733	64.363	ng	0.01
Target Compounds						
84) Bis(2-ethylhexyl)phtha...	14.145	149	2901	4.162	ng	Qvalue # 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF112322\
Data File : BF131358.D
Acq On : 23 Nov 2022 12:24
Operator : CG\JU
Sample : N5641-12
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SOIL-PILE-12

Quant Time: Nov 24 06:30:38 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF112122.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 24 05:53:32 2022
Response via : Initial Calibration

