



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF112322\
 Data File : BF131354.D
 Acq On : 23 Nov 2022 09:26
 Operator : CG\JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF112122.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Nov 24 05:53:32 2022






AutoFind: Scans 1644, 1645, 1646; Background Corrected with Scan 1639

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	41.4	60173	PASS
68	69	0.00	2	1.9	1020	PASS
69	198	0.00	100	36.3	52813	PASS
70	69	0.00	2	0.3	143	PASS
127	198	10	80	47.0	68317	PASS
197	198	0.00	2	0.6	848	PASS
198	198	100	100	100.0	145421	PASS
199	198	5	9	6.5	9501	PASS
275	198	10	60	29.8	43379	PASS
365	198	1	100	3.6	5300	PASS
441	198	0.01	100	13.7	19875	PASS
442	442	50	100	100.0	127011	PASS
443	442	15	24	19.1	24238	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF112322\
Data File : BF131354.D
Acq On : 23 Nov 2022 09:26
Operator : CG\JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Nov 24 07:18:37 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF112122.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 24 05:53:32 2022
Response via : Initial Calibration



TIC: BF131354.D\data.ms

(70) Pentachlorophenol (C)

11.310min (+ 0.012) 32798.69 ng




response 159438

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.20	65.04
264.00	63.00	62.59
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF112322\
Data File : BF131354.D
Acq On : 23 Nov 2022 09:26
Operator : CG\JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Nov 24 07:18:37 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF112122.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 24 05:53:32 2022
Response via : Initial Calibration



TIC: BF131354.D\data.ms

(77) Benzidine

12.863min (+ 0.018) 0.00 ng

response 812496

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	17.30	14.45
183.00	12.00	12.08
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
11/23/2022	BNA_F	<u>BF131354.D</u>
Compound Name	Response	Retention Time
DDT	333470	13.686
DDD	4486	13.416
DDE	191	13.045
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
4677	338147	1.38

Instrument :
BNA_F
ClientSampleId :
DFTPP