

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF112423\
 Data File : BF136210.D
 Acq On : 25 Nov 2023 08:13
 Operator : CG\JU
 Sample : 05418-09
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 STONE-3

Quant Time: Nov 27 02:20:12 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF110723.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 24 23:44:28 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.810	152	136072	20.000	ng	0.00
21) Naphthalene-d8	8.081	136	572724	20.000	ng	0.00
39) Acenaphthene-d10	9.839	164	290028	20.000	ng	0.00
64) Phenanthrene-d10	11.327	188	524443	20.000	ng	-0.01
76) Chrysene-d12	13.980	240	270873	20.000	ng	0.00
86) Perylene-d12	15.462	264	265661	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.451	112	1033684	120.956	ng	0.02
7) Phenol-d6	6.439	99	1114014	105.935	ng	0.00
23) Nitrobenzene-d5	7.369	82	895301	87.233	ng	0.00
42) 2,4,6-Tribromophenol	10.633	330	396653	130.582	ng	0.00
45) 2-Fluorobiphenyl	9.163	172	1843857	93.977	ng	0.00
79) Terphenyl-d14	12.927	244	1909662	100.073	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF112423\
Data File : BF136210.D
Acq On : 25 Nov 2023 08:13
Operator : CG\JU
Sample : 05418-09
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
STONE-3

Quant Time: Nov 27 02:20:12 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF110723.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 24 23:44:28 2023
Response via : Initial Calibration

