

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF112918\
 Data File : BF111092.D
 Acq On : 29 Nov 2018 16:03
 Operator : JU/SJ
 Sample : J6084-01
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 MS-SFVP8283LM

Quant Time: Nov 29 16:37:41 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF112618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 29 10:35:31 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.95	152	61464	20.00	ng	0.02
21) Naphthalene-d8	8.22	136	265599	20.00	ng	0.00
39) Acenaphthene-d10	9.97	164	122540	20.00	ng	0.00
64) Phenanthrene-d10	11.47	188	201193	20.00	ng	0.00
76) Chrysene-d12	14.13	240	132719	20.00	ng	0.00
87) Perylene-d12	15.67	264	136308	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.62	112	31058	8.38	ng	0.06
7) Phenol-d6	6.66	99	17854	3.70	ng	0.09
23) Nitrobenzene-d5	7.52	82	53010	11.42	ng	0.00
42) 2,4,6-Tribromophenol	10.77	330	7762	6.38	ng	0.00
45) 2-Fluorobiphenyl	9.29	172	97907	11.58	ng	0.00
79) Terphenyl-d14	13.05	244	70608	10.36	ng	0.00
Target Compounds						
6) Aniline	6.72	93	15158751	2538.723	ng	# 58
31) Naphthalene	8.25	128	25674	2.062	ng	# 83
33) 4-Chloroaniline	8.31	127	23679	4.468	ng	# 75
37) 2-Methylnaphthalene	8.94	142	37591	4.571	ng	# 98
38) 1-Methylnaphthalene	9.04	142	23821	2.978	ng	# 98
50) Dimethylphthalate	9.68	163	43921	4.887	ng	# 98
77) Benzidine	12.83	184	161827	47.671	ng	# 70
84) Bis(2-ethylhexyl)phthalate	14.06	149	48377	8.296	ng	# 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF112918\
Data File : BF111092.D
Acq On : 29 Nov 2018 16:03
Operator : JU/SJ
Sample : J6084-01
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
MS-SFVP8283LM

Quant Time: Nov 29 16:37:41 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF112618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 29 10:35:31 2018
Response via : Initial Calibration

