


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF112918\
 Data File : BF111086.D
 Acq On : 29 Nov 2018 12:41
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF112618.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Nov 29 10:35:31 2018



AutoFind: Scans 1632, 1633, 1634; Background Corrected with Scan 1627

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	37.9	29721	PASS
68	69	0.00	2	1.5	421	PASS
69	198	0.00	100	35.9	28141	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	48.8	38248	PASS
197	198	0.00	2	0.7	558	PASS
198	198	100	100	100.0	78352	PASS
199	198	5	9	6.3	4934	PASS
275	198	10	60	25.3	19862	PASS
365	198	1	100	2.9	2272	PASS
441	198	0.01	100	15.1	11855	PASS
442	198	50	100	92.3	72338	PASS
443	442	15	24	19.4	14003	PASS