

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF120423\
 Data File : BF136376.D
 Acq On : 04 Dec 2023 12:05
 Operator : CG\JU
 Sample : PB157389BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB157389BL

Quant Time: Dec 04 12:52:10 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF112723.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 28 03:04:20 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.798	152	155490	20.000	ng	-0.01
21) Naphthalene-d8	8.075	136	665770	20.000	ng	-0.01
39) Acenaphthene-d10	9.833	164	356574	20.000	ng	-0.01
64) Phenanthrene-d10	11.327	188	734407	20.000	ng	0.00
76) Chrysene-d12	13.974	240	489540	20.000	ng	-0.02
86) Perylene-d12	15.457	264	342601	20.000	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.451	112	1303075	116.643	ng	0.01
7) Phenol-d6	6.440	99	1641144	117.489	ng	0.00
23) Nitrobenzene-d5	7.363	82	999938	79.367	ng	-0.01
42) 2,4,6-Tribromophenol	10.628	330	563757	129.319	ng	-0.01
45) 2-Fluorobiphenyl	9.157	172	2094700	75.035	ng	-0.01
79) Terphenyl-d14	12.927	244	2923370	68.882	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF120423\
Data File : BF136376.D
Acq On : 04 Dec 2023 12:05
Operator : CG\JU
Sample : PB157389BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB157389BL

Quant Time: Dec 04 12:52:10 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF112723.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 28 03:04:20 2023
Response via : Initial Calibration

