

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF120519\
 Data File : BF118107.D
 Acq On : 5 Dec 2019 15:52
 Operator : JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Quant Time: Dec 06 03:32:58 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF111319.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 03 15:16:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.87	152	150670	20.00	ng	0.00
21) Naphthalene-d8	8.16	136	565317	20.00	ng	0.00
39) Acenaphthene-d10	9.92	164	291245	20.00	ng	-0.01
64) Phenanthrene-d10	11.42	188	526654	20.00	ng	0.00
76) Chrysene-d12	14.07	240	329062	20.00	ng	0.00
87) Perylene-d12	15.56	264	283227	20.00	ng	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
5) 2-Fluorophenol	5.49	112	678608	79.72	ng	0.00
7) Phenol-d6	6.50	99	808644	78.76	ng	0.00
23) Nitrobenzene-d5	7.44	82	765229	80.47	ng	0.00
42) 2,4,6-Tribromophenol	10.72	330	262357	85.09	ng	0.00
45) 2-Fluorobiphenyl	9.24	172	1510549	93.05	ng	0.00
79) Terphenyl-d14	13.00	244	1579181	87.71	ng	-0.01

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) 1,4-Dioxane	2.61	88	148483	37.319	ng	98
3) Pyridine	3.37	79	376281	36.053	ng	99
4) n-Nitrosodimethylamine	3.32	42	149647	32.846	ng	93
6) Aniline	6.53	93	520776	38.385	ng	99
8) 2-Chlorophenol	6.66	128	379005	41.035	ng	99
9) Benzaldehyde	6.42	77	214235	29.898	ng	98
10) Phenol	6.52	94	457128	42.848	ng	99
11) bis(2-Chloroethyl)ether	6.61	93	333647	38.944	ng	99
12) 1,3-Dichlorobenzene	6.82	146	446553	41.067	ng	98
13) 1,4-Dichlorobenzene	6.89	146	451772	41.103	ng	98
14) 1,2-Dichlorobenzene	7.05	146	421226	40.072	ng	99
15) Benzyl Alcohol	7.02	79	317888	40.105	ng	99
16) 2,2'-oxybis(1-Chloropropan	7.15	45	416209	31.554	ng	98
17) 2-Methylphenol	7.13	107	299631	38.308	ng	99
18) Hexachloroethane	7.39	117	165192	40.913	ng	98
19) n-Nitroso-di-n-propylamine	7.29	70	240674	37.999	ng	98
20) 3+4-Methylphenols	7.28	107	375973	38.723	ng	99
22) Acetophenone	7.29	105	516143	40.638	ng	98
24) Nitrobenzene	7.46	77	376180	38.344	ng	99
25) Isophorone	7.70	82	643230	36.904	ng	100
26) 2-Nitrophenol	7.77	139	200306	41.948	ng	100
27) 2,4-Dimethylphenol	7.81	122	271723	36.621	ng	100
28) bis(2-Chloroethoxy)methane	7.90	93	430823	38.951	ng	99
29) 2,4-Dichlorophenol	8.02	162	319221	39.662	ng	100
30) 1,2,4-Trichlorobenzene	8.10	180	379744	40.676	ng	99
31) Naphthalene	8.19	128	1102869	41.847	ng	100
32) Benzoic acid	7.93	122	232799	37.471	ng	98
33) 4-Chloroaniline	8.23	127	467727	39.643	ng	97
34) Hexachlorobutadiene	8.30	225	225603	41.332	ng	98
35) Caprolactam	8.60	113	94970	37.128	ng	94
36) 4-Chloro-3-methylphenol	8.71	107	323318	39.583	ng	99
37) 2-Methylnaphthalene	8.87	142	738628	41.480	ng	99
38) 1-Methylnaphthalene	8.97	142	694886	41.277	ng	99
40) 1,2,4,5-Tetrachlorobenzene	9.05	216	353348	41.777	ng	99


Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF120519\
 Data File : BF118107.D
 Acq On : 5 Dec 2019 15:52
 Operator : JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Quant Time: Dec 06 03:32:58 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF111319.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 03 15:16:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Hexachlorocyclopentadiene	9.03	237	146456	30.899	ng	100
43) 2,4,6-Trichlorophenol	9.15	196	231426	38.201	ng	100
44) 2,4,5-Trichlorophenol	9.20	196	237485	37.756	ng	98
46) 1,1'-Biphenyl	9.34	154	910135	42.121	ng	99
47) 2-Chloronaphthalene	9.37	162	721343	40.580	ng	98
48) 2-Nitroaniline	9.46	65	197212	36.748	ng	97
49) Acenaphthylene	9.79	152	1066245	41.142	ng	99
50) Dimethylphthalate	9.64	163	819117	40.107	ng	99
51) 2,6-Dinitrotoluene	9.70	165	177536	39.774	ng	94
52) Acenaphthene	9.96	154	637684	38.411	ng	98
53) 3-Nitroaniline	9.87	138	207353	38.823	ng	100
54) 2,4-Dinitrophenol	9.99	184	73181	36.252	ng	98
55) Dibenzofuran	10.13	168	946867	41.187	ng	99
56) 4-Nitrophenol	10.03	139	138017	34.619	ng	98
57) 2,4-Dinitrotoluene	10.11	165	233861	41.187	ng	# 86
58) Fluorene	10.47	166	742446	41.675	ng	100
59) 2,3,4,6-Tetrachlorophenol	10.25	232	192707	37.287	ng	99
60) Diethylphthalate	10.34	149	809223	40.641	ng	98
61) 4-Chlorophenyl-phenylether	10.46	204	378401	41.854	ng	97
62) 4-Nitroaniline	10.49	138	209747	39.222	ng	98
63) Azobenzene	10.63	77	700031	38.644	ng	96
65) 4,6-Dinitro-2-methylphenol	10.52	198	104743	42.603	ng	83
66) n-Nitrosodiphenylamine	10.58	169	649890	42.401	ng	99
67) 4-Bromophenyl-phenylether	10.96	248	238608	41.990	ng	95
68) Hexachlorobenzene	11.03	284	273223	41.234	ng	95
69) Atrazine	11.11	200	179049	35.512	ng	98
70) Pentachlorophenol	11.22	266	127700	32.987	ng	99
71) Phenanthrene	11.44	178	1090685	41.191	ng	100
72) Anthracene	11.49	178	1082981	41.252	ng	100
73) Carbazole	11.65	167	961025	41.891	ng	99
74) Di-n-butylphthalate	11.97	149	1214732	42.527	ng	100
75) Fluoranthene	12.63	202	1061236	43.814	ng	100
77) Benzidine	12.75	184	396462	36.782	ng	98
78) Pyrene	12.86	202	1067300	40.134	ng	100
80) Butylbenzylphthalate	13.48	149	470359	41.181	ng	98
81) Benzo(a)anthracene	14.06	228	873366	40.164	ng	99
82) 3,3'-Dichlorobenzidine	14.02	252	293300	38.561	ng	97
83) Chrysene	14.09	228	839630	39.848	ng	99
84) Bis(2-ethylhexyl)phthalate	14.04	149	593665	42.872	ng	98
85) Di-n-octyl phthalate	14.66	149	874789	43.002	ng	100
86) Indeno(1,2,3-cd)pyrene	17.07	276	692831	38.634	ng	99
88) Benzo(b)fluoranthene	15.12	252	700593	38.073	ng	99
89) Benzo(k)fluoranthene	15.15	252	707859	40.826	ng	99
90) Benzo(a)pyrene	15.50	252	640409	39.578	ng	98
91) Dibenzo(a,h)anthracene	17.09	278	566810	40.698	ng	99
92) Benzo(g,h,i)perylene	17.53	276	560909	40.435	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed




#1




1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 6.87 min Scan# 814
 Delta R.T. -0.01 min
 Lab File: BF118107.D
 Acq: 5 Dec 2019 15:52

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Tot Ion:152 Resp: 150670
 Ion Ratio Lower Upper
 152 100
 150 155.3 121.4 182.2
 115 59.6 48.8 73.2




Abundance Ion 152.00 (151.70 to 152.70): E
 Ion 150.00 (149.70 to 150.70): E
 Ion 115.00 (114.70 to 115.70): E





#2


1,4-Dioxane
 Concen: 37.319 ng
 RT: 2.61 min Scan# 89
 Delta R.T. -0.01 min
 Lab File: BF118107.D
 Acq: 5 Dec 2019 15:52

Tgt Ion: 88 Resp: 148483
 Ion Ratio Lower Upper
 88 100
 58 64.9 52.6 78.8
 43 25.6 21.8 32.6



Abundance Ion 88.00 (87.70 to 88.70): BF1
 Ion 58.00 (57.70 to 58.70): BF1
 Ion 43.00 (42.70 to 43.70): BF1





#3

Pvridine

Concen: 36.053 ng

RT: 3.37 min Scan# 218

Delta R.T. -0.01 min

Lab File: BF118107.D

Acq: 5 Dec 2019 15:52

Instrument :

BNA_F

ClientSampleId :

DFTPP


Tot Ion: 79 Resp: 376281

Ion Ratio Lower Upper




79 100

52 63.7 50.6 75.8

51 32.9 26.8 40.2



Abundance Ion 79.00 (78.70 to 79.70): BF1
Ion 52.00 (51.70 to 52.70): BF1
Ion 51.00 (50.70 to 51.70): BF1



#4

n-Nitrosodimethylamine

Concen: 32.846 ng

RT: 3.32 min Scan# 210

Delta R.T. -0.01 min

Lab File: BF118107.D

Acq: 5 Dec 2019 15:52


Tgt Ion: 42 Resp: 149647

Ion Ratio Lower Upper



42 100


74 137.0 102.8 154.2

44 6.5 4.7 7.1



Abundance Ion 42.00 (41.70 to 42.70): BF1
Ion 74.00 (73.70 to 74.70): BF1
Ion 44.00 (43.70 to 44.70): BF1





#5

2-Fluorophenol

Concen: 79.724 ng

RT: 5.49 min Scan# 578

Delta R.T. -0.00 min

Lab File: BF118107.D

Acq: 5 Dec 2019 15:52

Instrument :

BNA_F

ClientSampleId :

DFTPP


Tot Ion:112 Resp: 678608

Ion Ratio Lower Upper

112 100

64 50.3 45.1 67.7




63 27.2 24.6 37.0



Abundance Ion 112.00 (111.70 to 112.70): BF1

Ion 64.00 (63.70 to 64.70): BF1

Ion 63.00 (62.70 to 63.70): BF1



#6

Aniline

Concen: 38.385 ng

RT: 6.53 min Scan# 756

Delta R.T. -0.01 min

Lab File: BF118107.D

Acq: 5 Dec 2019 15:52


Tgt Ion: 93 Resp: 520776

Ion Ratio Lower Upper

93 100

66 38.2 31.0 46.6



65 19.4 15.8 23.8




Abundance Ion 93.00 (92.70 to 93.70): BF1

Ion 66.00 (65.70 to 66.70): BF1

Ion 65.00 (64.70 to 65.70): BF1





#7

Phenol-d6

Concen: 78.763 ng

RT: 6.50 min Scan# 751

Delta R.T. -0.01 min

Lab File: BF118107.D

Acq: 5 Dec 2019 15:52

Instrument :

BNA_F

ClientSampleId :

DFTPP


Tot Ion: 99 Resp: 808644

Ion Ratio Lower Upper




99 100

42 16.9 14.0 21.0

71 34.9 29.2 43.8



Abundance Ion 99.00 (98.70 to 99.70): BF1
Ion 42.00 (41.70 to 42.70): BF1
Ion 71.00 (70.70 to 71.70): BF1



#8

2-Chlorophenol

Concen: 41.035 ng

RT: 6.66 min Scan# 777

Delta R.T. -0.01 min

Lab File: BF118107.D

Acq: 5 Dec 2019 15:52


Tgt Ion:128 Resp: 379005

Ion Ratio Lower Upper



128 100


130 31.6 12.4 52.4

64 43.2 23.2 63.2



Abundance Ion 128.00 (127.70 to 128.70): E
Ion 130.00 (129.70 to 130.70): E
Ion 64.00 (63.70 to 64.70): BF1





#9

Benzaldehyde

Concen: 29.898 ng

RT: 6.42 min Scan# 737

Delta R.T. -0.01 min

Lab File: BF118107.D

Acq: 5 Dec 2019 15:52

Instrument :

BNA_F

ClientSampleId :

DFTPP





Tot Ion: 77 Resp: 214235

Ion Ratio Lower Upper

77 100

105 101.4 83.3 123.3

106 99.6 80.9 120.9



#10

Phenol

Concen: 42.848 ng

RT: 6.52 min Scan# 753

Delta R.T. -0.01 min

Lab File: BF118107.D

Acq: 5 Dec 2019 15:52




Tgt Ion: 94 Resp: 457128


Ion Ratio Lower Upper

94 100

65 27.5 6.9 46.9

66 38.7 18.6 58.6









#11
bis(2-Chloroethyl)ether
Concen: 38.944 ng
RT: 6.61 min Scan# 769
Delta R.T. -0.01 min
Lab File: BF118107.D
Acq: 5 Dec 2019 15:52




Instrument :
BNA_F
ClientSampled :
DFTPP


Tot Ion: 93 Resp: 333647
Ion Ratio Lower Upper
93 100
63 67.4 46.8 86.8
95 32.3 12.0 52.0



#12
1,3-Dichlorobenzene
Concen: 41.067 ng
RT: 6.82 min Scan# 804
Delta R.T. -0.01 min
Lab File: BF118107.D
Acq: 5 Dec 2019 15:52

Tgt Ion: 146 Resp: 446553
Ion Ratio Lower Upper
146 100
148 64.7 52.7 79.1
75 29.4 24.5 36.7









#13
 1,4-Dichlorobenzene
 Concen: 41.103 ng
 RT: 6.89 min Scan# 817
 Delta R.T. -0.01 min
 Lab File: BF118107.D
 Acq: 5 Dec 2019 15:52

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Tot Ion:146 Resp: 451772
 Ion Ratio Lower Upper
 146 100
 148 64.5 50.8 76.2
 111 41.2 34.6 51.8




Abundance Ion 146.00 (145.70 to 146.70): E
 Ion 148.00 (147.70 to 148.70): E
 Ion 111.00 (110.70 to 111.70): E






#14
 1,2-Dichlorobenzene
 Concen: 40.072 ng
 RT: 7.05 min Scan# 843
 Delta R.T. -0.01 min
 Lab File: BF118107.D
 Acq: 5 Dec 2019 15:52

Tgt Ion:146 Resp: 421226
 Ion Ratio Lower Upper
 146 100
 148 65.5 51.6 77.4
 111 44.7 35.8 53.8



Abundance Ion 146.00 (145.70 to 146.70): E
 Ion 148.00 (147.70 to 148.70): E
 Ion 111.00 (110.70 to 111.70): E





#15

Benzyl Alcohol

Concen: 40.105 ng

RT: 7.02 min Scan# 838

Delta R.T. -0.01 min

Lab File: BF118107.D

Acq: 5 Dec 2019 15:52

Instrument :

BNA_F

ClientSampleId :

DFTPP


Tot Ion: 79 Resp: 317888

Ion Ratio Lower Upper

79 100

108 80.1 63.1 94.7




77 62.7 50.6 76.0



Abundance Ion 79.00 (78.70 to 79.70): BF1

Ion 108.00 (107.70 to 108.70): BF1

Ion 77.00 (76.70 to 77.70): BF1



#16

2,2'-oxybis(1-chloropropane)

Concen: 31.554 ng

RT: 7.15 min Scan# 861

Delta R.T. -0.01 min

Lab File: BF118107.D

Acq: 5 Dec 2019 15:52


Tgt Ion: 45 Resp: 416209

Ion Ratio Lower Upper

45 100

77 17.7 0.0 36.8



79 13.9 0.0 33.5




Abundance Ion 45.00 (44.70 to 45.70): BF1

Ion 77.00 (76.70 to 77.70): BF1

Ion 79.00 (78.70 to 79.70): BF1









#17
2-Methylphenol
Concen: 38.308 ng
RT: 7.13 min Scan# 857
Delta R.T. -0.01 min
Lab File: BF118107.D
Acq: 5 Dec 2019 15:52




Instrument :
BNA_F
ClientSampleId :
DFTPP


Tot Ion:107 Resp: 299631
Ion Ratio Lower Upper
107 100
108 111.0 89.8 134.8
77 44.0 35.0 52.6
79 43.8 35.0 52.6



#18
Hexachloroethane
Concen: 40.913 ng
RT: 7.39 min Scan# 901
Delta R.T. -0.01 min
Lab File: BF118107.D
Acq: 5 Dec 2019 15:52

Tgt Ion:117 Resp: 165192
Ion Ratio Lower Upper
117 100
119 96.6 76.4 114.6
201 93.3 71.7 107.5









#19
n-Nitroso-di-n-propylamine
Concen: 37.999 ng
RT: 7.29 min Scan# 884
Delta R.T. -0.01 min
Lab File: BF118107.D
Acq: 5 Dec 2019 15:52

Instrument :
BNA_F
ClientSampleId :
DFTPP

Tot Ion:	70	Resp:	240674
Ion	Ratio	Lower	Upper
70	100		
42	55.7	46.0	69.0
101	9.9	8.8	13.2
130	23.6	19.8	29.6




Abundance Ion 70.00 (69.70 to 70.70): BF1
Ion 42.00 (41.70 to 42.70): BF1
Ion 101.00 (100.70 to 101.70): E
Ion 130.00 (129.70 to 130.70): E






#20
3+4-Methylphenols
Concen: 38.723 ng
RT: 7.28 min Scan# 883
Delta R.T. -0.01 min
Lab File: BF118107.D
Acq: 5 Dec 2019 15:52

Tgt Ion:	107	Resp:	375973
Ion	Ratio	Lower	Upper
107	100		
108	90.0	71.9	111.9
77	111.6	91.6	131.6
79	29.9	11.3	51.3



Abundance Ion 107.00 (106.70 to 107.70): E
Ion 108.00 (107.70 to 108.70): E
Ion 77.00 (76.70 to 77.70): BF1
Ion 79.00 (78.70 to 79.70): BF1





#21

Naphthalene-d8

Concen: 20.000 ng

RT: 8.16 min Scan# 1033

Delta R.T. -0.01 min

Lab File: BF118107.D

Acq: 5 Dec 2019 15:52

Instrument :

BNA_F

ClientSampleId :

DFTPP

Tot Ion:136 Resp: 565317



Ion Ratio Lower Upper

136 100

137 11.0 9.1 13.7

54 6.0 4.9 7.3

68 4.5 3.6 5.4





Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BF1

Ion 68.00 (67.70 to 68.70): BF1



#22

Acetophenone

Concen: 40.638 ng

RT: 7.29 min Scan# 884

Delta R.T. -0.01 min

Lab File: BF118107.D

Acq: 5 Dec 2019 15:52

Tgt Ion:105 Resp: 516143



Ion Ratio Lower Upper

105 100

71 8.1 5.9 8.9

51 26.3 20.2 30.2

120 22.7 17.9 26.9





Abundance Ion 105.00 (104.70 to 105.70): E

Ion 71.00 (70.70 to 71.70): BF1

Ion 51.00 (50.70 to 51.70): BF1

Ion 120.00 (119.70 to 120.70): E









#23
 Nitrobenzene-d5
 Concen: 80.468 ng
 RT: 7.44 min Scan# 910
 Delta R.T. -0.01 min
 Lab File: BF118107.D
 Acq: 5 Dec 2019 15:52




Instrument :
 BNA_F
 ClientSampleID :
 DFTPP


Tot Ion	Ratio	Lower	Upper
82	100		
128	46.9	38.2	57.4
54	46.5	39.0	58.4



#24
 Nitrobenzene
 Concen: 38.344 ng
 RT: 7.46 min Scan# 913
 Delta R.T. -0.01 min
 Lab File: BF118107.D
 Acq: 5 Dec 2019 15:52

Tgt Ion	Ratio	Lower	Upper
77	100		
123	48.6	38.2	57.2
65	13.4	11.1	16.7





#25

Isophorone

Concen: 36.904 ng

RT: 7.70 min Scan# 954

Delta R.T. -0.01 min

Lab File: BF118107.D

Acq: 5 Dec 2019 15:52

Instrument :

BNA_F

ClientSampleId :

DFTPP


Tot Ion: 82 Resp: 643230

Ion Ratio Lower Upper

82 100

95 7.7 6.2 9.2




138 20.2 15.9 23.9



Abundance Ion 82.00 (81.70 to 82.70): BF1

Ion 95.00 (94.70 to 95.70): BF1

Ion 138.00 (137.70 to 138.70): BF1



#26

2-Nitrophenol

Concen: 41.948 ng

RT: 7.77 min Scan# 967

Delta R.T. -0.01 min

Lab File: BF118107.D

Acq: 5 Dec 2019 15:52


Tgt Ion: 139 Resp: 200306

Ion Ratio Lower Upper

139 100

109 30.1 23.8 35.6



65 45.1 36.2 54.4




Abundance Ion 139.00 (138.70 to 139.70): BF1

Ion 109.00 (108.70 to 109.70): BF1

Ion 65.00 (64.70 to 65.70): BF1









#27
 2,4-Dimethylphenol
 Concen: 36.621 ng
 RT: 7.81 min Scan# 973
 Delta R.T. -0.01 min
 Lab File: BF118107.D
 Acq: 5 Dec 2019 15:52




Instrument :
 BNA_F
 ClientSampleId :
 DFTPP


Tot Ion	Ratio	Lower	Upper
122	100		
107	115.0	92.6	138.8
121	58.8	46.9	70.3



#28
 bis(2-Chloroethoxy)methane
 Concen: 38.951 ng
 RT: 7.90 min Scan# 989
 Delta R.T. -0.01 min
 Lab File: BF118107.D
 Acq: 5 Dec 2019 15:52

Tgt Ion	Ratio	Lower	Upper
93	100		
95	32.4	26.4	39.6
123	12.3	10.2	15.4









#29
2,4-Dichlorophenol
Concen: 39.662 ng
RT: 8.02 min Scan# 1008
Delta R.T. -0.01 min
Lab File: BF118107.D
Acq: 5 Dec 2019 15:52




Instrument :
BNA_F
ClientSampleId :
DFTPP


Tot Ion:162 Resp: 319221
Ion Ratio Lower Upper
162 100
164 62.8 42.9 82.9
98 37.1 16.7 56.7



#30
1,2,4-Trichlorobenzene
Concen: 40.676 ng
RT: 8.10 min Scan# 1023
Delta R.T. -0.01 min
Lab File: BF118107.D
Acq: 5 Dec 2019 15:52

Tgt Ion:180 Resp: 379744
Ion Ratio Lower Upper
180 100
182 94.9 76.6 114.8
145 29.8 25.1 37.7





#31

Naphthalene

Concen: 41.847 ng

RT: 8.19 min Scan# 1037

Delta R.T. -0.01 min

Lab File: BF118107.D

Acq: 5 Dec 2019 15:52

Instrument :

BNA_F

ClientSampleId :

DFTPP


Tot Ion:128 Resp: 1102869

Ion Ratio Lower Upper

128 100

129 11.0 8.8 13.2




127 13.3 10.6 16.0



Abundance Ion 128.00 (127.70 to 128.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 127.00 (126.70 to 127.70): E



#32

Benzoic acid

Concen: 37.471 ng

RT: 7.93 min Scan# 994

Delta R.T. -0.01 min

Lab File: BF118107.D

Acq: 5 Dec 2019 15:52


Tgt Ion:122 Resp: 232799

Ion Ratio Lower Upper

122 100

105 122.5 101.4 141.4



77 90.6 66.4 106.4




Abundance Ion 122.00 (121.70 to 122.70): E

Ion 105.00 (104.70 to 105.70): E

Ion 77.00 (76.70 to 77.70): BF1









#33
4-Chloroaniline
Concen: 39.643 ng
RT: 8.23 min Scan# 1045
Delta R.T. -0.00 min
Lab File: BF118107.D
Acq: 5 Dec 2019 15:52




Instrument :
BNA_F
ClientSampleId :
DFTPP


Tot Ion:	127	Resp:	467727
Ion	Ratio	Lower	Upper
127	100		
129	32.3	25.4	38.0
65	25.6	23.4	35.0
92	18.2	15.1	22.7



#34
Hexachlorobutadiene
Concen: 41.332 ng
RT: 8.30 min Scan# 1056
Delta R.T. -0.01 min
Lab File: BF118107.D
Acq: 5 Dec 2019 15:52

Tgt Ion:	225	Resp:	225603
Ion	Ratio	Lower	Upper
225	100		
223	62.0	51.4	77.2
227	64.9	52.6	78.8









#35
 Caprolactam
 Concen: 37.128 ng
 RT: 8.60 min Scan# 1107
 Delta R.T. -0.01 min
 Lab File: BF118107.D
 Acq: 5 Dec 2019 15:52

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Tot Ion:113 Resp: 94970
 Ion Ratio Lower Upper
 113 100
 55 133.3 125.5 165.5
 56 106.1 88.0 128.0




Abundance Ion 113.00 (112.70 to 113.70): E
 Ion 55.00 (54.70 to 55.70): BF1
 Ion 56.00 (55.70 to 56.70): BF1






#36
 4-Chloro-3-methylphenol
 Concen: 39.583 ng
 RT: 8.71 min Scan# 1126
 Delta R.T. -0.01 min
 Lab File: BF118107.D
 Acq: 5 Dec 2019 15:52

Tgt Ion:107 Resp: 323318
 Ion Ratio Lower Upper
 107 100
 144 25.9 20.6 30.8
 142 78.7 62.5 93.7



Abundance Ion 107.00 (106.70 to 107.70): E
 Ion 144.00 (143.70 to 144.70): E
 Ion 142.00 (141.70 to 142.70): E





#37

2-Methylnaphthalene

Concen: 41.480 ng

RT: 8.87 min Scan# 1154

Delta R.T. -0.01 min

Lab File: BF118107.D

Acq: 5 Dec 2019 15:52

Instrument :

BNA_F

ClientSampleId :

DFTPP


Tot Ion:142 Resp: 738628

Ion Ratio Lower Upper

142 100

141 88.2 70.2 105.4




115 35.9 29.2 43.8



Abundance Ion 142.00 (141.70 to 142.70): E

Ion 141.00 (140.70 to 141.70): E

Ion 115.00 (114.70 to 115.70): E



#38

1-Methylnaphthalene

Concen: 41.277 ng

RT: 8.97 min Scan# 1171

Delta R.T. -0.01 min

Lab File: BF118107.D

Acq: 5 Dec 2019 15:52


Tgt Ion:142 Resp: 694886

Ion Ratio Lower Upper

142 100

141 91.3 73.6 110.4



116 3.9 3.1 4.7




Abundance Ion 142.00 (141.70 to 142.70): E

Ion 141.00 (140.70 to 141.70): E

Ion 116.00 (115.70 to 116.70): E









#39
 Acenaphthene-d10
 Concen: 20.000 ng
 RT: 9.92 min Scan# 1332
 Delta R.T. -0.01 min
 Lab File: BF118107.D
 Acq: 5 Dec 2019 15:52




Instrument :
 BNA_F
 ClientSampleId :
 DFTPP


Tot Ion	Ratio	Lower	Upper
164	100		
162	100.7	81.8	122.8
160	44.3	35.9	53.9



#40
 1,2,4,5-Tetrachlorobenzene
 Concen: 41.777 ng
 RT: 9.05 min Scan# 1183
 Delta R.T. -0.01 min
 Lab File: BF118107.D
 Acq: 5 Dec 2019 15:52

Tgt Ion	Ratio	Lower	Upper
216	100		
214	78.4	63.0	94.4
179	22.7	18.5	27.7
108	18.4	15.4	23.2





#41

Hexachlorocyclopentadiene

Concen: 30.899 ng

RT: 9.03 min Scan# 1180

Delta R.T. -0.01 min

Lab File: BF118107.D

Acq: 5 Dec 2019 15:52

Instrument :

BNA_F

ClientSampleId :

DFTPP


Tot Ion:237 Resp: 146456

Ion Ratio Lower Upper




237 100

235 63.6 43.7 83.7

272 13.0 0.0 32.7



Abundance Ion 236.80 (236.50 to 237.50): E
Ion 234.90 (234.60 to 235.60): E
Ion 271.75 (271.45 to 272.45): E



#42

2,4,6-Tribromophenol

Concen: 85.087 ng

RT: 10.72 min Scan# 1467

Delta R.T. -0.01 min

Lab File: BF118107.D

Acq: 5 Dec 2019 15:52


Tgt Ion:330 Resp: 262357

Ion Ratio Lower Upper



330 100


332 96.2 78.3 117.5

141 32.5 26.4 39.6



Abundance Ion 329.80 (329.50 to 330.50): E
Ion 331.80 (331.50 to 332.50): E
Ion 141.00 (140.70 to 141.70): E









#43
 2,4,6-Trichlorophenol
 Concen: 38.201 ng
 RT: 9.15 min Scan# 1201
 Delta R.T. -0.01 min
 Lab File: BF118107.D
 Acq: 5 Dec 2019 15:52

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Tot Ion:196 Resp: 231426
 Ion Ratio Lower Upper
 196 100
 198 95.7 76.9 115.3
 200 30.0 24.2 36.4




Abundance Ion 196.00 (195.70 to 196.70): E
 Ion 198.00 (197.70 to 198.70): E
 Ion 200.00 (199.70 to 200.70): E






#44
 2,4,5-Trichlorophenol
 Concen: 37.756 ng
 RT: 9.20 min Scan# 1209
 Delta R.T. -0.01 min
 Lab File: BF118107.D
 Acq: 5 Dec 2019 15:52

Tgt Ion:196 Resp: 237485
 Ion Ratio Lower Upper
 196 100
 198 99.5 78.1 117.1
 97 44.4 37.4 56.2
 132 27.5 22.7 34.1



Abundance Ion 196.00 (195.70 to 196.70): E
 Ion 198.00 (197.70 to 198.70): E
 Ion 97.00 (96.70 to 97.70): BF1
 Ion 132.00 (131.70 to 132.70): E









#45
2-Fluorobiphenyl
Concen: 93.048 ng
RT: 9.24 min Scan# 1216
Delta R.T. -0.01 min
Lab File: BF118107.D
Acq: 5 Dec 2019 15:52




Instrument :
BNA_F
ClientSampleId :
DFTPP


Tot Ion:172 Resp: 1510549
Ion Ratio Lower Upper
172 100
171 36.3 29.2 43.8
170 24.2 19.3 28.9



#46
1,1'-Biphenyl
Concen: 42.121 ng
RT: 9.34 min Scan# 1233
Delta R.T. -0.01 min
Lab File: BF118107.D
Acq: 5 Dec 2019 15:52

Tgt Ion:154 Resp: 910135
Ion Ratio Lower Upper
154 100
153 41.6 21.1 61.1
76 12.3 0.0 30.8









#47
2-Chloronaphthalene
Concen: 40.580 ng
RT: 9.37 min Scan# 1238
Delta R.T. -0.01 min
Lab File: BF118107.D
Acq: 5 Dec 2019 15:52

Instrument :
BNA_F
ClientSampleId :
DFTPP

Tot Ion:162 Resp: 721343
Ion Ratio Lower Upper
162 100
127 41.5 34.6 52.0
164 32.9 26.0 39.0




Abundance Ion 162.00 (161.70 to 162.70): E
Ion 127.00 (126.70 to 127.70): E
Ion 164.00 (163.70 to 164.70): E






#48
2-Nitroaniline
Concen: 36.748 ng
RT: 9.46 min Scan# 1254
Delta R.T. -0.01 min
Lab File: BF118107.D
Acq: 5 Dec 2019 15:52

Tgt Ion: 65 Resp: 197212
Ion Ratio Lower Upper
65 100
92 75.6 57.9 86.9
138 122.2 96.2 144.2



Abundance Ion 65.00 (64.70 to 65.70): BF1
Ion 92.00 (91.70 to 92.70): BF1
Ion 138.00 (137.70 to 138.70): E





#49

Acenaphthylene

Concen: 41.142 ng

RT: 9.79 min Scan# 1309

Delta R.T. -0.01 min

Lab File: BF118107.D

Acq: 5 Dec 2019 15:52

Instrument :

BNA_F

ClientSampleId :

DFTPP


Tot Ion:152 Resp: 1066245

Ion Ratio Lower Upper



152 100

151 20.5 16.7 25.1

153 13.1 11.1 16.7



Abundance Ion 152.00 (151.70 to 152.70): E
Ion 151.00 (150.70 to 151.70): E
Ion 153.00 (152.70 to 153.70): E



#50

Dimethylphthalate

Concen: 40.107 ng

RT: 9.64 min Scan# 1284

Delta R.T. -0.01 min

Lab File: BF118107.D

Acq: 5 Dec 2019 15:52


Tgt Ion:163 Resp: 819117

Ion Ratio Lower Upper



163 100


194 3.9 3.1 4.7

164 10.6 7.9 11.9



Abundance Ion 163.00 (162.70 to 163.70): E
Ion 194.00 (193.70 to 194.70): E
Ion 164.00 (163.70 to 164.70): E









#51
 2,6-Dinitrotoluene
 Concen: 39.774 ng
 RT: 9.70 min Scan# 1295
 Delta R.T. -0.01 min
 Lab File: BF118107.D
 Acq: 5 Dec 2019 15:52

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Tot Ion:165 Resp: 177536
 Ion Ratio Lower Upper
 165 100
 63 50.2 43.9 65.9
 89 53.3 45.5 68.3




Abundance Ion 165.00 (164.70 to 165.70): E
 Ion 63.00 (62.70 to 63.70): BF1
 Ion 89.00 (88.70 to 89.70): BF1






#52
 Acenaphthene
 Concen: 38.411 ng
 RT: 9.96 min Scan# 1338
 Delta R.T. -0.01 min
 Lab File: BF118107.D
 Acq: 5 Dec 2019 15:52

Tgt Ion:154 Resp: 637684
 Ion Ratio Lower Upper
 154 100
 153 112.2 91.2 136.8
 152 52.9 43.6 65.4



Abundance Ion 154.00 (153.70 to 154.70): E
 Ion 153.00 (152.70 to 153.70): E
 Ion 152.00 (151.70 to 152.70): E









#53
3-Nitroaniline
Concen: 38.823 ng
RT: 9.87 min Scan# 1324
Delta R.T. -0.01 min
Lab File: BF118107.D
Acq: 5 Dec 2019 15:52

Instrument :
BNA_F
ClientSampleId :
DFTPP

Tot Ion:138 Resp: 207353
Ion Ratio Lower Upper
138 100
108 10.4 8.7 13.1
92 108.8 87.4 131.0




Abundance Ion 138.00 (137.70 to 138.70): E
Ion 108.00 (107.70 to 108.70): E
Ion 92.00 (91.70 to 92.70): BF1






#54
2,4-Dinitrophenol
Concen: 36.252 ng
RT: 9.99 min Scan# 1343
Delta R.T. -0.01 min
Lab File: BF118107.D
Acq: 5 Dec 2019 15:52

Tgt Ion:184 Resp: 73181
Ion Ratio Lower Upper
184 100
63 53.1 45.3 67.9
154 64.4 51.6 77.4



Abundance Ion 184.00 (183.70 to 184.70): E
Ion 63.00 (62.70 to 63.70): BF1
Ion 154.00 (153.70 to 154.70): E





#55

Dibenzofuran

Concen: 41.187 ng

RT: 10.13 min Scan# 1367

Delta R.T. -0.01 min

Lab File: BF118107.D

Acq: 5 Dec 2019 15:52

Instrument :

BNA_F

ClientSampled :

DFTPP


Tot Ion:168 Resp: 946867

Ion Ratio Lower Upper

168 100

139 41.3 32.5 48.7




169 14.1 10.8 16.2



Abundance Ion 168.00 (167.70 to 168.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 169.00 (168.70 to 169.70): E



#56

4-Nitrophenol

Concen: 34.619 ng

RT: 10.03 min Scan# 1351

Delta R.T. -0.01 min

Lab File: BF118107.D

Acq: 5 Dec 2019 15:52


Tgt Ion:139 Resp: 138017

Ion Ratio Lower Upper

139 100

109 73.0 55.2 95.2



65 90.1 68.7 108.7




Abundance Ion 139.00 (138.70 to 139.70): E

Ion 109.00 (108.70 to 109.70): E

Ion 65.00 (64.70 to 65.70): BF1









#57
2,4-Dinitrotoluene
Concen: 41.187 ng
RT: 10.11 min Scan# 1364
Delta R.T. -0.01 min
Lab File: BF118107.D
Acq: 5 Dec 2019 15:52




Instrument :
BNA_F
ClientSampleId :
DFTPP


Tot Ion	Ratio	Lower	Upper
165	100		
63	41.4	26.5	39.7#
89	70.3	47.2	70.8
182	3.0	2.3	3.5



#58
Fluorene
Concen: 41.675 ng
RT: 10.47 min Scan# 1426
Delta R.T. -0.01 min
Lab File: BF118107.D
Acq: 5 Dec 2019 15:52

Tgt Ion	Ratio	Lower	Upper
166	100		
165	96.8	77.4	116.2
167	13.4	10.7	16.1





#59

2,3,4,6-Tetrachlorophenol

Concen: 37.287 ng

RT: 10.25 min Scan# 1387

Delta R.T. -0.01 min

Lab File: BF118107.D

Acq: 5 Dec 2019 15:52

Instrument :

BNA_F

ClientSampleId :

DFTPP

Tot Ion:232 Resp: 192707



Ion Ratio Lower Upper

232 100

131 48.3 38.8 58.2

130 2.3 1.9 2.9

166 31.7 25.9 38.9





Abundance Ion 232.00 (231.70 to 232.70): E

Ion 131.00 (130.70 to 131.70): E

Ion 130.00 (129.70 to 130.70): E

Ion 166.00 (165.70 to 166.70): E



#60

Diethylphthalate

Concen: 40.641 ng

RT: 10.34 min Scan# 1403

Delta R.T. -0.01 min

Lab File: BF118107.D

Acq: 5 Dec 2019 15:52



Tgt Ion:149 Resp: 809223

Ion Ratio Lower Upper

149 100

177 21.0 17.6 26.4


150 12.5 9.8 14.8




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 177.00 (176.70 to 177.70): E

Ion 150.00 (149.70 to 150.70): E









#61
4-Chlorophenyl-phenylether
Concen: 41.854 ng
RT: 10.46 min Scan# 1424
Delta R.T. -0.01 min
Lab File: BF118107.D
Acq: 5 Dec 2019 15:52

Instrument :
BNA_F
ClientSampleId :
DFTPP

Tot Ion:204 Resp: 378401
Ion Ratio Lower Upper
204 100
206 32.8 26.0 39.0
141 66.0 55.1 82.7




Abundance Ion 204.00 (203.70 to 204.70): E
Ion 206.00 (205.70 to 206.70): E
Ion 141.00 (140.70 to 141.70): E






#62
4-Nitroaniline
Concen: 39.222 ng
RT: 10.49 min Scan# 1429
Delta R.T. -0.01 min
Lab File: BF118107.D
Acq: 5 Dec 2019 15:52

Tgt Ion:138 Resp: 209747
Ion Ratio Lower Upper
138 100
92 46.8 27.5 67.5
108 76.2 57.9 97.9



Abundance Ion 138.00 (137.70 to 138.70): E
Ion 92.00 (91.70 to 92.70): BF1
Ion 108.00 (107.70 to 108.70): E





#63

Azobenzene

Concen: 38.644 ng

RT: 10.63 min Scan# 1452

Delta R.T. -0.01 min

Lab File: BF118107.D

Acq: 5 Dec 2019 15:52

Instrument :

BNA_F

ClientSampleId :

DFTPP

Tot Ion: 77 Resp: 700031



Ion Ratio Lower Upper

77 100

182 31.0 8.2 48.2

105 23.4 2.1 42.1

51 26.6 8.7 48.7





Abundance Ion 77.00 (76.70 to 77.70): BF1

Ion 182.00 (181.70 to 182.70): BF1

Ion 105.00 (104.70 to 105.70): BF1

Ion 51.00 (50.70 to 51.70): BF1



#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 11.42 min Scan# 1586

Delta R.T. -0.01 min

Lab File: BF118107.D

Acq: 5 Dec 2019 15:52



Tgt Ion: 188 Resp: 526654

Ion Ratio Lower Upper

188 100

94 8.1 7.0 10.6


80 8.7 7.2 10.8




Abundance Ion 188.00 (187.70 to 188.70): BF1

Ion 94.00 (93.70 to 94.70): BF1

Ion 80.00 (79.70 to 80.70): BF1









#65
4,6-Dinitro-2-methylphenol
Concen: 42.603 ng
RT: 10.52 min Scan# 1434
Delta R.T. -0.01 min
Lab File: BF118107.D
Acq: 5 Dec 2019 15:52

Instrument :
BNA_F
ClientSampleId :
DFTPP

Tot Ion:198 Resp: 104743
Ion Ratio Lower Upper
198 100
51 42.3 12.6 52.6
105 47.0 17.0 57.0




Abundance Ion 198.00 (197.70 to 198.70): E
Ion 51.00 (50.70 to 51.70): BF1
Ion 105.00 (104.70 to 105.70): E






#66
n-Nitrosodiphenylamine
Concen: 42.401 ng
RT: 10.58 min Scan# 1444
Delta R.T. -0.01 min
Lab File: BF118107.D
Acq: 5 Dec 2019 15:52

Tgt Ion:169 Resp: 649890
Ion Ratio Lower Upper
169 100
168 66.8 52.5 78.7
167 35.8 28.4 42.6



Abundance Ion 169.00 (168.70 to 169.70): E
Ion 168.00 (167.70 to 168.70): E
Ion 167.00 (166.70 to 167.70): E









#67
4-Bromophenyl-phenylether
Concen: 41.990 ng
RT: 10.96 min Scan# 1508
Delta R.T. -0.01 min
Lab File: BF118107.D
Acq: 5 Dec 2019 15:52

Instrument :
BNA_F
ClientSampleId :
DFTPP

Tot Ion:248 Resp: 238608
Ion Ratio Lower Upper
248 100
250 95.2 78.0 117.0
141 72.6 64.0 96.0




Abundance Ion 248.00 (247.70 to 248.70): E
Ion 250.00 (249.70 to 250.70): E
Ion 141.00 (140.70 to 141.70): E






#68
Hexachlorobenzene
Concen: 41.234 ng
RT: 11.03 min Scan# 1520
Delta R.T. -0.01 min
Lab File: BF118107.D
Acq: 5 Dec 2019 15:52

Tgt Ion:284 Resp: 273223
Ion Ratio Lower Upper
284 100
142 29.9 27.3 40.9
249 29.9 24.8 37.2



Abundance Ion 283.80 (283.50 to 284.50): E
Ion 142.00 (141.70 to 142.70): E
Ion 249.00 (248.70 to 249.70): E









#69
Atrazine
Concen: 35.512 ng
RT: 11.11 min Scan# 1534
Delta R.T. -0.01 min
Lab File: BF118107.D
Acq: 5 Dec 2019 15:52




Instrument :
BNA_F
ClientSampleId :
DFTPP


Tot Ion:200 Resp: 179049
Ion Ratio Lower Upper
200 100
173 27.7 7.9 47.9
215 44.5 27.0 67.0



#70
Pentachlorophenol
Concen: 32.987 ng
RT: 11.22 min Scan# 1553
Delta R.T. -0.01 min
Lab File: BF118107.D
Acq: 5 Dec 2019 15:52

Tgt Ion:266 Resp: 127700
Ion Ratio Lower Upper
266 100
268 63.7 50.3 75.5
264 61.7 50.3 75.5









#71
Phenanthrene
Concen: 41.191 ng
RT: 11.44 min Scan# 1590
Delta R.T. -0.01 min
Lab File: BF118107.D
Acq: 5 Dec 2019 15:52




Instrument :
BNA_F
ClientSampleId :
DFTPP


Tot Ion:178 Resp: 1090685
Ion Ratio Lower Upper
178 100
176 19.2 15.4 23.0
179 15.8 12.3 18.5



#72
Anthracene
Concen: 41.252 ng
RT: 11.49 min Scan# 1599
Delta R.T. -0.01 min
Lab File: BF118107.D
Acq: 5 Dec 2019 15:52

Tgt Ion:178 Resp: 1082981
Ion Ratio Lower Upper
178 100
176 18.8 15.0 22.6
179 16.0 12.8 19.2





#73

Carbazole

Concen: 41.891 ng

RT: 11.65 min Scan# 1626

Delta R.T. -0.00 min

Lab File: BF118107.D

Acq: 5 Dec 2019 15:52

Instrument :

BNA_F

ClientSampleId :

DFTPP


Tot Ion:167 Resp: 961025

Ion Ratio Lower Upper




167 100

166 22.0 17.6 26.4

139 13.5 11.5 17.3



Abundance Ion 167.00 (166.70 to 167.70): E
Ion 166.00 (165.70 to 166.70): E
Ion 139.00 (138.70 to 139.70): E



#74

Di-n-butylphthalate

Concen: 42.527 ng

RT: 11.97 min Scan# 1681

Delta R.T. -0.01 min

Lab File: BF118107.D

Acq: 5 Dec 2019 15:52


Tgt Ion:149 Resp: 1214732

Ion Ratio Lower Upper



149 100


150 9.4 7.5 11.3

104 5.1 4.2 6.4



Abundance Ion 149.00 (148.70 to 149.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 104.00 (103.70 to 104.70): E









#75
 Fluoranthene
 Concen: 43.814 ng
 RT: 12.63 min Scan# 1793
 Delta R.T. -0.01 min
 Lab File: BF118107.D
 Acq: 5 Dec 2019 15:52




Instrument :
 BNA_F
 ClientSampleId :
 DFTPP


Tot Ion:202 Resp: 1061236
 Ion Ratio Lower Upper
 202 100
 101 11.6 0.0 31.5
 203 17.8 0.0 37.6



#76
 Chrysene-d12
 Concen: 20.000 ng
 RT: 14.07 min Scan# 2037
 Delta R.T. -0.01 min
 Lab File: BF118107.D
 Acq: 5 Dec 2019 15:52

Tgt Ion:240 Resp: 329062
 Ion Ratio Lower Upper
 240 100
 120 10.4 8.4 12.6
 236 25.9 20.6 31.0





#77

Benzidine

Concen: 36.782 ng

RT: 12.75 min Scan# 1813

Delta R.T. -0.01 min

Lab File: BF118107.D

Acq: 5 Dec 2019 15:52

Instrument :

BNA_F

ClientSampleId :

DFTPP


Tot Ion:184 Resp: 396462

Ion Ratio Lower Upper

184 100

185 15.2 11.6 17.4




183 11.6 9.7 14.5



Abundance Ion 184.00 (183.70 to 184.70): E

Ion 185.00 (184.70 to 185.70): E

Ion 183.00 (182.70 to 183.70): E



#78

Pyrene

Concen: 40.134 ng

RT: 12.86 min Scan# 1832

Delta R.T. -0.01 min

Lab File: BF118107.D

Acq: 5 Dec 2019 15:52


Tgt Ion:202 Resp: 1067300

Ion Ratio Lower Upper

202 100

200 21.0 17.0 25.6



203 17.9 14.2 21.4




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E





#79

Terphenyl-d14

Concen: 87.708 ng

RT: 13.00 min Scan# 1856

Delta R.T. -0.01 min

Lab File: BF118107.D

Acq: 5 Dec 2019 15:52

Instrument :

BNA_F

ClientSampleId :

DFTPP


Tot Ion:244 Resp: 1579181

Ion Ratio Lower Upper

244 100

212 7.1 5.2 7.8




122 11.7 7.2 10.8#



Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E



#80

Butylbenzylphthalate

Concen: 41.181 ng

RT: 13.48 min Scan# 1937

Delta R.T. -0.01 min

Lab File: BF118107.D

Acq: 5 Dec 2019 15:52


Tgt Ion:149 Resp: 470359

Ion Ratio Lower Upper

149 100

91 64.0 53.1 79.7



206 19.7 15.8 23.6




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 91.00 (90.70 to 91.70): BF1

Ion 206.00 (205.70 to 206.70): E









#81
Benzo(a)anthracene
Concen: 40.164 ng
RT: 14.06 min Scan# 2035
Delta R.T. -0.01 min
Lab File: BF118107.D
Acq: 5 Dec 2019 15:52




Instrument :
BNA_F
ClientSampleId :
DFTPP


Tot Ion:228 Resp: 873366
Ion Ratio Lower Upper
228 100
226 26.9 21.9 32.9
229 20.1 16.2 24.2



#82
3,3'-Dichlorobenzidine
Concen: 38.561 ng
RT: 14.02 min Scan# 2028
Delta R.T. -0.01 min
Lab File: BF118107.D
Acq: 5 Dec 2019 15:52

Tgt Ion:252 Resp: 293300
Ion Ratio Lower Upper
252 100
254 63.0 52.2 78.2
126 12.3 9.3 13.9





#83

Chrysene

Concen: 39.848 ng

RT: 14.09 min Scan# 2041

Delta R.T. -0.01 min

Lab File: BF118107.D

Acq: 5 Dec 2019 15:52

Instrument :

BNA_F

ClientSampleId :

DFTPP





Tot Ion:228 Resp: 839630

Ion Ratio Lower Upper

228 100

226 29.9 23.6 35.4

229 20.4 16.2 24.2



#84

Bis(2-ethylhexyl)phthalate

Concen: 42.872 ng

RT: 14.04 min Scan# 2032

Delta R.T. -0.01 min

Lab File: BF118107.D

Acq: 5 Dec 2019 15:52




Tgt Ion:149 Resp: 593665


Ion Ratio Lower Upper

149 100

167 26.1 21.8 32.8

279 4.1 3.4 5.0





#85

Di-n-octyl phthalate

Concen: 43.002 ng

RT: 14.66 min Scan# 2137

Delta R.T. -0.01 min

Lab File: BF118107.D

Acq: 5 Dec 2019 15:52

Instrument :

BNA_F

ClientSampleId :

DFTPP


Tot Ion:149 Resp: 874789

Ion Ratio Lower Upper

149 100

167 1.5 1.3 1.9




43 7.9 6.4 9.6



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 43.00 (42.70 to 43.70): BF1



#86

Indeno(1,2,3-cd)pyrene

Concen: 38.634 ng

RT: 17.07 min Scan# 2548

Delta R.T. -0.01 min

Lab File: BF118107.D

Acq: 5 Dec 2019 15:52


Tgt Ion:276 Resp: 692831

Ion Ratio Lower Upper

276 100

138 24.1 19.0 28.6



277 25.1 20.7 31.1




Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E





#87

Pervlene-d12

Concen: 20.000 ng

RT: 15.56 min Scan# 2291

Delta R.T. -0.01 min

Lab File: BF118107.D

Acq: 5 Dec 2019 15:52

Instrument :

BNA_F

ClientSampleId :

DFTPP


Tot Ion:264 Resp: 283227

Ion Ratio Lower Upper

264 100

260 24.1 19.0 28.6




265 22.3 16.4 24.6



Abundance Ion 264.00 (263.70 to 264.70): E

Ion 260.00 (259.70 to 260.70): E

Ion 265.00 (264.70 to 265.70): E



#88

Benzo(b)fluoranthene

Concen: 38.073 ng

RT: 15.12 min Scan# 2216

Delta R.T. -0.01 min

Lab File: BF118107.D

Acq: 5 Dec 2019 15:52


Tgt Ion:252 Resp: 700593

Ion Ratio Lower Upper

252 100

253 22.0 18.1 27.1



125 9.9 7.7 11.5




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E









#89
Benzo(k)fluoranthene
Concen: 40.826 ng
RT: 15.15 min Scan# 2221
Delta R.T. -0.01 min
Lab File: BF118107.D
Acq: 5 Dec 2019 15:52




Instrument :
BNA_F
ClientSampleId :
DFTPP


Tot Ion	Ratio	Lower	Upper
252	100		
253	22.3	17.3	25.9
125	10.1	8.5	12.7



#90
Benzo(a)pyrene
Concen: 39.578 ng
RT: 15.50 min Scan# 2280
Delta R.T. -0.01 min
Lab File: BF118107.D
Acq: 5 Dec 2019 15:52

Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.5	17.1	25.7
125	10.7	8.9	13.3





#91

Dibenzo(a,h)anthracene

Concen: 40.698 ng

RT: 17.09 min Scan# 2550

Delta R.T. -0.01 min

Lab File: BF118107.D

Acq: 5 Dec 2019 15:52

Instrument :

BNA_F

ClientSampleId :

DFTPP


Tot Ion:278 Resp: 566810

Ion Ratio Lower Upper




278 100

139 15.6 12.2 18.4

279 23.4 19.0 28.6



Abundance Ion 278.00 (277.70 to 278.70): E
Ion 139.00 (138.70 to 139.70): E
Ion 279.00 (278.70 to 279.70): E



#92

Benzo(g,h,i)perylene

Concen: 40.435 ng

RT: 17.53 min Scan# 2625

Delta R.T. -0.01 min

Lab File: BF118107.D

Acq: 5 Dec 2019 15:52


Tgt Ion:276 Resp: 560909

Ion Ratio Lower Upper

276 100

277 23.4 19.0 28.4

138 22.2 16.6 25.0



Abundance Ion 276.00 (275.70 to 276.70): E
Ion 277.00 (276.70 to 277.70): E
Ion 138.00 (137.70 to 138.70): E

