

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF120522\
 Data File : BF131511.D
 Acq On : 05 Dec 2022 20:19
 Operator : CG\JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF120522

Quant Time: Dec 06 01:36:39 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF120522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 06 01:30:30 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	119	0.00
2	1,4-Dioxane	40.000	36.457	8.9	113	0.01
3	Pyridine	40.000	36.276	9.3	112	0.00
4	n-Nitrosodimethylamine	40.000	33.906	15.2	103	0.00
5 S	2-Fluorophenol	80.000	75.838	5.2	118	0.00
6	Aniline	40.000	36.427	8.9	114	0.00
7 S	Phenol-d6	80.000	74.238	7.2	116	0.00
8	2-Chlorophenol	40.000	38.674	3.3	123	0.00
9	Benzaldehyde	40.000	33.424	16.4	109	0.00
10 C	Phenol	40.000	37.577	6.1	118	0.00
11	bis(2-Chloroethyl)ether	40.000	36.230	9.4	113	0.00
12	1,3-Dichlorobenzene	40.000	37.932	5.2	120	0.00
13 C	1,4-Dichlorobenzene	40.000	37.853	5.4	120	0.00
14	1,2-Dichlorobenzene	40.000	38.272	4.3	120	0.00
15	Benzyl Alcohol	40.000	36.717	8.2	111	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	35.172	12.1	109	0.00
17	2-Methylphenol	40.000	37.654	5.9	117	0.00
18	Hexachloroethane	40.000	37.733	5.7	118	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	34.600	13.5	109	0.00
20	3+4-Methylphenols	40.000	37.553	6.1	115	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	120	0.00
22	Acetophenone	40.000	37.688	5.8	114	0.00
23 S	Nitrobenzene-d5	80.000	73.377	8.3	110	0.00
24	Nitrobenzene	40.000	36.016	10.0	108	0.00
25	Isophorone	40.000	36.768	8.1	110	0.00
26 C	2-Nitrophenol	40.000	43.155	-7.9	130	0.00
27	2,4-Dimethylphenol	40.000	40.643	-1.6	121	0.00
28	bis(2-Chloroethoxy)methane	40.000	37.519	6.2	113	0.00
29 C	2,4-Dichlorophenol	40.000	40.164	-0.4	122	0.00
30	1,2,4-Trichlorobenzene	40.000	38.923	2.7	120	0.00
31	Naphthalene	40.000	37.808	5.5	121	0.00
32	Benzoic acid	40.000	46.309	-15.8	136	0.00
33	4-Chloroaniline	40.000	38.305	4.2	120	0.00
34 C	Hexachlorobutadiene	40.000	38.241	4.4	119	0.00
35	Caprolactam	40.000	38.938	2.7	120	0.00
36 C	4-Chloro-3-methylphenol	40.000	38.130	4.7	117	0.00
37	2-Methylnaphthalene	40.000	38.375	4.1	120	0.00
38	1-Methylnaphthalene	40.000	38.410	4.0	121	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	121	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	37.905	5.2	121	0.00
41 P	Hexachlorocyclopentadiene	40.000	40.358	-0.9	123	0.00
42 S	2,4,6-Tribromophenol	80.000	82.777	-3.5	130	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.642	0.9	125	0.00
44	2,4,5-Trichlorophenol	40.000	39.804	0.5	123	0.00
45 S	2-Fluorobiphenyl	80.000	75.333	5.8	119	0.00
46	1,1'-Biphenyl	40.000	37.619	6.0	121	0.00
47	2-Chloronaphthalene	40.000	37.674	5.8	122	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF120522\
 Data File : BF131511.D
 Acq On : 05 Dec 2022 20:19
 Operator : CG\JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF120522

Quant Time: Dec 06 01:36:39 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF120522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 06 01:30:30 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	36.635	8.4	112	0.00
49	Acenaphthylene	40.000	37.477	6.3	120	0.00
50	Dimethylphthalate	40.000	38.043	4.9	122	0.00
51	2,6-Dinitrotoluene	40.000	39.206	2.0	122	0.00
52 C	Acenaphthene	40.000	38.154	4.6	119	0.00
53	3-Nitroaniline	40.000	38.369	4.1	120	0.00
54 P	2,4-Dinitrophenol	40.000	42.115	-5.3	136	0.00
55	Dibenzofuran	40.000	37.262	6.8	119	0.00
56 P	4-Nitrophenol	40.000	41.186	-3.0	122	0.00
57	2,4-Dinitrotoluene	40.000	40.362	-0.9	122	0.00
58	Fluorene	40.000	37.821	5.4	119	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.028	-0.1	124	0.00
60	Diethylphthalate	40.000	38.247	4.4	121	0.00
61	4-Chlorophenyl-phenylether	40.000	37.203	7.0	117	0.00
62	4-Nitroaniline	40.000	38.931	2.7	121	0.00
63	Azobenzene	40.000	35.016	12.5	108	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	118	0.00
65	4,6-Dinitro-2-methylphenol	40.000	43.963	-9.9	133	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.666	3.3	120	0.00
67	4-Bromophenyl-phenylether	40.000	39.041	2.4	121	0.00
68	Hexachlorobenzene	40.000	39.932	0.2	125	0.00
69	Atrazine	40.000	40.912	-2.3	122	0.00
70 C	Pentachlorophenol	40.000	41.308	-3.3	127	0.00
71	Phenanthrene	40.000	38.838	2.9	120	0.00
72	Anthracene	40.000	38.153	4.6	119	0.00
73	Carbazole	40.000	39.198	2.0	119	0.00
74	Di-n-butylphthalate	40.000	39.783	0.5	118	0.00
75 C	Fluoranthene	40.000	38.005	5.0	116	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	118	0.00
77	Benzydine	40.000	35.732	10.7	106	0.00
78	Pyrene	40.000	38.173	4.6	117	0.00
79 S	Terphenyl-d14	80.000	79.053	1.2	118	0.00
80	Butylbenzylphthalate	40.000	44.392	-11.0	127	0.00
81	Benzo(a)anthracene	40.000	37.921	5.2	115	0.00
82	3,3'-Dichlorobenzidine	40.000	38.291	4.3	117	0.00
83	Chrysene	40.000	37.760	5.6	117	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	42.877	-7.2	128	0.00
85 c	Di-n-octyl phthalate	40.000	41.397	-3.5	136	0.00
86 I	Perylene-d12	20.000	20.000	0.0	125	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	36.979	7.6	107	0.00
88	Benzo(b)fluoranthene	40.000	37.085	7.3	123	0.00
89	Benzo(k)fluoranthene	40.000	37.387	6.5	124	0.00
90 C	Benzo(a)pyrene	40.000	39.044	2.4	124	0.00
91	Dibenzo(a,h)anthracene	40.000	37.264	6.8	110	0.00
92	Benzo(g,h,i)perylene	40.000	40.027	-0.1	106	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF120522\
Data File : BF131511.D
Acq On : 05 Dec 2022 20:19
Operator : CG\JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF120522

Quant Time: Dec 06 01:36:39 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF120522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 06 01:30:30 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0