

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF120523\
 Data File : BF136417.D
 Acq On : 05 Dec 2023 20:53
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 06 09:42:43 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF120523.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 06 09:40:27 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	100	0.00
2	1,4-Dioxane	40.000	41.575	-3.9	100	0.00
3	Pyridine	40.000	43.309	-8.3	101	0.00
4	n-Nitrosodimethylamine	40.000	41.876	-4.7	99	0.00
5 S	2-Fluorophenol	80.000	83.408	-4.3	100	0.00
6	Aniline	40.000	42.260	-5.6	100	0.00
7 S	Phenol-d6	80.000	83.161	-4.0	100	0.00
8	2-Chlorophenol	40.000	41.890	-4.7	101	0.00
9	Benzaldehyde	40.000	46.836	-17.1	106	0.00
10 C	Phenol	40.000	41.516	-3.8	100	0.00
11	bis(2-Chloroethyl)ether	40.000	41.178	-2.9	100	0.00
12	1,3-Dichlorobenzene	40.000	41.985	-5.0	101	0.00
13 C	1,4-Dichlorobenzene	40.000	41.416	-3.5	100	0.00
14	1,2-Dichlorobenzene	40.000	41.734	-4.3	101	0.00
15	Benzyl Alcohol	40.000	41.578	-3.9	98	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	41.404	-3.5	100	0.00
17	2-Methylphenol	40.000	41.797	-4.5	100	0.00
18	Hexachloroethane	40.000	42.060	-5.2	100	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	41.988	-5.0	101	0.00
20	3+4-Methylphenols	40.000	41.527	-3.8	101	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	102	0.00
22	Acetophenone	40.000	40.900	-2.2	100	0.00
23 S	Nitrobenzene-d5	80.000	82.965	-3.7	100	0.00
24	Nitrobenzene	40.000	41.488	-3.7	101	0.00
25	Isophorone	40.000	41.825	-4.6	101	0.00
26 C	2-Nitrophenol	40.000	43.236	-8.1	102	0.00
27	2,4-Dimethylphenol	40.000	42.047	-5.1	101	0.00
28	bis(2-Chloroethoxy)methane	40.000	41.073	-2.7	100	0.00
29 C	2,4-Dichlorophenol	40.000	41.915	-4.8	100	0.00
30	1,2,4-Trichlorobenzene	40.000	41.919	-4.8	102	0.00
31	Naphthalene	40.000	40.972	-2.4	100	0.00
32	Benzoic acid	40.000	44.752	-11.9	103	0.00
33	4-Chloroaniline	40.000	42.019	-5.0	101	0.00
34 C	Hexachlorobutadiene	40.000	41.411	-3.5	101	0.00
35	Caprolactam	40.000	42.358	-5.9	102	0.00
36 C	4-Chloro-3-methylphenol	40.000	41.811	-4.5	101	0.00
37	2-Methylnaphthalene	40.000	41.652	-4.1	102	0.00
38	1-Methylnaphthalene	40.000	41.239	-3.1	101	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	103	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	41.709	-4.3	104	0.00
41 P	Hexachlorocyclopentadiene	40.000	43.639	-9.1	104	0.00
42 S	2,4,6-Tribromophenol	80.000	82.700	-3.4	101	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.643	-4.1	102	0.00
44	2,4,5-Trichlorophenol	40.000	42.065	-5.2	102	0.00
45 S	2-Fluorobiphenyl	80.000	80.754	-0.9	101	0.00
46	1,1'-Biphenyl	40.000	40.670	-1.7	101	0.00
47	2-Chloronaphthalene	40.000	40.543	-1.4	101	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF120523\
 Data File : BF136417.D
 Acq On : 05 Dec 2023 20:53
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 06 09:42:43 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF120523.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 06 09:40:27 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	42.352	-5.9	102	0.00
49	Acenaphthylene	40.000	40.988	-2.5	102	0.00
50	Dimethylphthalate	40.000	40.811	-2.0	101	0.00
51	2,6-Dinitrotoluene	40.000	40.990	-2.5	100	0.00
52 C	Acenaphthene	40.000	40.721	-1.8	101	0.00
53	3-Nitroaniline	40.000	41.455	-3.6	101	0.00
54 P	2,4-Dinitrophenol	40.000	44.336	-10.8	101	0.00
55	Dibenzofuran	40.000	40.604	-1.5	101	0.00
56 P	4-Nitrophenol	40.000	43.112	-7.8	100	0.00
57	2,4-Dinitrotoluene	40.000	40.469	-1.2	98	0.00
58	Fluorene	40.000	40.240	-0.6	101	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	41.937	-4.8	102	0.00
60	Diethylphthalate	40.000	40.944	-2.4	101	0.00
61	4-Chlorophenyl-phenylether	40.000	40.170	-0.4	102	0.00
62	4-Nitroaniline	40.000	41.972	-4.9	99	0.00
63	Azobenzene	40.000	40.923	-2.3	101	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	99	0.00
65	4,6-Dinitro-2-methylphenol	40.000	46.039	-15.1	100	0.00
66 c	n-Nitrosodiphenylamine	40.000	41.647	-4.1	101	0.00
67	4-Bromophenyl-phenylether	40.000	42.303	-5.8	100	0.00
68	Hexachlorobenzene	40.000	42.459	-6.1	102	0.00
69	Atrazine	40.000	39.352	1.6	101	0.00
70 C	Pentachlorophenol	40.000	43.624	-9.1	101	0.00
71	Phenanthrene	40.000	41.256	-3.1	99	0.00
72	Anthracene	40.000	41.686	-4.2	100	0.00
73	Carbazole	40.000	41.758	-4.4	100	0.00
74	Di-n-butylphthalate	40.000	42.680	-6.7	101	0.00
75 C	Fluoranthene	40.000	41.088	-2.7	98	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	98	0.00
77	Benzydine	40.000	39.825	0.4	104	0.00
78	Pyrene	40.000	42.004	-5.0	98	0.00
79 S	Terphenyl-d14	80.000	84.542	-5.7	100	0.00
80	Butylbenzylphthalate	40.000	44.246	-10.6	101	0.00
81	Benzo(a)anthracene	40.000	40.994	-2.5	98	0.00
82	3,3'-Dichlorobenzidine	40.000	43.045	-7.6	104	0.00
83	Chrysene	40.000	41.165	-2.9	97	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	44.983	-12.5	103	0.00
85 c	Di-n-octyl phthalate	40.000	46.434	-16.1	107	0.00
86 I	Perylene-d12	20.000	20.000	0.0	105	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	43.407	-8.5	103	0.00
88	Benzo(b)fluoranthene	40.000	39.582	1.0	103	0.00
89	Benzo(k)fluoranthene	40.000	41.307	-3.3	102	0.00
90 C	Benzo(a)pyrene	40.000	41.008	-2.5	103	0.00
91	Dibenzo(a,h)anthracene	40.000	42.923	-7.3	102	0.00
92	Benzo(g,h,i)perylene	40.000	43.403	-8.5	103	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF120523\
Data File : BF136417.D
Acq On : 05 Dec 2023 20:53
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Dec 06 09:42:43 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF120523.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 06 09:40:27 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0