

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF120618\
 Data File : BF111150.D
 Acq On : 6 Dec 2018 20:49
 Operator : JU/SJ
 Sample : J6204-06
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 HLSP-SB-2A-(0-2)

Quant Time: Dec 06 22:51:13 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF120518.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 05 17:30:55 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.80	152	501320	20.00	ng	0.00
21) Naphthalene-d8	8.09	136	1997466	20.00	ng	0.00
39) Acenaphthene-d10	9.84	164	954824	20.00	ng	0.00
64) Phenanthrene-d10	11.32	188	1634183	20.00	ng	0.00
76) Chrysene-d12	13.96	240	964160	20.00	ng	0.00
87) Perylene-d12	15.39	264	954426	20.00	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.43	112	3976286	123.77	ng	0.01
7) Phenol-d6	6.44	99	4997580	120.76	ng	0.00
23) Nitrobenzene-d5	7.36	82	3283838	97.33	ng	0.00
42) 2,4,6-Tribromophenol	10.63	330	1065094	140.33	ng	0.00
45) 2-Fluorobiphenyl	9.17	172	4663708	91.51	ng	0.00
79) Terphenyl-d14	12.91	244	3909746	83.05	ng	0.00
Target Compounds						
10) Phenol	6.45	94	330329	7.320	ng	90
50) Dimethylphthalate	9.56	163	946790	14.248	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF120618\
Data File : BF111150.D
Acq On : 6 Dec 2018 20:49
Operator : JU/SJ
Sample : J6204-06
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
HLSP-SB-2A-(0-2)

Quant Time: Dec 06 22:51:13 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF120518.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 05 17:30:55 2018
Response via : Initial Calibration

