

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF120623\
 Data File : BF136419.D
 Acq On : 06 Dec 2023 10:06
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 06 11:00:24 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF120523.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 06 09:40:27 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	106	0.00
2	1,4-Dioxane	0.536	0.695	-29.7#	131	0.00
3	Pyridine	1.297	1.580	-21.8	120	0.00
4	n-Nitrosodimethylamine	0.564	0.578	-2.5	102	0.00
5 S	2-Fluorophenol	1.354	1.311	3.2	99	0.00
6	Aniline	1.708	2.094	-22.6	123	0.00
7 S	Phenol-d6	1.690	1.642	2.8	99	0.00
8	2-Chlorophenol	1.478	1.458	1.4	101	0.00
9	Benzaldehyde	0.896	0.905	-1.0	105	0.00
10 C	Phenol	1.798	1.784	0.8	101	0.00
11	bis(2-Chloroethyl)ether	1.340	1.326	1.0	102	0.00
12	1,3-Dichlorobenzene	1.663	1.497	10.0	92	0.00
13 C	1,4-Dichlorobenzene	1.687	1.536	9.0	93	0.00
14	1,2-Dichlorobenzene	1.586	1.431	9.8	93	0.00
15	Benzyl Alcohol	1.131	1.121	0.9	99	0.00
16	2,2'-oxybis(1-Chloropropane	1.724	1.671	3.1	99	0.00
17	2-Methylphenol	1.194	1.300	-8.9	110	0.00
18	Hexachloroethane	0.589	0.545	7.5	93	0.00
19 P	n-Nitroso-di-n-propylamine	0.975	0.961	1.4	100	0.00
20	3+4-Methylphenols	1.483	1.517	-2.3	105	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	107	0.00
22	Acetophenone	0.494	0.478	3.2	100	0.00
23 S	Nitrobenzene-d5	0.370	0.365	1.4	100	0.00
24	Nitrobenzene	0.372	0.356	4.3	98	0.00
25	Isophorone	0.665	0.651	2.1	100	0.00
26 C	2-Nitrophenol	0.183	0.182	0.5	99	0.00
27	2,4-Dimethylphenol	0.226	0.292	-29.2#	131	0.00
28	bis(2-Chloroethoxy)methane	0.408	0.398	2.5	100	0.00
29 C	2,4-Dichlorophenol	0.298	0.289	3.0	97	0.00
30	1,2,4-Trichlorobenzene	0.347	0.319	8.1	94	0.00
31	Naphthalene	1.103	1.045	5.3	97	0.00
32	Benzoic acid	0.224	0.224	0.0	97	0.00
33	4-Chloroaniline	0.393	0.438	-11.5	112	0.00
34 C	Hexachlorobutadiene	0.206	0.184	10.7	91	0.00
35	Caprolactam	0.092	0.093	-1.1	103	0.00
36 C	4-Chloro-3-methylphenol	0.316	0.311	1.6	99	0.00
37	2-Methylnaphthalene	0.702	0.691	1.6	102	0.00
38	1-Methylnaphthalene	0.689	0.645	6.4	96	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	105	0.00
40	1,2,4,5-Tetrachlorobenzene	0.622	0.620	0.3	101	0.00
41 P	Hexachlorocyclopentadiene	0.231	0.304	-31.6#	128	0.00
42 S	2,4,6-Tribromophenol	0.247	0.247	0.0	100	0.00
43 C	2,4,6-Trichlorophenol	0.405	0.393	3.0	97	0.00
44	2,4,5-Trichlorophenol	0.452	0.472	-4.4	104	0.00
45 S	2-Fluorobiphenyl	1.466	1.428	2.6	99	0.00
46	1,1'-Biphenyl	1.702	1.687	0.9	100	0.00
47	2-Chloronaphthalene	1.325	1.238	6.6	95	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF120623\
 Data File : BF136419.D
 Acq On : 06 Dec 2023 10:06
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 06 11:00:24 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF120523.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 06 09:40:27 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.355	0.379	-6.8	105	0.00
49	Acenaphthylene	1.887	1.947	-3.2	104	0.00
50	Dimethylphthalate	1.472	1.513	-2.8	104	0.00
51	2,6-Dinitrotoluene	0.331	0.322	2.7	97	0.00
52 C	Acenaphthene	1.209	1.206	0.2	101	0.00
53	3-Nitroaniline	0.338	0.354	-4.7	104	0.00
54 P	2,4-Dinitrophenol	0.167	0.209	-25.1#	116	0.00
55	Dibenzofuran	1.766	1.796	-1.7	103	0.00
56 P	4-Nitrophenol	0.254	0.252	0.8	94	0.00
57	2,4-Dinitrotoluene	0.438	0.438	0.0	99	0.00
58	Fluorene	1.418	1.391	1.9	100	0.00
59	2,3,4,6-Tetrachlorophenol	0.362	0.369	-1.9	101	0.00
60	Diethylphthalate	1.448	1.459	-0.8	101	0.00
61	4-Chlorophenyl-phenylether	0.691	0.690	0.1	103	0.00
62	4-Nitroaniline	0.336	0.354	-5.4	102	0.00
63	Azobenzene	1.318	1.321	-0.2	101	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	104	0.00
65	4,6-Dinitro-2-methylphenol	0.124	0.131	-5.6	97	0.00
66 c	n-Nitrosodiphenylamine	0.642	0.651	-1.4	104	0.00
67	4-Bromophenyl-phenylether	0.233	0.231	0.9	100	0.00
68	Hexachlorobenzene	0.269	0.246	8.6	93	0.00
69	Atrazine	0.183	0.150	18.0	88	0.00
70 C	Pentachlorophenol	0.152	0.147	3.3	94	0.00
71	Phenanthrene	1.099	1.112	-1.2	103	0.00
72	Anthracene	1.107	1.134	-2.4	104	0.00
73	Carbazole	0.980	0.958	2.2	98	0.00
74	Di-n-butylphthalate	1.201	1.226	-2.1	102	0.00
75 C	Fluoranthene	1.152	1.137	1.3	99	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	111	0.00
77	Benzydine	0.402	0.337	16.2	88	0.00
78	Pyrene	1.959	1.829	6.6	99	0.00
79 S	Terphenyl-d14	1.517	1.462	3.6	103	0.00
80	Butylbenzylphthalate	0.697	0.726	-4.2	107	0.00
81	Benzo(a)anthracene	1.393	1.362	2.2	106	0.00
82	3,3'-Dichlorobenzidine	0.391	0.436	-11.5	122	0.00
83	Chrysene	1.371	1.361	0.7	106	0.00
84	Bis(2-ethylhexyl)phthalate	0.855	0.921	-7.7	112	0.00
85 c	Di-n-octyl phthalate	1.263	1.395	-10.5	115	0.00
86 I	Perylene-d12	1.000	1.000	0.0	107	0.00
87	Indeno(1,2,3-cd)pyrene	1.389	1.405	-1.2	98	0.00
88	Benzo(b)fluoranthene	1.367	1.260	7.8	98	0.00
89	Benzo(k)fluoranthene	1.289	1.326	-2.9	104	0.00
90 C	Benzo(a)pyrene	1.129	1.184	-4.9	107	0.00
91	Dibenzo(a,h)anthracene	1.138	1.159	-1.8	99	0.00
92	Benzo(g,h,i)perylene	1.168	1.171	-0.3	97	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF120623\
Data File : BF136419.D
Acq On : 06 Dec 2023 10:06
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Dec 06 11:00:24 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF120523.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 06 09:40:27 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0