

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF120723\
 Data File : BF136443.D
 Acq On : 07 Dec 2023 12:50
 Operator : CG\JU
 Sample : PB157532BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB157532BL

Quant Time: Dec 07 18:57:29 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF120523.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 06 09:40:27 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.798	152	119333	20.000	ng	0.00
21) Naphthalene-d8	8.075	136	490071	20.000	ng	0.00
39) Acenaphthene-d10	9.834	164	257715	20.000	ng	0.00
64) Phenanthrene-d10	11.328	188	503363	20.000	ng	0.00
76) Chrysene-d12	13.974	240	344242	20.000	ng	# 0.00
86) Perylene-d12	15.457	264	269154	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.451	112	1141933	141.322	ng	0.02
7) Phenol-d6	6.440	99	1416850	140.540	ng	0.00
23) Nitrobenzene-d5	7.363	82	865657	95.492	ng	0.00
42) 2,4,6-Tribromophenol	10.633	330	456754	143.740	ng	0.00
45) 2-Fluorobiphenyl	9.157	172	1778191	94.148	ng	0.00
79) Terphenyl-d14	12.927	244	2339045	89.558	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF120723\
Data File : BF136443.D
Acq On : 07 Dec 2023 12:50
Operator : CG\JU
Sample : PB157532BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB157532BL

Quant Time: Dec 07 18:57:29 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF120523.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 06 09:40:27 2023
Response via : Initial Calibration

