





Method Name: Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF120818.M
Calibration Table Last Updated: Mon Dec 10 01:15:27 2018

2-Fluorobiphenyl

Response Ratio




$R = -6.69e-002 A^2 + 1.37e+000 A + 2.76e-001$
Coef of Det (r^2) = 0.993 Curve Fit: Quadratic

Method Name: Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF120818.M
Calibration Table Last Updated: Mon Dec 10 01:15:27 2018

1,1'-Biphenyl


Response Ratio



$R = -1.74e-001 A^2 + 1.94e+000 A + 7.52e-002$
Coef of Det (r^2) = 0.998 Curve Fit: Quadratic

Method Name: Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF120818.M
Calibration Table Last Updated: Mon Dec 10 01:15:27 2018

4,6-Dinitro-2-methylphenol



Resp Ratio = $1.21 \times 10^{-1} \times \text{Amt} - 5.97 \times 10^{-3}$
Coef of Det (r^2) = 0.996 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF120818.M
Calibration Table Last Updated: Mon Dec 10 01:15:27 2018