

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF121018\
 Data File : BF111264.D
 Acq On : 11 Dec 2018 1:21
 Operator : JU/SJ
 Sample : J6196-03 2X
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 EO-03-120718-B

Quant Time: Dec 11 06:31:47 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF120818.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 10 01:15:27 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.80	152	456851	20.00	ng	0.00
21) Naphthalene-d8	8.09	136	1641032	20.00	ng	0.00
39) Acenaphthene-d10	9.84	164	648955	20.00	ng	0.00
64) Phenanthrene-d10	11.32	188	1019610	20.00	ng	0.00
76) Chrysene-d12	13.96	240	699562	20.00	ng	0.00
87) Perylene-d12	15.40	264	411827	20.00	ng	0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.42	112	1517227	52.44	ng	0.00
7) Phenol-d6	6.43	99	1857794	51.67	ng	0.00
23) Nitrobenzene-d5	7.36	82	1056592	36.11	ng	0.00
42) 2,4,6-Tribromophenol	10.62	330	264296	45.98	ng	0.00
45) 2-Fluorobiphenyl	9.16	172	1605119	35.08	ng	0.00
79) Terphenyl-d14	12.91	244	1220449	41.85	ng	0.00
Target Compounds						
10) Phenol	6.44	94	103945	2.472	ng	94
50) Dimethylphthalate	9.55	163	117721	2.458	ng	98
71) Phenanthrene	11.35	178	122109	2.374	ng	93
75) Fluoranthene	12.53	202	162072	3.450	ng	100
78) Pyrene	12.76	202	170079	3.402	ng	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF121018\
Data File : BF111264.D
Acq On : 11 Dec 2018 1:21
Operator : JU/SJ
Sample : J6196-03 2X
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
EO-03-120718-B

Quant Time: Dec 11 06:31:47 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF120818.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 10 01:15:27 2018
Response via : Initial Calibration

