

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF121019\  
 Data File : BF118165.D  
 Acq On : 10 Dec 2019 14:14  
 Operator : JU  
 Sample : K6119-05RE  
 Misc :  
 ALS Vial : 7 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 SB-4-(0-2)RE

Quant Time: Dec 10 15:43:16 2019  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF120619.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Fri Dec 06 17:34:43 2019  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units | Dev(Min)     |
|-----------------------------|-------|------|----------|--------|-------|--------------|
| 1) 1,4-Dichlorobenzene-d4   | 6.87  | 152  | 85427    | 20.00  | ng    | 0.00         |
| 21) Naphthalene-d8          | 8.16  | 136  | 313208   | 20.00  | ng    | 0.00         |
| 39) Acenaphthene-d10        | 9.92  | 164  | 142059   | 20.00  | ng    | 0.00         |
| 64) Phenanthrene-d10        | 11.41 | 188  | 239705   | 20.00  | ng    | 0.00         |
| 76) Chrysene-d12            | 14.06 | 240  | 212079   | 20.00  | ng    | 0.00         |
| 87) Perylene-d12            | 15.56 | 264  | 166291   | 20.00  | ng    | 0.01         |
| System Monitoring Compounds |       |      |          |        |       |              |
| 5) 2-Fluorophenol           | 5.49  | 112  | 526033   | 107.85 | ng    | 0.00         |
| 7) Phenol-d6                | 6.50  | 99   | 642773   | 108.96 | ng    | 0.00         |
| 23) Nitrobenzene-d5         | 7.43  | 82   | 377896   | 67.88  | ng    | 0.00         |
| 42) 2,4,6-Tribromophenol    | 10.71 | 330  | 160320   | 92.90  | ng    | 0.00         |
| 45) 2-Fluorobiphenyl        | 9.23  | 172  | 731015   | 78.30  | ng    | 0.00         |
| 79) Terphenyl-d14           | 13.00 | 244  | 778148   | 63.10  | ng    | 0.00         |
| Target Compounds            |       |      |          |        |       |              |
| 50) Dimethylphthalate       | 9.63  | 163  | 116436   | 11.096 | ng    | Qvalue<br>99 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF121019\  
Data File : BF118165.D  
Acq On : 10 Dec 2019 14:14  
Operator : JU  
Sample : K6119-05RE  
Misc :  
ALS Vial : 7 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
SB-4-(0-2)RE

Quant Time: Dec 10 15:43:16 2019  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF120619.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Fri Dec 06 17:34:43 2019  
Response via : Initial Calibration

