


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF121118\
 Data File : BF111270.D
 Acq On : 11 Dec 2018 10:43
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF120818.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Dec 10 01:15:27 2018



AutoFind: Scans 1614, 1615, 1616; Background Corrected with Scan 1609

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	45.4	99784	PASS
68	69	0.00	2	1.7	1465	PASS
69	198	0.00	100	38.5	84607	PASS
70	69	0.00	2	0.5	384	PASS
127	198	10	80	50.9	111848	PASS
197	198	0.00	2	0.3	663	PASS
198	198	100	100	100.0	219792	PASS
199	198	5	9	6.4	14159	PASS
275	198	10	60	23.2	50912	PASS
365	198	1	100	2.4	5343	PASS
441	198	0.01	100	14.6	31984	PASS
442	198	50	100	93.9	206485	PASS
443	442	15	24	19.6	40498	PASS