

Data Path : Z:\HPCHEM1\BNA F\DATA\BF121217\
 Data File : BF101329.D
 Acq On : 12 Dec 2017 15:00
 Operator : SJ/JU
 Sample : PB104934BL
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB104934BL

Quant Time: Dec 13 01:28:50 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF113017.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 08 18:17:31 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.82	152	41673	20.00	ng	0.00
21) Naphthalene-d8	8.10	136	166089	20.00	ng	0.00
38) Acenaphthene-d10	9.86	164	81609	20.00	ng	-0.01
63) Phenanthrene-d10	11.35	188	171424	20.00	ng	0.00
75) Chrysene-d12	14.00	240	162462	20.00	ng	0.00
86) Perylene-d12	15.48	264	153478	20.00	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.45	112	340984	135.21	ng	0.01
7) Phenol-d6	6.46	99	410536	134.08	ng	0.00
23) Nitrobenzene-d5	7.39	82	256395	92.09	ng	-0.01
41) 2,4,6-Tribromophenol	10.66	330	135137	137.84	ng	0.00
44) 2-Fluorobiphenyl	9.18	172	571628	95.84	ng	0.00
78) Terphenyl-d14	12.95	244	664012	89.40	ng	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF121217\
Data File : BF101329.D
Acq On : 12 Dec 2017 15:00
Operator : SJ/JU
Sample : PB104934BL
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB104934BL

Quant Time: Dec 13 01:28:50 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF113017.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 08 18:17:31 2017
Response via : Initial Calibration

