

Data Path : Z:\HPCHEM1\BNA F\DATA\BF121217\
 Data File : BF101335.D
 Acq On : 12 Dec 2017 17:43
 Operator : SJ/JU
 Sample : I6822-01
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-1(0-2)

Manual Integrations
 APPROVED

Sohil
 12/13/2017 12:07:38 PM

Quant Time: Dec 13 11:12:01 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF113017.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 08 18:17:31 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.82	152	45854	20.00	ng	0.00
21) Naphthalene-d8	8.10	136	173753	20.00	ng	0.00
38) Acenaphthene-d10	9.87	164	72791	20.00	ng	0.00
63) Phenanthrene-d10	11.36	188	120762	20.00	ng	0.00
75) Chrysene-d12	14.01	240	102439	20.00	ng	0.00
86) Perylene-d12	15.49	264	81615	20.00	ng	0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.45	112	13604	4.90	ng	0.01
7) Phenol-d6	6.46	99	249478	74.05	ng	0.00
23) Nitrobenzene-d5	7.39	82	212781	73.05	ng	0.00
41) 2,4,6-Tribromophenol	10.67	330	219	0.25	ng	0.00
44) 2-Fluorobiphenyl	9.19	172	444985	83.64	ng	0.00
78) Terphenyl-d14	12.94	244	355624	75.93	ng	0.00
Target Compounds						
10) Phenol	6.47	94	10856	2.898	ng	79
31) Naphthalene	8.13	128	60980	7.121	ng	98
37) 2-Methylnaphthalene	8.82	142	110942	19.066	ng	95
49) Dimethylphthalate	9.57	163	47553	8.100	ng	# 93
70) Phenanthrene	11.39	178	79258	11.918	ng	99
74) Fluoranthene	12.57	202	24546	3.330	ng	93
77) Pyrene	12.80	202	22375	3.165	ng	98
87) Benzo(b)fluoranthene	15.05	252	11818m	2.418	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF121217\
Data File : BF101335.D
Acq On : 12 Dec 2017 17:43
Operator : SJ/JU
Sample : I6822-01
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB-1(0-2)

Quant Time: Dec 13 11:12:01 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF113017.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 08 18:17:31 2017
Response via : Initial Calibration

Manual Integrations
APPROVED

Sohil
12/13/2017 12:07:38 PM

