



Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF121323\  
 Data File : BF136530.D  
 Acq On : 13 Dec 2023 10:50  
 Operator : CG\JU  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA\_F\Methods\8270-BF120523.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Wed Dec 06 09:40:27 2023






AutoFind: Scans 1613, 1614, 1615; Background Corrected with Scan 1608

| Target<br>Mass | Rel. to<br>Mass | Lower<br>Limit% | Upper<br>Limit% | Rel.<br>Abn% | Raw<br>Abn | Result<br>Pass/Fail |
|----------------|-----------------|-----------------|-----------------|--------------|------------|---------------------|
| 51             | 198             | 10              | 80              | 30.5         | 47456      | PASS                |
| 68             | 69              | 0.00            | 2               | 1.9          | 963        | PASS                |
| 69             | 198             | 0.00            | 100             | 31.8         | 49437      | PASS                |
| 70             | 69              | 0.00            | 2               | 0.5          | 258        | PASS                |
| 127            | 198             | 10              | 80              | 47.2         | 73477      | PASS                |
| 197            | 198             | 0.00            | 2               | 0.4          | 668        | PASS                |
| 198            | 198             | 100             | 100             | 100.0        | 155595     | PASS                |
| 199            | 198             | 5               | 9               | 6.7          | 10351      | PASS                |
| 275            | 198             | 10              | 60              | 26.3         | 40971      | PASS                |
| 365            | 198             | 1               | 100             | 3.5          | 5374       | PASS                |
| 441            | 198             | 0.01            | 100             | 14.1         | 21971      | PASS                |
| 442            | 442             | 50              | 100             | 100.0        | 138427     | PASS                |
| 443            | 442             | 15              | 24              | 19.6         | 27065      | PASS                |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF121323\  
 Data File : BF136530.D  
 Acq On : 13 Dec 2023 10:50  
 Operator : CG\JU  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 DFTPP

Quant Time: Dec 13 12:05:01 2023  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_F\Methods\8270-BF120523.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Wed Dec 06 09:40:27 2023  
 Response via : Initial Calibration



TIC: BF136530.D\data.ms

**(70) Pentachlorophenol (C)**

11.127min (-0.000) 23588.07 ng




response 192488

| Ion    | Exp%   | Act%   |
|--------|--------|--------|
| 265.70 | 100.00 | 100.00 |
| 268.00 | 63.80  | 63.01  |
| 264.00 | 65.30  | 61.77  |
| 0.00   | 0.00   | 0.00   |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF121323\  
Data File : BF136530.D  
Acq On : 13 Dec 2023 10:50  
Operator : CG\JU  
Sample : DFTPP  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
DFTPP

Quant Time: Dec 13 12:05:01 2023  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_F\Methods\8270-BF120523.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Wed Dec 06 09:40:27 2023  
Response via : Initial Calibration



TIC: BF136530.D\data.ms

(77) Benzidine

12.674min (+ 0.005) 1638.06 ng

response 932869

| Ion    | Exp%   | Act%   |
|--------|--------|--------|
| 184.00 | 100.00 | 100.00 |
| 185.00 | 15.00  | 13.71  |
| 183.00 | 11.70  | 12.44  |
| 0.00   | 0.00   | 0.00   |

### DDT Breakdown

| Date          | Instrument Name  | DFTPP Data File    |
|---------------|------------------|--------------------|
| 12/13/2023    | BNA_F            | <u>BF136530.D</u>  |
|               |                  |                    |
|               |                  |                    |
| Compound Name | Response         | Retention Time     |
|               |                  |                    |
| DDT           | 484892           | 13.498             |
| DDD           | 7917             | 13.221             |
| DDE           | 732              | 12.857             |
|               |                  |                    |
|               |                  |                    |
| SUM(DDD+DDE)  | SUM(DDT+DDD+DDE) | % Breakdown Of DDT |
|               |                  |                    |
| 8649          | 493541           | 1.75               |
|               |                  |                    |

Instrument :  
BNA\_F  
ClientSampleId :  
DFTPP