

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF121418\
 Data File : BF111418.D
 Acq On : 14 Dec 2018 15:00
 Operator : JU/SJ
 Sample : J6319-03
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 057_BT-2-1

Quant Time: Dec 14 15:57:38 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF121418.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 14 13:26:17 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.79	152	157673	20.00	ng	0.00
21) Naphthalene-d8	8.07	136	645037	20.00	ng	0.00
39) Acenaphthene-d10	9.83	164	339595	20.00	ng	0.00
64) Phenanthrene-d10	11.31	188	612066	20.00	ng	0.00
76) Chrysene-d12	13.94	240	420723	20.00	ng	0.00
87) Perylene-d12	15.38	264	346027	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.42	112	1079371	113.92	ng	0.00
7) Phenol-d6	6.43	99	1411164	111.96	ng	0.00
23) Nitrobenzene-d5	7.35	82	870676	80.38	ng	0.00
42) 2,4,6-Tribromophenol	10.62	330	324204	106.57	ng	0.00
45) 2-Fluorobiphenyl	9.15	172	1457097	66.28	ng	0.00
79) Terphenyl-d14	12.90	244	1392096	77.70	ng	0.00
Target Compounds						
10) Phenol	6.45	94	95398	6.795	ng	97
38) 1-Methylnaphthalene	8.89	142	166280	8.431	ng	99
50) Dimethylphthalate	9.54	163	116277	4.754	ng	97
78) Pyrene	12.75	202	62780	2.116	ng	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF121418\
Data File : BF111418.D
Acq On : 14 Dec 2018 15:00
Operator : JU/SJ
Sample : J6319-03
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
057_BT-2-1

Quant Time: Dec 14 15:57:38 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF121418.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 14 13:26:17 2018
Response via : Initial Calibration

