

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF121418\
 Data File : BF111416.D
 Acq On : 14 Dec 2018 14:01
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Quant Time: Dec 14 15:41:03 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF121418.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 14 13:26:17 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.79	152	190986	20.00	ng	0.00
21) Naphthalene-d8	8.07	136	799796	20.00	ng	0.00
39) Acenaphthene-d10	9.83	164	383783	20.00	ng	0.00
64) Phenanthrene-d10	11.32	188	694558	20.00	ng	0.00
76) Chrysene-d12	13.95	240	533712	20.00	ng	0.00
87) Perylene-d12	15.39	264	403456	20.00	ng	0.00

System Monitoring Compounds

5) 2-Fluorophenol	5.42	112	926719	80.74	ng	0.00
7) Phenol-d6	6.44	99	1169715	76.62	ng	0.00
23) Nitrobenzene-d5	7.36	82	1098108	81.76	ng	0.00
42) 2,4,6-Tribromophenol	10.62	330	270376	78.65	ng	0.00
45) 2-Fluorobiphenyl	9.16	172	1986535	79.95	ng	0.00
79) Terphenyl-d14	12.90	244	1764996	77.66	ng	0.00

Target Compounds

						Qvalue
2) 1,4-Dioxane	2.50	88	209664	37.824	ng	97
3) Pyridine	3.23	79	549083	39.440	ng	92
4) n-Nitrosodimethylamine	3.17	42	257110	40.658	ng	88
6) Aniline	6.46	93	731185	38.611	ng	# 87
8) 2-Chlorophenol	6.58	128	546160	39.827	ng	99
9) Benzaldehyde	6.34	77	334976	36.175	ng	95
10) Phenol	6.45	94	659397	38.773	ng	# 73
11) bis(2-Chloroethyl)ether	6.53	93	518314	38.878	ng	98
12) 1,3-Dichlorobenzene	6.73	146	582062	38.559	ng	98
13) 1,4-Dichlorobenzene	6.81	146	601355	39.247	ng	97
14) 1,2-Dichlorobenzene	6.97	146	558182	38.723	ng	98
15) Benzyl Alcohol	6.94	79	452725	38.973	ng	96
16) 2,2'-oxybis(1-Chloropropan	7.07	45	1020728	39.213	ng	100
17) 2-Methylphenol	7.06	107	424689	38.481	ng	97
18) Hexachloroethane	7.31	117	215965	37.874	ng	91
19) n-Nitroso-di-n-propylamine	7.21	70	380623	37.790	ng	93
20) 3+4-Methylphenols	7.21	107	532695	38.479	ng	# 81
22) Acetophenone	7.20	105	731180	40.953	ng	# 94
24) Nitrobenzene	7.37	77	560363	40.899	ng	99
25) Isophorone	7.62	82	899397	39.577	ng	99
26) 2-Nitrophenol	7.69	139	288825	40.649	ng	99
27) 2,4-Dimethylphenol	7.73	122	423263	41.090	ng	96
28) bis(2-Chloroethoxy)methane	7.83	93	622529	39.681	ng	98
29) 2,4-Dichlorophenol	7.94	162	451582	41.067	ng	95
30) 1,2,4-Trichlorobenzene	8.02	180	458263	39.810	ng	98
31) Naphthalene	8.10	128	1488516	39.791	ng	97
32) Benzoic acid	7.86	122	303567	34.103	ng	96
33) 4-Chloroaniline	8.15	127	652677	39.237	ng	99
34) Hexachlorobutadiene	8.22	225	254750	40.169	ng	97
35) Caprolactam	8.52	113	161292	39.507	ng	98
36) 4-Chloro-3-methylphenol	8.63	107	478516	39.534	ng	98
37) 2-Methylnaphthalene	8.79	142	957010	37.807	ng	98
38) 1-Methylnaphthalene	8.89	142	929375	38.004	ng	97
40) 1,2,4,5-Tetrachlorobenzene	8.96	216	413191	39.258	ng	99

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF121418\
 Data File : BF111416.D
 Acq On : 14 Dec 2018 14:01
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Quant Time: Dec 14 15:41:03 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF121418.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 14 13:26:17 2018
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Hexachlorocyclopentadiene	8.95	237	223852	41.426	ng	97
43) 2,4,6-Trichlorophenol	9.07	196	299664	40.228	ng	96
44) 2,4,5-Trichlorophenol	9.12	196	317588	40.065	ng	94
46) 1,1'-Biphenyl	9.26	154	1285077	39.518	ng	94
47) 2-Chloronaphthalene	9.28	162	976021	39.486	ng	95
48) 2-Nitroaniline	9.37	65	323148	40.264	ng	93
49) Acenaphthylene	9.69	152	1460474	39.806	ng	96
50) Dimethylphthalate	9.56	163	1078607	39.024	ng	98
51) 2,6-Dinitrotoluene	9.62	165	251891	40.488	ng	99
52) Acenaphthene	9.87	154	911347	39.299	ng	97
53) 3-Nitroaniline	9.79	138	296470	39.233	ng	91
54) 2,4-Dinitrophenol	9.89	184	93332	39.432	ng	88
55) Dibenzofuran	10.04	168	1325284	39.372	ng	96
56) 4-Nitrophenol	9.95	139	214470	41.011	ng	99
57) 2,4-Dinitrotoluene	10.02	165	329269	40.324	ng	98
58) Fluorene	10.38	166	976778	40.007	ng	99
59) 2,3,4,6-Tetrachlorophenol	10.16	232	242613	39.174	ng	98
60) Diethylphthalate	10.26	149	1103873	39.428	ng	98
61) 4-Chlorophenyl-phenylether	10.37	204	475960	39.845	ng	96
62) 4-Nitroaniline	10.40	138	291351	38.948	ng	98
63) Azobenzene	10.53	77	1080092	39.132	ng	97
65) 4,6-Dinitro-2-methylphenol	10.43	198	150412	38.360	ng	91
66) n-Nitrosodiphenylamine	10.49	169	935569	39.061	ng	98
67) 4-Bromophenyl-phenylether	10.86	248	293660	39.228	ng	95
68) Hexachlorobenzene	10.93	284	301179	39.112	ng	# 91
69) Atrazine	11.02	200	276541	39.951	ng	96
70) Pentachlorophenol	11.12	266	189758	40.177	ng	96
71) Phenanthrene	11.34	178	1386187	38.732	ng	95
72) Anthracene	11.39	178	1444267	39.460	ng	96
73) Carbazole	11.55	167	1482409	39.167	ng	96
74) Di-n-butylphthalate	11.88	149	1676912	39.782	ng	96
75) Fluoranthene	12.53	202	1405644	40.147	ng	97
77) Benzidine	12.65	184	743087	37.875	ng	100
78) Pyrene	12.75	202	1457110	38.724	ng	95
80) Butylbenzylphthalate	13.37	149	712508	38.847	ng	94
81) Benzo(a)anthracene	13.94	228	1109300	38.225	ng	98
82) 3,3'-Dichlorobenzidine	13.90	252	444279	37.659	ng	99
83) Chrysene	13.98	228	1158877	37.908	ng	96
84) Bis(2-ethylhexyl)phthalate	13.94	149	919392	39.246	ng	99
85) Di-n-octyl phthalate	14.55	149	1521339	38.500	ng	99
86) Indeno(1,2,3-cd)pyrene	16.80	276	719800	32.623	ng	96
88) Benzo(b)fluoranthene	14.97	252	917132	38.990	ng	97
89) Benzo(k)fluoranthene	15.00	252	943331	41.971	ng	100
90) Benzo(a)pyrene	15.33	252	838327	39.523	ng	96
91) Dibenzo(a,h)anthracene	16.82	278	613588	36.671	ng	# 95
92) Benzo(g,h,i)perylene	17.23	276	595151	36.237	ng	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF121418\
Data File : BF111416.D
Acq On : 14 Dec 2018 14:01
Operator : JU/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Quant Time: Dec 14 15:41:03 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF121418.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 14 13:26:17 2018
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 20.000 ng

RT: 6.79 min Scan# 800

Delta R.T. 0.00 min

Lab File: BF111416.D

Acq: 14 Dec 2018 14:01

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tot Ion:152 Resp: 190986

Ion Ratio Lower Upper

152 100

150 160.2 126.6 189.8




115 64.7 50.7 76.1



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 115.00 (114.70 to 115.70): E



#2

1,4-Dioxane

Concen: 37.824 ng

RT: 2.50 min Scan# 70

Delta R.T. -0.01 min

Lab File: BF111416.D

Acq: 14 Dec 2018 14:01


Tgt Ion: 88 Resp: 209664

Ion Ratio Lower Upper

88 100

58 87.8 68.9 103.3



43 35.5 25.8 38.6




Abundance Ion 88.00 (87.70 to 88.70): BF1

Ion 58.00 (57.70 to 58.70): BF1

Ion 43.00 (42.70 to 43.70): BF1









#3
 Pvridine
 Concen: 39.440 ng
 RT: 3.23 min Scan# 194
 Delta R.T. -0.01 min
 Lab File: BF111416.D
 Acq: 14 Dec 2018 14:01




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Tot Ion: 79 Resp: 549083
 Ion Ratio Lower Upper
 79 100
 52 94.3 68.3 102.5
 51 44.5 32.6 49.0



#4
 n-Nitrosodimethylamine
 Concen: 40.658 ng
 RT: 3.17 min Scan# 184
 Delta R.T. -0.01 min
 Lab File: BF111416.D
 Acq: 14 Dec 2018 14:01

Tgt Ion: 42 Resp: 257110
 Ion Ratio Lower Upper
 42 100
 74 111.9 101.5 152.3
 44 7.3 6.0 9.0





#5

2-Fluorophenol

Concen: 80.745 ng

RT: 5.42 min Scan# 566

Delta R.T. 0.00 min

Lab File: BF111416.D

Acq: 14 Dec 2018 14:01

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tot Ion:112 Resp: 926719

Ion Ratio Lower Upper

112 100

64 58.2 51.8 77.6




63 30.1 26.8 40.2



Abundance Ion 112.00 (111.70 to 112.70): E

Ion 64.00 (63.70 to 64.70): BF1

Ion 63.00 (62.70 to 63.70): BF1



#6

Aniline

Concen: 38.611 ng

RT: 6.46 min Scan# 743

Delta R.T. 0.00 min

Lab File: BF111416.D

Acq: 14 Dec 2018 14:01


Tgt Ion: 93 Resp: 731185

Ion Ratio Lower Upper

93 100

66 49.0 33.2 49.8



65 28.5 17.0 25.4#




Abundance Ion 93.00 (92.70 to 93.70): BF1

Ion 66.00 (65.70 to 66.70): BF1

Ion 65.00 (64.70 to 65.70): BF1





#7

Phenol-d6

Concen: 76.619 ng

RT: 6.44 min Scan# 740

Delta R.T. 0.00 min

Lab File: BF111416.D

Acq: 14 Dec 2018 14:01

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tot Ion: 99 Resp: 1169715

Ion Ratio Lower Upper

99 100

42 22.0 17.4 26.0




71 32.0 26.1 39.1



Abundance Ion 99.00 (98.70 to 99.70): BF1

Ion 42.00 (41.70 to 42.70): BF1

Ion 71.00 (70.70 to 71.70): BF1



#8

2-Chlorophenol

Concen: 39.827 ng

RT: 6.58 min Scan# 764

Delta R.T. -0.01 min

Lab File: BF111416.D

Acq: 14 Dec 2018 14:01


Tgt Ion:128 Resp: 546160

Ion Ratio Lower Upper

128 100

130 31.8 12.7 52.7



64 50.3 30.0 70.0




Abundance Ion 128.00 (127.70 to 128.70): E

Ion 130.00 (129.70 to 130.70): E

Ion 64.00 (63.70 to 64.70): BF1





#9

Benzaldehyde

Concen: 36.175 ng

RT: 6.34 min Scan# 723

Delta R.T. 0.00 min

Lab File: BF111416.D

Acq: 14 Dec 2018 14:01

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tot Ion: 77 Resp: 334976

Ion Ratio Lower Upper




77 100

105 106.6 81.4 121.4

106 102.5 77.9 117.9



Abundance Ion 77.00 (76.70 to 77.70): BF1
Ion 105.00 (104.70 to 105.70): BF1
Ion 106.00 (105.70 to 106.70): BF1



#10

Phenol

Concen: 38.773 ng

RT: 6.45 min Scan# 742

Delta R.T. 0.00 min

Lab File: BF111416.D

Acq: 14 Dec 2018 14:01


Tgt Ion: 94 Resp: 659397

Ion Ratio Lower Upper



94 100


65 41.4 11.7 51.7

66 62.6 21.0 61.0#



Abundance Ion 94.00 (93.70 to 94.70): BF1
Ion 65.00 (64.70 to 65.70): BF1
Ion 66.00 (65.70 to 66.70): BF1









#11
 bis(2-Chloroethyl)ether
 Concen: 38.878 ng
 RT: 6.53 min Scan# 755
 Delta R.T. -0.01 min
 Lab File: BF111416.D
 Acq: 14 Dec 2018 14:01




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Tot Ion:	93	Resp:	518314
Ion	Ratio	Lower	Upper
93	100		
63	82.3	60.0	100.0
95	32.7	13.4	53.4



#12
 1,3-Dichlorobenzene
 Concen: 38.559 ng
 RT: 6.73 min Scan# 790
 Delta R.T. 0.00 min
 Lab File: BF111416.D
 Acq: 14 Dec 2018 14:01

Tgt Ion:	146	Resp:	582062
Ion	Ratio	Lower	Upper
146	100		
148	63.7	52.1	78.1
75	30.9	23.9	35.9





#13

1,4-Dichlorobenzene

Concen: 39.247 ng

RT: 6.81 min Scan# 803

Delta R.T. 0.00 min

Lab File: BF111416.D

Acq: 14 Dec 2018 14:01

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tot Ion:146 Resp: 601355

Ion Ratio Lower Upper

146 100

148 63.9 52.8 79.2




111 43.8 33.4 50.0



Abundance Ion 146.00 (145.70 to 146.70): E

Ion 148.00 (147.70 to 148.70): E

Ion 111.00 (110.70 to 111.70): E



#14

1,2-Dichlorobenzene

Concen: 38.723 ng

RT: 6.97 min Scan# 830

Delta R.T. 0.00 min

Lab File: BF111416.D

Acq: 14 Dec 2018 14:01


Tgt Ion:146 Resp: 558182

Ion Ratio Lower Upper

146 100

148 65.4 51.9 77.9



111 43.1 36.2 54.4




Abundance Ion 146.00 (145.70 to 146.70): E

Ion 148.00 (147.70 to 148.70): E

Ion 111.00 (110.70 to 111.70): E





#15

Benzyl Alcohol

Concen: 38.973 ng

RT: 6.94 min Scan# 825

Delta R.T. 0.00 min

Lab File: BF111416.D

Acq: 14 Dec 2018 14:01

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040





Tot Ion: 79 Resp: 452725

Ion Ratio Lower Upper

79 100

108 82.2 69.4 104.2

77 63.4 49.7 74.5



#16

2,2'-oxybis(1-chloropropane)

Concen: 39.213 ng

RT: 7.07 min Scan# 848

Delta R.T. 0.00 min

Lab File: BF111416.D

Acq: 14 Dec 2018 14:01




Tgt Ion: 45 Resp: 1020728


Ion Ratio Lower Upper

45 100

77 11.2 0.0 31.5

79 8.9 0.0 28.9





#17

2-Methylphenol

Concen: 38.481 ng

RT: 7.06 min Scan# 845

Delta R.T. 0.00 min

Lab File: BF111416.D

Acq: 14 Dec 2018 14:01

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

Tot Ion:107 Resp: 424689


Ion Ratio Lower Upper

107 100

108 113.7 91.0 136.4

77 46.6 33.5 50.3

79 44.5 33.3 49.9






Abundance Ion 107.00 (106.70 to 107.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 77.00 (76.70 to 77.70): BF1

Ion 79.00 (78.70 to 79.70): BF1



#18

Hexachloroethane

Concen: 37.874 ng

RT: 7.31 min Scan# 888

Delta R.T. 0.00 min

Lab File: BF111416.D

Acq: 14 Dec 2018 14:01


Tgt Ion:117 Resp: 215965

Ion Ratio Lower Upper

117 100

119 96.6 75.4 113.0



201 91.8 62.4 93.6




Abundance Ion 117.00 (116.70 to 117.70): E

Ion 119.00 (118.70 to 119.70): E

Ion 201.00 (200.70 to 201.70): E






#19
n-Nitroso-di-n-propylamine
Concen: 37.790 ng
RT: 7.21 min Scan# 871
Delta R.T. -0.01 min
Lab File: BF111416.D
Acq: 14 Dec 2018 14:01

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tot Ion:	70	Resp:	380623
Ion	Ratio	Lower	Upper
70	100		
42	73.4	53.2	79.8
101	10.0	8.2	12.4
130	20.3	18.1	27.1





Abundance

Ion 70.00 (69.70 to 70.70): BF1

Ion 42.00 (41.70 to 42.70): BF1

Ion 101.00 (100.70 to 101.70): E

Ion 130.00 (129.70 to 130.70): E



#20
3+4-Methylphenols
Concen: 38.479 ng
RT: 7.21 min Scan# 871
Delta R.T. 0.00 min
Lab File: BF111416.D
Acq: 14 Dec 2018 14:01

Tgt Ion:	107	Resp:	532695
Ion	Ratio	Lower	Upper
107	100		
108	93.9	72.1	112.1
77	72.9	94.1	134.1#
79	29.5	9.8	49.8





Abundance


Ion 107.00 (106.70 to 107.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 77.00 (76.70 to 77.70): BF1

Ion 79.00 (78.70 to 79.70): BF1






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.07 min Scan# 1018
Delta R.T. -0.01 min
Lab File: BF111416.D
Acq: 14 Dec 2018 14:01

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tot Ion:	136	Resp:	799796
Ion	Ratio	Lower	Upper
136	100		
137	11.4	9.1	13.7
54	10.3	7.7	11.5
68	5.1	4.9	7.3






Abundance

Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E


Ion 54.00 (53.70 to 54.70): BF1

Ion 68.00 (67.70 to 68.70): BF1



#22
Acetophenone
Concen: 40.953 ng
RT: 7.20 min Scan# 870
Delta R.T. 0.00 min
Lab File: BF111416.D
Acq: 14 Dec 2018 14:01

Tgt Ion:	105	Resp:	731180
Ion	Ratio	Lower	Upper
105	100		
71	4.9	3.1	4.7#
51	39.9	36.6	54.8
120	22.1	18.2	27.4





Abundance


Ion 105.00 (104.70 to 105.70): E

Ion 71.00 (70.70 to 71.70): BF1

Ion 51.00 (50.70 to 51.70): BF1

Ion 120.00 (119.70 to 120.70): E









#23
Nitrobenzene-d5
Concen: 81.759 ng
RT: 7.36 min Scan# 896
Delta R.T. 0.00 min
Lab File: BF111416.D
Acq: 14 Dec 2018 14:01




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tot Ion: 82 Resp: 1098108
Ion Ratio Lower Upper
82 100
128 45.5 37.4 56.2
54 59.6 47.6 71.4



#24
Nitrobenzene
Concen: 40.899 ng
RT: 7.37 min Scan# 899
Delta R.T. 0.00 min
Lab File: BF111416.D
Acq: 14 Dec 2018 14:01

Tgt Ion: 77 Resp: 560363
Ion Ratio Lower Upper
77 100
123 47.7 37.2 55.8
65 14.4 11.5 17.3





#25

Isophorone

Concen: 39.577 ng

RT: 7.62 min Scan# 940

Delta R.T. 0.00 min

Lab File: BF111416.D

Acq: 14 Dec 2018 14:01

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tot Ion: 82 Resp: 899397

Ion Ratio Lower Upper

82 100

95 6.6 5.4 8.2




138 18.6 15.1 22.7



Abundance Ion 82.00 (81.70 to 82.70): BF1

Ion 95.00 (94.70 to 95.70): BF1

Ion 138.00 (137.70 to 138.70): BF1



#26

2-Nitrophenol

Concen: 40.649 ng

RT: 7.69 min Scan# 953

Delta R.T. 0.00 min

Lab File: BF111416.D

Acq: 14 Dec 2018 14:01


Tgt Ion: 139 Resp: 288825

Ion Ratio Lower Upper

139 100

109 25.7 20.2 30.4



65 49.6 40.7 61.1




Abundance Ion 139.00 (138.70 to 139.70): E

Ion 109.00 (108.70 to 109.70): E

Ion 65.00 (64.70 to 65.70): BF1









#27
2,4-Dimethylphenol
Concen: 41.090 ng
RT: 7.73 min Scan# 960
Delta R.T. -0.01 min
Lab File: BF111416.D
Acq: 14 Dec 2018 14:01




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tot Ion:122 Resp: 423263
Ion Ratio Lower Upper
122 100
107 112.9 85.5 128.3
121 57.5 46.6 70.0



#28
bis(2-Chloroethoxy)methane
Concen: 39.681 ng
RT: 7.83 min Scan# 976
Delta R.T. 0.00 min
Lab File: BF111416.D
Acq: 14 Dec 2018 14:01

Tgt Ion: 93 Resp: 622529
Ion Ratio Lower Upper
93 100
95 32.6 27.1 40.7
123 11.8 9.9 14.9









#29
 2,4-Dichlorophenol
 Concen: 41.067 ng
 RT: 7.94 min Scan# 995
 Delta R.T. 0.00 min
 Lab File: BF111416.D
 Acq: 14 Dec 2018 14:01

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Tot Ion	Ratio	Lower	Upper
162	100		
164	62.6	44.4	84.4
98	35.0	20.3	60.3




Abundance Ion 162.00 (161.70 to 162.70): E
 Ion 164.00 (163.70 to 164.70): E
 Ion 98.00 (97.70 to 98.70): BF1






#30
 1,2,4-Trichlorobenzene
 Concen: 39.810 ng
 RT: 8.02 min Scan# 1009
 Delta R.T. 0.00 min
 Lab File: BF111416.D
 Acq: 14 Dec 2018 14:01

Tgt Ion	Ratio	Lower	Upper
180	100		
182	92.9	75.6	113.4
145	31.6	26.7	40.1



Abundance Ion 180.00 (179.70 to 180.70): E
 Ion 182.00 (181.70 to 182.70): E
 Ion 145.00 (144.70 to 145.70): E









#31
Naphthalene
Concen: 39.791 ng
RT: 8.10 min Scan# 1022
Delta R.T. 0.00 min
Lab File: BF111416.D
Acq: 14 Dec 2018 14:01

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tot Ion:128 Resp: 1488516
Ion Ratio Lower Upper
128 100
129 11.5 9.8 14.8
127 13.2 12.0 18.0




Abundance Ion 128.00 (127.70 to 128.70): E
Ion 129.00 (128.70 to 129.70): E
Ion 127.00 (126.70 to 127.70): E






#32
Benzoic acid
Concen: 34.103 ng
RT: 7.86 min Scan# 981
Delta R.T. -0.01 min
Lab File: BF111416.D
Acq: 14 Dec 2018 14:01

Tgt Ion:122 Resp: 303567
Ion Ratio Lower Upper
122 100
105 122.6 97.0 137.0
77 83.2 65.7 105.7



Abundance Ion 122.00 (121.70 to 122.70): E
Ion 105.00 (104.70 to 105.70): E
Ion 77.00 (76.70 to 77.70): BF1









#33
4-Chloroaniline
Concen: 39.237 ng
RT: 8.15 min Scan# 1030
Delta R.T. 0.00 min
Lab File: BF111416.D
Acq: 14 Dec 2018 14:01




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tot Ion:	127	Resp:	652677
Ion	Ratio	Lower	Upper
127	100		
129	32.5	26.6	39.8
65	30.9	25.7	38.5
92	18.9	15.4	23.2



#34
Hexachlorobutadiene
Concen: 40.169 ng
RT: 8.22 min Scan# 1043
Delta R.T. 0.00 min
Lab File: BF111416.D
Acq: 14 Dec 2018 14:01

Tgt Ion:	225	Resp:	254750
Ion	Ratio	Lower	Upper
225	100		
223	62.2	50.7	76.1
227	62.8	52.6	78.8









#35
Caprolactam
Concen: 39.507 ng
RT: 8.52 min Scan# 1094
Delta R.T. -0.01 min
Lab File: BF111416.D
Acq: 14 Dec 2018 14:01

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tot Ion:113 Resp: 161292
Ion Ratio Lower Upper
113 100
55 199.5 184.1 224.1
56 145.7 126.0 166.0




Abundance Ion 113.00 (112.70 to 113.70): E
Ion 55.00 (54.70 to 55.70): BF1
Ion 56.00 (55.70 to 56.70): BF1






#36
4-Chloro-3-methylphenol
Concen: 39.534 ng
RT: 8.63 min Scan# 1113
Delta R.T. -0.01 min
Lab File: BF111416.D
Acq: 14 Dec 2018 14:01

Tgt Ion:107 Resp: 478516
Ion Ratio Lower Upper
107 100
144 24.3 19.8 29.6
142 75.7 62.3 93.5



Abundance Ion 107.00 (106.70 to 107.70): E
Ion 144.00 (143.70 to 144.70): E
Ion 142.00 (141.70 to 142.70): E





#37

2-Methylnaphthalene

Concen: 37.807 ng

RT: 8.79 min Scan# 1140

Delta R.T. 0.00 min

Lab File: BF111416.D

Acq: 14 Dec 2018 14:01

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tot Ion:142 Resp: 957010

Ion Ratio Lower Upper

142 100

141 89.0 70.6 105.8




115 31.2 27.0 40.6



Abundance Ion 142.00 (141.70 to 142.70): E

Ion 141.00 (140.70 to 141.70): E

Ion 115.00 (114.70 to 115.70): E



#38

1-Methylnaphthalene

Concen: 38.004 ng

RT: 8.89 min Scan# 1156

Delta R.T. -0.01 min

Lab File: BF111416.D

Acq: 14 Dec 2018 14:01


Tgt Ion:142 Resp: 929375

Ion Ratio Lower Upper

142 100

141 90.2 74.2 111.4



116 3.7 2.9 4.3




Abundance Ion 142.00 (141.70 to 142.70): E

Ion 141.00 (140.70 to 141.70): E

Ion 116.00 (115.70 to 116.70): E





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.83 min Scan# 1317

Delta R.T. 0.00 min

Lab File: BF111416.D

Acq: 14 Dec 2018 14:01

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tot Ion:164 Resp: 383783

Ion Ratio Lower Upper

164 100

162 102.4 81.4 122.0




160 45.6 35.7 53.5



Abundance Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E



#40

1,2,4,5-Tetrachlorobenzene

Concen: 39.258 ng

RT: 8.96 min Scan# 1168

Delta R.T. 0.00 min

Lab File: BF111416.D

Acq: 14 Dec 2018 14:01

Tgt Ion:216 Resp: 413191


Ion Ratio Lower Upper

216 100

214 80.3 63.8 95.8

179 22.4 17.9 26.9

108 20.9 18.4 27.6






Abundance Ion 216.00 (215.70 to 216.70): E

Ion 214.00 (213.70 to 214.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 108.00 (107.70 to 108.70): E





#41

Hexachlorocyclopentadiene

Concen: 41.426 ng

RT: 8.95 min Scan# 1167

Delta R.T. 0.00 min

Lab File: BF111416.D

Acq: 14 Dec 2018 14:01

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tot Ion:237 Resp: 223852

Ion Ratio Lower Upper




237 100

235 65.9 43.1 83.1

272 12.9 0.0 33.0



Abundance Ion 236.80 (236.50 to 237.50): E
Ion 234.90 (234.60 to 235.60): E
Ion 271.75 (271.45 to 272.45): E



#42

2,4,6-Tribromophenol

Concen: 78.646 ng

RT: 10.62 min Scan# 1451

Delta R.T. 0.00 min

Lab File: BF111416.D

Acq: 14 Dec 2018 14:01


Tgt Ion:330 Resp: 270376

Ion Ratio Lower Upper



330 100


332 97.8 76.0 114.0

141 40.9 31.0 46.6



Abundance Ion 329.80 (329.50 to 330.50): E
Ion 331.80 (331.50 to 332.50): E
Ion 141.00 (140.70 to 141.70): E









#43
 2,4,6-Trichlorophenol
 Concen: 40.228 ng
 RT: 9.07 min Scan# 1187
 Delta R.T. 0.00 min
 Lab File: BF111416.D
 Acq: 14 Dec 2018 14:01




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Tot Ion:196 Resp: 299664
 Ion Ratio Lower Upper
 196 100
 198 91.8 76.3 114.5
 200 28.9 24.9 37.3



#44
 2,4,5-Trichlorophenol
 Concen: 40.065 ng
 RT: 9.12 min Scan# 1195
 Delta R.T. 0.00 min
 Lab File: BF111416.D
 Acq: 14 Dec 2018 14:01

Tgt Ion:196 Resp: 317588
 Ion Ratio Lower Upper
 196 100
 198 93.9 72.9 109.3
 97 49.2 45.3 67.9
 132 29.0 25.5 38.3









#45
2-Fluorobiphenyl
Concen: 79.953 ng
RT: 9.16 min Scan# 1202
Delta R.T. 0.00 min
Lab File: BF111416.D
Acq: 14 Dec 2018 14:01

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tot Ion:172 Resp: 1986535
Ion Ratio Lower Upper
172 100
171 36.3 33.0 49.4
170 24.3 22.3 33.5




Abundance Ion 172.00 (171.70 to 172.70): E
Ion 171.00 (170.70 to 171.70): E
Ion 170.00 (169.70 to 170.70): E






#46
1,1'-Biphenyl
Concen: 39.518 ng
RT: 9.26 min Scan# 1219
Delta R.T. 0.00 min
Lab File: BF111416.D
Acq: 14 Dec 2018 14:01

Tgt Ion:154 Resp: 1285077
Ion Ratio Lower Upper
154 100
153 40.8 23.7 63.7
76 11.3 0.0 35.0



Abundance Ion 154.00 (153.70 to 154.70): E
Ion 153.00 (152.70 to 153.70): E
Ion 76.00 (75.70 to 76.70): BF1





#47

2-Chloronaphthalene

Concen: 39.486 ng

RT: 9.28 min Scan# 1223

Delta R.T. 0.00 min

Lab File: BF111416.D

Acq: 14 Dec 2018 14:01

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tot Ion:162 Resp: 976021

Ion Ratio Lower Upper




162 100

127 40.0 35.0 52.4

164 33.5 28.4 42.6



Abundance Ion 162.00 (161.70 to 162.70): E
Ion 127.00 (126.70 to 127.70): E
Ion 164.00 (163.70 to 164.70): E



#48

2-Nitroaniline

Concen: 40.264 ng

RT: 9.37 min Scan# 1239

Delta R.T. 0.00 min

Lab File: BF111416.D

Acq: 14 Dec 2018 14:01


Tgt Ion: 65 Resp: 323148

Ion Ratio Lower Upper



65 100


92 68.9 53.5 80.3

138 111.8 80.9 121.3



Abundance Ion 65.00 (64.70 to 65.70): BF1
Ion 92.00 (91.70 to 92.70): BF1
Ion 138.00 (137.70 to 138.70): E





#49

Acenaphthylene

Concen: 39.806 ng

RT: 9.69 min Scan# 1293

Delta R.T. 0.00 min

Lab File: BF111416.D

Acq: 14 Dec 2018 14:01

Instrument :


BNA_F

ClientSampleId :

SSTDCCC040

Tot Ion:152 Resp: 1460474




Ion	Ratio	Lower	Upper
152	100		
151	20.5	18.0	27.0
153	13.5	12.3	18.5



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 151.00 (150.70 to 151.70): E

Ion 153.00 (152.70 to 153.70): E



#50

Dimethylphthalate

Concen: 39.024 ng

RT: 9.56 min Scan# 1270


Delta R.T. 0.00 min

Lab File: BF111416.D

Acq: 14 Dec 2018 14:01

Tgt Ion:163 Resp: 1078607



Ion	Ratio	Lower	Upper
163	100		
194	3.8	3.0	4.6
164	10.2	8.9	13.3




Abundance Ion 163.00 (162.70 to 163.70): E

Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E









#51
 2,6-Dinitrotoluene
 Concen: 40.488 ng
 RT: 9.62 min Scan# 1280
 Delta R.T. 0.00 min
 Lab File: BF111416.D
 Acq: 14 Dec 2018 14:01




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Tot Ion	Ratio	Lower	Upper
165	100		
63	50.2	40.5	60.7
89	48.6	40.0	60.0



#52
 Acenaphthene
 Concen: 39.299 ng
 RT: 9.87 min Scan# 1323
 Delta R.T. 0.00 min
 Lab File: BF111416.D
 Acq: 14 Dec 2018 14:01

Tgt Ion	Ratio	Lower	Upper
154	100		
153	113.2	87.8	131.8
152	53.5	43.4	65.2









#53
3-Nitroaniline
Concen: 39.233 ng
RT: 9.79 min Scan# 1309
Delta R.T. 0.00 min
Lab File: BF111416.D
Acq: 14 Dec 2018 14:01

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tot Ion:138 Resp: 296470
Ion Ratio Lower Upper
138 100
108 9.8 8.2 12.4
92 106.1 93.4 140.2




Abundance Ion 138.00 (137.70 to 138.70): E
Ion 108.00 (107.70 to 108.70): E
Ion 92.00 (91.70 to 92.70): BF1






#54
2,4-Dinitrophenol
Concen: 39.432 ng
RT: 9.89 min Scan# 1326
Delta R.T. 0.00 min
Lab File: BF111416.D
Acq: 14 Dec 2018 14:01

Tgt Ion:184 Resp: 93332
Ion Ratio Lower Upper
184 100
63 65.7 62.3 93.5
154 61.5 56.2 84.2



Abundance Ion 184.00 (183.70 to 184.70): E
Ion 63.00 (62.70 to 63.70): BF1
Ion 154.00 (153.70 to 154.70): E









#55
Dibenzofuran
Concen: 39.372 ng
RT: 10.04 min Scan# 1352
Delta R.T. 0.00 min
Lab File: BF111416.D
Acq: 14 Dec 2018 14:01

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tot Ion:168 Resp: 1325284
Ion Ratio Lower Upper
168 100
139 37.7 32.6 49.0
169 13.5 11.8 17.6




Abundance Ion 168.00 (167.70 to 168.70): E
2000000 Ion 139.00 (138.70 to 139.70): E
Ion 169.00 (168.70 to 169.70): E






#56
4-Nitrophenol
Concen: 41.011 ng
RT: 9.95 min Scan# 1337
Delta R.T. -0.01 min
Lab File: BF111416.D
Acq: 14 Dec 2018 14:01

Tgt Ion:139 Resp: 214470
Ion Ratio Lower Upper
139 100
109 64.1 41.8 81.8
65 99.1 79.6 119.6



Abundance Ion 139.00 (138.70 to 139.70): E
Ion 109.00 (108.70 to 109.70): E
Ion 65.00 (64.70 to 65.70): BF1









#57
2,4-Dinitrotoluene
Concen: 40.324 ng
RT: 10.02 min Scan# 1348
Delta R.T. 0.00 min
Lab File: BF111416.D
Acq: 14 Dec 2018 14:01




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tot Ion	Ratio	Lower	Upper
165	100		
63	45.9	34.6	52.0
89	71.4	56.2	84.4
182	2.5	1.8	2.6



#58
Fluorene
Concen: 40.007 ng
RT: 10.38 min Scan# 1410
Delta R.T. 0.00 min
Lab File: BF111416.D
Acq: 14 Dec 2018 14:01

Tgt Ion	Ratio	Lower	Upper
166	100		
165	97.1	78.4	117.6
167	13.4	11.2	16.8









#59
 2,3,4,6-Tetrachlorophenol
 Concen: 39.174 ng
 RT: 10.16 min Scan# 1372
 Delta R.T. 0.00 min
 Lab File: BF111416.D
 Acq: 14 Dec 2018 14:01




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Tot Ion:232 Resp: 242613
 Ion Ratio Lower Upper
 232 100
 131 52.5 40.9 61.3
 130 2.3 2.0 3.0
 166 34.3 26.8 40.2



#60
 Diethylphthalate
 Concen: 39.428 ng
 RT: 10.26 min Scan# 1389
 Delta R.T. 0.00 min
 Lab File: BF111416.D
 Acq: 14 Dec 2018 14:01

Tgt Ion:149 Resp: 1103873
 Ion Ratio Lower Upper
 149 100
 177 21.8 18.3 27.5
 150 12.3 10.6 15.8









#61
 4-Chlorophenyl-phenylether
 Concen: 39.845 ng
 RT: 10.37 min Scan# 1408
 Delta R.T. -0.01 min
 Lab File: BF111416.D
 Acq: 14 Dec 2018 14:01

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Tot Ion:204 Resp: 475960
 Ion Ratio Lower Upper
 204 100
 206 32.0 26.6 39.8
 141 70.1 52.6 78.8




Abundance Ion 204.00 (203.70 to 204.70): E
 Ion 206.00 (205.70 to 206.70): E
 Ion 141.00 (140.70 to 141.70): E






#62
 4-Nitroaniline
 Concen: 38.948 ng
 RT: 10.40 min Scan# 1413
 Delta R.T. 0.00 min
 Lab File: BF111416.D
 Acq: 14 Dec 2018 14:01

Tgt Ion:138 Resp: 291351
 Ion Ratio Lower Upper
 138 100
 92 48.2 29.9 69.9
 108 61.9 43.2 83.2



Abundance Ion 138.00 (137.70 to 138.70): E
 Ion 92.00 (91.70 to 92.70): BF1
 Ion 108.00 (107.70 to 108.70): E









#63
Azobenzene
Concen: 39.132 ng
RT: 10.53 min Scan# 1436
Delta R.T. 0.00 min
Lab File: BF111416.D
Acq: 14 Dec 2018 14:01




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tot Ion: 77 Resp: 1080092
Ion Ratio Lower Upper
77 100
182 26.0 4.9 44.9
105 23.8 6.0 46.0
51 33.2 14.7 54.7



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.32 min Scan# 1569
Delta R.T. 0.00 min
Lab File: BF111416.D
Acq: 14 Dec 2018 14:01

Tgt Ion: 188 Resp: 694558
Ion Ratio Lower Upper
188 100
94 9.3 9.6 14.4#
80 9.7 9.8 14.6#









#65
4,6-Dinitro-2-methylphenol
Concen: 38.360 ng
RT: 10.43 min Scan# 1418
Delta R.T. 0.00 min
Lab File: BF111416.D
Acq: 14 Dec 2018 14:01




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tot Ion:198 Resp: 150412
Ion Ratio Lower Upper
198 100
51 58.6 48.0 88.0
105 50.2 34.0 74.0



#66
n-Nitrosodiphenylamine
Concen: 39.061 ng
RT: 10.49 min Scan# 1429
Delta R.T. 0.00 min
Lab File: BF111416.D
Acq: 14 Dec 2018 14:01

Tgt Ion:169 Resp: 935569
Ion Ratio Lower Upper
169 100
168 66.5 54.4 81.6
167 37.1 30.5 45.7









#67
 4-Bromophenyl-phenylether
 Concen: 39.228 ng
 RT: 10.86 min Scan# 1492
 Delta R.T. 0.00 min
 Lab File: BF111416.D
 Acq: 14 Dec 2018 14:01




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Tot Ion:248 Resp: 293660
 Ion Ratio Lower Upper
 248 100
 250 94.9 77.0 115.4
 141 70.2 63.0 94.6



#68
 Hexachlorobenzene
 Concen: 39.112 ng
 RT: 10.93 min Scan# 1503
 Delta R.T. -0.01 min
 Lab File: BF111416.D
 Acq: 14 Dec 2018 14:01

Tgt Ion:284 Resp: 301179
 Ion Ratio Lower Upper
 284 100
 142 42.8 27.4 41.2#
 249 31.5 24.4 36.6









#69
Atrazine
Concen: 39.951 ng
RT: 11.02 min Scan# 1519
Delta R.T. 0.00 min
Lab File: BF111416.D
Acq: 14 Dec 2018 14:01

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tot Ion:200 Resp: 276541
Ion Ratio Lower Upper
200 100
173 26.1 10.1 50.1
215 47.6 26.9 66.9




Abundance Ion 200.00 (199.70 to 200.70): E
Ion 173.00 (172.70 to 173.70): E
Ion 215.00 (214.70 to 215.70): E






#70
Pentachlorophenol
Concen: 40.177 ng
RT: 11.12 min Scan# 1536
Delta R.T. 0.00 min
Lab File: BF111416.D
Acq: 14 Dec 2018 14:01

Tgt Ion:266 Resp: 189758
Ion Ratio Lower Upper
266 100
268 63.8 47.9 71.9
264 64.1 49.1 73.7



Abundance Ion 265.70 (265.40 to 266.40): E
Ion 268.00 (267.70 to 268.70): E
Ion 264.00 (263.70 to 264.70): E





#71

Phenanthrene

Concen: 38.732 ng

RT: 11.34 min Scan# 1573

Delta R.T. 0.00 min

Lab File: BF111416.D

Acq: 14 Dec 2018 14:01

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tot Ion:178 Resp: 1386187

Ion Ratio Lower Upper

178 100

176 19.9 18.1 27.1




179 16.2 14.0 21.0



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E



#72

Anthracene

Concen: 39.460 ng

RT: 11.39 min Scan# 1582

Delta R.T. 0.00 min

Lab File: BF111416.D

Acq: 14 Dec 2018 14:01


Tgt Ion:178 Resp: 1444267

Ion Ratio Lower Upper

178 100

176 19.1 17.2 25.8



179 15.9 13.8 20.8




Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E





#73

Carbazole

Concen: 39.167 ng

RT: 11.55 min Scan# 1608

Delta R.T. 0.00 min





Lab File: BF111416.D

Acq: 14 Dec 2018 14:01

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tot Ion:167 Resp: 1482409

Ion	Ratio	Lower	Upper
167	100		
166	22.2	19.3	28.9
139	13.6	11.9	17.9



#74

Di-n-butylphthalate

Concen: 39.782 ng

RT: 11.88 min Scan# 1665




Delta R.T. 0.00 min


Lab File: BF111416.D

Acq: 14 Dec 2018 14:01

Tgt Ion:149 Resp: 1676912

Ion	Ratio	Lower	Upper
149	100		
150	9.3	8.8	13.2
104	4.7	4.5	6.7





#75

Fluoranthene

Concen: 40.147 ng

RT: 12.53 min Scan# 1775

Delta R.T. 0.00 min

Lab File: BF111416.D

Acq: 14 Dec 2018 14:01

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tot Ion:202 Resp: 1405644

Ion Ratio Lower Upper

202 100

101 11.8 0.0 33.8




203 17.6 0.0 38.6



Abundance Ion 202.00 (201.70 to 202.70): E

2000000 Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E



#76

Chrysene-d12

Concen: 20.000 ng

RT: 13.95 min Scan# 2017

Delta R.T. 0.00 min

Lab File: BF111416.D

Acq: 14 Dec 2018 14:01


Tgt Ion:240 Resp: 533712

Ion Ratio Lower Upper

240 100

120 14.8 12.2 18.2



236 25.4 20.2 30.2




Abundance Ion 240.00 (239.70 to 240.70): E

800000 Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E





#77

Benzidine

Concen: 37.875 ng

RT: 12.65 min Scan# 1795

Delta R.T. 0.00 min

Lab File: BF111416.D

Acq: 14 Dec 2018 14:01

Instrument :


BNA_F

ClientSampleId :

SSTDCCC040

Tot Ion:184 Resp: 743087




Ion	Ratio	Lower	Upper
184	100		
185	16.4	13.0	19.6
183	11.9	9.3	13.9



Abundance Ion 184.00 (183.70 to 184.70): E

Ion 185.00 (184.70 to 185.70): E

Ion 183.00 (182.70 to 183.70): E



#78

Pyrene

Concen: 38.724 ng

RT: 12.75 min Scan# 1813


Delta R.T. -0.01 min

Lab File: BF111416.D

Acq: 14 Dec 2018 14:01

Tgt Ion:202 Resp: 1457110



Ion	Ratio	Lower	Upper
202	100		
200	20.7	19.0	28.4
203	17.8	15.2	22.8




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E





#79

Terphenyl-d14

Concen: 77.657 ng

RT: 12.90 min Scan# 1838

Delta R.T. 0.00 min

Lab File: BF111416.D

Acq: 14 Dec 2018 14:01

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tot Ion:244 Resp: 1764996

Ion Ratio Lower Upper

244 100

212 6.9 5.7 8.5




122 14.7 13.1 19.7



Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E



#80

Butylbenzylphthalate

Concen: 38.847 ng

RT: 13.37 min Scan# 1919

Delta R.T. 0.00 min

Lab File: BF111416.D

Acq: 14 Dec 2018 14:01


Tgt Ion:149 Resp: 712508

Ion Ratio Lower Upper

149 100

91 73.7 63.8 95.6



206 18.6 15.1 22.7




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 91.00 (90.70 to 91.70): BF1

Ion 206.00 (205.70 to 206.70): E





#81

Benzo(a)anthracene

Concen: 38.225 ng

RT: 13.94 min Scan# 2015

Delta R.T. -0.01 min

Lab File: BF111416.D

Acq: 14 Dec 2018 14:01

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tot Ion:228 Resp: 1109300

Ion Ratio Lower Upper

228 100

226 27.3 22.6 34.0




229 20.0 16.3 24.5



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#82

3,3'-Dichlorobenzidine

Concen: 37.659 ng

RT: 13.90 min Scan# 2009

Delta R.T. 0.00 min

Lab File: BF111416.D

Acq: 14 Dec 2018 14:01


Tgt Ion:252 Resp: 444279

Ion Ratio Lower Upper

252 100

254 66.6 52.7 79.1



126 13.5 11.1 16.7




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 254.00 (253.70 to 254.70): E

Ion 126.00 (125.70 to 126.70): E





#83

Chrvsene

Concen: 37.908 ng

RT: 13.98 min Scan# 2022

Delta R.T. 0.00 min

Lab File: BF111416.D

Acq: 14 Dec 2018 14:01

Instrument :





BNA_F

ClientSampleId :

SSTDCCC040

Tot Ion:228 Resp: 1158877

Ion	Ratio	Lower	Upper
228	100		
226	29.8	25.3	37.9
229	19.6	17.2	25.8



#84

Bis(2-ethylhexyl)phthalate

Concen: 39.246 ng

RT: 13.94 min Scan# 1015




Delta R.T. 0.00 min


Lab File: BF111416.D

Acq: 14 Dec 2018 14:01

Tgt Ion:149 Resp: 919392

Ion	Ratio	Lower	Upper
149	100		
167	27.5	22.5	33.7
279	3.7	2.6	4.0





#85

Di-n-octyl phthalate

Concen: 38.500 ng

RT: 14.55 min Scan# 2119

Delta R.T. 0.00 min

Lab File: BF111416.D

Acq: 14 Dec 2018 14:01

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040





Tot Ion:149 Resp: 1521339

Ion Ratio Lower Upper

149 100

167 1.5 1.3 1.9

43 13.3 10.9 16.3



#86

Indeno(1,2,3-cd)pyrene

Concen: 32.623 ng

RT: 16.80 min Scan# 2501

Delta R.T. 0.00 min

Lab File: BF111416.D

Acq: 14 Dec 2018 14:01




Tgt Ion:276 Resp: 719800


Ion Ratio Lower Upper

276 100

138 30.7 27.8 41.6

277 25.2 20.1 30.1





#87

Pervlene-d12

Concen: 20.000 ng

RT: 15.39 min Scan# 2261

Delta R.T. 0.00 min

Lab File: BF111416.D

Acq: 14 Dec 2018 14:01

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tot Ion:264 Resp: 403456

Ion Ratio Lower Upper

264 100

260 22.8 18.3 27.5




265 22.1 17.7 26.5



Abundance Ion 264.00 (263.70 to 264.70): E

Ion 260.00 (259.70 to 260.70): E

Ion 265.00 (264.70 to 265.70): E



#88

Benzo(b)fluoranthene

Concen: 38.990 ng

RT: 14.97 min Scan# 2191

Delta R.T. 0.00 min

Lab File: BF111416.D

Acq: 14 Dec 2018 14:01


Tgt Ion:252 Resp: 917132

Ion Ratio Lower Upper

252 100

253 21.2 18.2 27.2



125 11.5 10.4 15.6




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E









#89
Benzo(k)fluoranthene
Concen: 41.971 ng
RT: 15.00 min Scan# 2196
Delta R.T. 0.00 min
Lab File: BF111416.D
Acq: 14 Dec 2018 14:01

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tot Ion:252 Resp: 943331
Ion Ratio Lower Upper
252 100
253 22.4 18.1 27.1
125 12.5 9.8 14.8




Abundance Ion 252.00 (251.70 to 252.70): E
Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E






#90
Benzo(a)pyrene
Concen: 39.523 ng
RT: 15.33 min Scan# 2251
Delta R.T. 0.00 min
Lab File: BF111416.D
Acq: 14 Dec 2018 14:01

Tgt Ion:252 Resp: 838327
Ion Ratio Lower Upper
252 100
253 21.0 17.8 26.8
125 12.7 12.3 18.5



Abundance Ion 252.00 (251.70 to 252.70): E
Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E









#91
Dibenzo(a,h)anthracene
Concen: 36.671 ng
RT: 16.82 min Scan# 2504
Delta R.T. 0.00 min
Lab File: BF111416.D
Acq: 14 Dec 2018 14:01

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tot Ion:278 Resp: 613588
Ion Ratio Lower Upper
278 100
139 18.0 18.2 27.4#
279 23.2 19.0 28.6




Abundance Ion 278.00 (277.70 to 278.70): E
Ion 139.00 (138.70 to 139.70): E
Ion 279.00 (278.70 to 279.70): E



#92
Benzo(g,h,i)perylene
Concen: 36.237 ng
RT: 17.23 min Scan# 2574
Delta R.T. 0.00 min
Lab File: BF111416.D
Acq: 14 Dec 2018 14:01

Tgt Ion:276 Resp: 595151
Ion Ratio Lower Upper
276 100
277 23.3 19.2 28.8
138 28.0 24.9 37.3



Abundance Ion 276.00 (275.70 to 276.70): E
Ion 277.00 (276.70 to 277.70): E
Ion 138.00 (137.70 to 138.70): E

