

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF121518\
 Data File : BF111468.D
 Acq On : 15 Dec 2018 15:03
 Operator : JU/SJ
 Sample : J6316-01
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 BORING-SAMPLE-SIGNAL-TOWER-

Manual Integrations
 APPROVED

Sohil
 12/17/2018 3:11:58 PM

Quant Time: Dec 16 01:15:30 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF121418.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 14 13:26:17 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.79	152	174585	20.00	ng	0.00
21) Naphthalene-d8	8.07	136	698822	20.00	ng	0.00
39) Acenaphthene-d10	9.83	164	315407	20.00	ng	0.00
64) Phenanthrene-d10	11.31	188	456826	20.00	ng	0.00
76) Chrysene-d12	13.95	240	335915	20.00	ng	0.00
87) Perylene-d12	15.39	264	320703	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.42	112	1048971	99.98	ng	0.00
7) Phenol-d6	6.44	99	1366682	97.93	ng	0.00
23) Nitrobenzene-d5	7.35	82	833211	71.00	ng	0.00
42) 2,4,6-Tribromophenol	10.62	330	218595	77.37	ng	0.00
45) 2-Fluorobiphenyl	9.15	172	1490139	72.98	ng	0.00
79) Terphenyl-d14	12.90	244	948525	66.31	ng	0.00
Target Compounds						
10) Phenol	6.45	94	78325	5.038	ng	91
31) Naphthalene	8.10	128	205454	6.286	ng	95
37) 2-Methylnaphthalene	8.79	142	54716	2.474	ng	# 75
50) Dimethylphthalate	9.55	163	141710	6.239	ng	98
71) Phenanthrene	11.33	178	389615	16.552	ng	95
72) Anthracene	11.39	178	109239	4.538	ng	95
75) Fluoranthene	12.53	202	486998	21.148	ng	95
78) Pyrene	12.75	202	585098	24.705	ng	96
81) Benzo(a)anthracene	13.94	228	317185	17.366	ng	100
83) Chrysene	13.97	228	302004	15.696	ng	98
86) Indeno(1,2,3-cd)pyrene	16.80	276	139358m	10.035	ng	
88) Benzo(b)fluoranthene	14.98	252	409979m	21.927	ng	
89) Benzo(k)fluoranthene	15.00	252	97302m	5.446	ng	
90) Benzo(a)pyrene	15.33	252	334987	19.868	ng	98
91) Dibenzo(a,h)anthracene	16.82	278	37465	2.817	ng	95
92) Benzo(g,h,i)perylene	17.23	276	137821	10.557	ng	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF121518\
Data File : BF111468.D
Acq On : 15 Dec 2018 15:03
Operator : JU/SJ
Sample : J6316-01
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
BORING-SAMPLE-SIGNAL-TOWER-

Manual Integrations
APPROVED

Sohil
12/17/2018 3:11:58 PM

Quant Time: Dec 16 01:15:30 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF121418.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 14 13:26:17 2018
Response via : Initial Calibration

