

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF121522\
 Data File : BF131641.D
 Acq On : 15 Dec 2022 10:36
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 16 01:46:08 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF120822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 16 01:44:46 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	134	0.00
2	1,4-Dioxane	0.541	0.571	-5.5	144	0.00
3	Pyridine	1.502	1.613	-7.4	144	0.00
4	n-Nitrosodimethylamine	0.829	0.765	7.7	126	0.00
5 S	2-Fluorophenol	1.199	1.205	-0.5	135	0.00
6	Aniline	1.879	1.923	-2.3	138	0.00
7 S	Phenol-d6	1.571	1.569	0.1	137	0.00
8	2-Chlorophenol	1.270	1.264	0.5	137	0.00
9	Benzaldehyde	1.076	0.974	9.5	132	0.00
10 C	Phenol	1.752	1.822	-4.0	140	0.00
11	bis(2-Chloroethyl)ether	1.314	1.341	-2.1	139	0.00
12	1,3-Dichlorobenzene	1.473	1.460	0.9	136	0.00
13 C	1,4-Dichlorobenzene	1.486	1.470	1.1	136	0.00
14	1,2-Dichlorobenzene	1.419	1.369	3.5	133	0.00
15	Benzyl Alcohol	1.440	1.424	1.1	131	0.00
16	2,2'-oxybis(1-Chloropropane	1.697	1.657	2.4	133	0.00
17	2-Methylphenol	1.145	1.163	-1.6	136	0.00
18	Hexachloroethane	0.599	0.594	0.8	133	0.00
19 P	n-Nitroso-di-n-propylamine	1.178	1.112	5.6	126	0.00
20	3+4-Methylphenols	1.560	1.555	0.3	133	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	135	0.00
22	Acetophenone	0.627	0.601	4.1	128	0.00
23 S	Nitrobenzene-d5	0.529	0.519	1.9	131	0.00
24	Nitrobenzene	0.534	0.525	1.7	131	0.00
25	Isophorone	0.852	0.844	0.9	132	0.00
26 C	2-Nitrophenol	0.182	0.203	-11.5	145	0.00
27	2,4-Dimethylphenol	0.279	0.289	-3.6	137	0.00
28	bis(2-Chloroethoxy)methane	0.441	0.445	-0.9	135	0.00
29 C	2,4-Dichlorophenol	0.345	0.345	0.0	133	0.00
30	1,2,4-Trichlorobenzene	0.423	0.413	2.4	132	0.00
31	Naphthalene	1.126	1.103	2.0	130	0.00
32	Benzoic acid	0.202	0.203	-0.5	131	0.00
33	4-Chloroaniline	0.422	0.430	-1.9	137	0.00
34 C	Hexachlorobutadiene	0.298	0.289	3.0	129	0.00
35	Caprolactam	0.089	0.095	-6.7	144	0.00
36 C	4-Chloro-3-methylphenol	0.387	0.387	0.0	132	0.00
37	2-Methylnaphthalene	0.770	0.752	2.3	130	0.00
38	1-Methylnaphthalene	0.740	0.717	3.1	130	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	130	0.00
40	1,2,4,5-Tetrachlorobenzene	0.740	0.752	-1.6	128	0.00
41 P	Hexachlorocyclopentadiene	0.376	0.326	13.3	109	0.00
42 S	2,4,6-Tribromophenol	0.222	0.223	-0.5	128	0.00
43 C	2,4,6-Trichlorophenol	0.458	0.482	-5.2	133	0.00
44	2,4,5-Trichlorophenol	0.492	0.517	-5.1	131	0.00
45 S	2-Fluorobiphenyl	1.586	1.555	2.0	125	0.00
46	1,1'-Biphenyl	1.581	1.613	-2.0	129	0.00
47	2-Chloronaphthalene	1.288	1.322	-2.6	130	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF121522\
 Data File : BF131641.D
 Acq On : 15 Dec 2022 10:36
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 16 01:46:08 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF120822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 16 01:44:46 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.456	0.473	-3.7	130	0.00
49	Acenaphthylene	1.892	1.964	-3.8	131	0.00
50	Dimethylphthalate	1.525	1.560	-2.3	129	0.00
51	2,6-Dinitrotoluene	0.315	0.342	-8.6	136	0.00
52 C	Acenaphthene	1.267	1.311	-3.5	130	0.00
53	3-Nitroaniline	0.302	0.330	-9.3	139	0.00
54 P	2,4-Dinitrophenol	0.158	0.173	-9.5	134	0.00
55	Dibenzofuran	1.791	1.837	-2.6	131	0.00
56 P	4-Nitrophenol	0.210	0.221	-5.2	132	0.00
57	2,4-Dinitrotoluene	0.418	0.451	-7.9	135	0.00
58	Fluorene	1.456	1.472	-1.1	129	0.00
59	2,3,4,6-Tetrachlorophenol	0.417	0.420	-0.7	127	0.00
60	Diethylphthalate	1.463	1.478	-1.0	129	0.00
61	4-Chlorophenyl-phenylether	0.804	0.780	3.0	123	0.00
62	4-Nitroaniline	0.299	0.315	-5.4	134	0.00
63	Azobenzene	1.669	1.635	2.0	127	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	132	0.00
65	4,6-Dinitro-2-methylphenol	0.127	0.141	-11.0	142	0.00
66 c	n-Nitrosodiphenylamine	0.647	0.651	-0.6	130	0.00
67	4-Bromophenyl-phenylether	0.257	0.252	1.9	125	0.00
68	Hexachlorobenzene	0.278	0.271	2.5	125	0.00
69	Atrazine	0.205	0.211	-2.9	127	0.00
70 C	Pentachlorophenol	0.143	0.130	9.1	115	0.00
71	Phenanthrene	1.115	1.128	-1.2	133	0.00
72	Anthracene	1.085	1.087	-0.2	129	0.00
73	Carbazole	0.885	0.901	-1.8	132	0.00
74	Di-n-butylphthalate	1.086	1.117	-2.9	134	0.00
75 C	Fluoranthene	1.148	1.157	-0.8	132	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	121	0.00
77	Benzydine	0.536	0.421	21.5	78	0.00
78	Pyrene	1.972	2.304	-16.8	128	0.00
79 S	Terphenyl-d14	1.458	1.599	-9.7	123	0.00
80	Butylbenzylphthalate	0.571	0.680	-19.1	129	0.00
81	Benzo(a)anthracene	1.444	1.555	-7.7	122	0.00
82	3,3'-Dichlorobenzidine	0.411	0.423	-2.9	113	0.00
83	Chrysene	1.373	1.477	-7.6	122	0.00
84	Bis(2-ethylhexyl)phthalate	0.691	0.703	-1.7	107	0.00
85 c	Di-n-octyl phthalate	1.164	0.958	17.7	93	0.00
86 I	Perylene-d12	1.000	1.000	0.0	118	0.00
87	Indeno(1,2,3-cd)pyrene	1.632	1.741	-6.7	117	0.00
88	Benzo(b)fluoranthene	1.318	1.368	-3.8	124	0.00
89	Benzo(k)fluoranthene	1.376	1.273	7.5	113	0.00
90 C	Benzo(a)pyrene	1.125	1.139	-1.2	120	0.00
91	Dibenzo(a,h)anthracene	1.343	1.423	-6.0	116	0.00
92	Benzo(g,h,i)perylene	1.350	1.429	-5.9	116	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF121522\
Data File : BF131641.D
Acq On : 15 Dec 2022 10:36
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Dec 16 01:46:08 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF120822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 16 01:44:46 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0