

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF121622\
 Data File : BF131664.D
 Acq On : 16 Dec 2022 13:45
 Operator : CG\JU
 Sample : N6038-06RE
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 RB-31-(6.5-8.5)RE

Quant Time: Dec 16 16:58:07 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF120822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 16 01:44:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.881	152	39471	20.000	ng	0.00
21) Naphthalene-d8	8.169	136	142609	20.000	ng	# 0.00
39) Acenaphthene-d10	9.933	164	82963	20.000	ng	0.00
64) Phenanthrene-d10	11.427	188	160443	20.000	ng	0.00
76) Chrysene-d12	14.074	240	103618	20.000	ng	# 0.00
86) Perylene-d12	15.580	264	91317	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.492	112	401444	169.675	ng	0.00
7) Phenol-d6	6.510	99	512662	165.352	ng	0.00
23) Nitrobenzene-d5	7.445	82	355602	94.273	ng	0.00
42) 2,4,6-Tribromophenol	10.727	330	139304	151.131	ng	0.00
45) 2-Fluorobiphenyl	9.251	172	591201	89.851	ng	0.00
79) Terphenyl-d14	13.021	244	661442	87.543	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF121622\
Data File : BF131664.D
Acq On : 16 Dec 2022 13:45
Operator : CG\JU
Sample : N6038-06RE
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
RB-31-(6.5-8.5)RE

Quant Time: Dec 16 16:58:07 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF120822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 16 01:44:46 2022
Response via : Initial Calibration

