

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF121718\
 Data File : BF111537.D
 Acq On : 17 Dec 2018 14:11
 Operator : JU/SJ
 Sample : J6386-11
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 P001-SS006-1218-01

Quant Time: Dec 18 04:48:54 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF121418.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 14 13:26:17 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.79	152	155144	20.00	ng	0.00
21) Naphthalene-d8	8.07	136	582813	20.00	ng	-0.01
39) Acenaphthene-d10	9.82	164	230477	20.00	ng	-0.01
64) Phenanthrene-d10	11.31	188	305292	20.00	ng	0.00
76) Chrysene-d12	13.94	240	293327	20.00	ng	0.00
87) Perylene-d12	15.39	264	240711	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.42	112	853563	91.55	ng	0.00
7) Phenol-d6	6.44	99	1074194	86.62	ng	0.00
23) Nitrobenzene-d5	7.35	82	645000	65.90	ng	-0.01
42) 2,4,6-Tribromophenol	10.61	330	158922	76.98	ng	-0.01
45) 2-Fluorobiphenyl	9.15	172	1037117	69.51	ng	-0.01
79) Terphenyl-d14	12.89	244	643676	51.53	ng	0.00
Target Compounds						
9) Benzaldehyde	6.33	77	19930	2.650	ng	91
10) Phenol	6.45	94	65037	4.708	ng	92
50) Dimethylphthalate	9.54	163	85635	5.159	ng	96
71) Phenanthrene	11.33	178	33986	2.160	ng	92
78) Pyrene	12.75	202	45077	2.180	ng	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF121718\
Data File : BF111537.D
Acq On : 17 Dec 2018 14:11
Operator : JU/SJ
Sample : J6386-11
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
P001-SS006-1218-01

Quant Time: Dec 18 04:48:54 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF121418.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 14 13:26:17 2018
Response via : Initial Calibration

