

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF121718\
 Data File : BF111543.D
 Acq On : 17 Dec 2018 17:13
 Operator : JU/SJ
 Sample : J6440-05
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 COMP-3

Quant Time: Dec 18 05:51:08 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF121418.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 14 13:26:17 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.79	152	172096	20.00	ng	0.00
21) Naphthalene-d8	8.07	136	613767	20.00	ng	0.00
39) Acenaphthene-d10	9.83	164	341350	20.00	ng	0.00
64) Phenanthrene-d10	11.32	188	349136	20.00	ng	0.00
76) Chrysene-d12	13.95	240	366610	20.00	ng	0.00
87) Perylene-d12	15.40	264	391661	20.00	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.43	112	577913	55.88	ng	0.01
7) Phenol-d6	6.46	99	711107	51.69	ng	0.02
23) Nitrobenzene-d5	7.36	82	1017789	98.75	ng	0.00
42) 2,4,6-Tribromophenol	10.64	330	287477	94.02	ng	0.02
45) 2-Fluorobiphenyl	9.16	172	2018510	91.34	ng	0.00
79) Terphenyl-d14	12.90	244	1127169	72.20	ng	0.00
Target Compounds						
9) Benzaldehyde	6.33	77	148151	17.755	ng	97
10) Phenol	6.47	94	151414	9.881	ng	97
17) 2-Methylphenol	7.07	107	113967	11.460	ng	99
20) 3+4-Methylphenols	7.23	107	172262	13.809	ng	# 62
27) 2,4-Dimethylphenol	7.75	122	90746	11.480	ng	# 93
31) Naphthalene	8.10	128	1173458	40.877	ng	97
37) 2-Methylnaphthalene	8.79	142	177251	9.125	ng	96
38) 1-Methylnaphthalene	8.89	142	112208	5.979	ng	# 98
50) Dimethylphthalate	9.56	163	387953	15.781	ng	98
74) Di-n-butylphthalate	11.88	149	53538	2.527	ng	# 83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF121718\
Data File : BF111543.D
Acq On : 17 Dec 2018 17:13
Operator : JU/SJ
Sample : J6440-05
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
COMP-3

Quant Time: Dec 18 05:51:08 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF121418.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 14 13:26:17 2018
Response via : Initial Calibration

