

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF121718\
 Data File : BF111544.D
 Acq On : 17 Dec 2018 17:41
 Operator : JU/SJ
 Sample : J6440-08
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 COMP-4

Quant Time: Dec 18 06:00:46 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF121418.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 14 13:26:17 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.79	152	187888	20.00	ng	0.00
21) Naphthalene-d8	8.07	136	756044	20.00	ng	-0.01
39) Acenaphthene-d10	9.82	164	358552	20.00	ng	-0.01
64) Phenanthrene-d10	11.31	188	616286	20.00	ng	0.00
76) Chrysene-d12	13.95	240	255143	20.00	ng	0.00
87) Perylene-d12	15.38	264	233805	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.43	112	1009178	89.38	ng	0.01
7) Phenol-d6	6.44	99	1153166	76.78	ng	0.00
23) Nitrobenzene-d5	7.35	82	1023463	80.61	ng	-0.01
42) 2,4,6-Tribromophenol	10.62	330	368779	114.82	ng	0.00
45) 2-Fluorobiphenyl	9.15	172	1847941	79.61	ng	-0.01
79) Terphenyl-d14	12.90	244	1276271	117.46	ng	0.00
Target Compounds						
20) 3+4-Methylphenols	7.20	107	60239	4.423	ng	Qvalue # 57
50) Dimethylphthalate	9.54	163	223835	8.668	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF121718\
Data File : BF111544.D
Acq On : 17 Dec 2018 17:41
Operator : JU/SJ
Sample : J6440-08
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
COMP-4

Quant Time: Dec 18 06:00:46 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF121418.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 14 13:26:17 2018
Response via : Initial Calibration

