

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF121718\
 Data File : BF111556.D
 Acq On : 17 Dec 2018 23:12
 Operator : JU/SJ
 Sample : J6403-29
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 TP-4-B

Manual Integrations
 APPROVED

Sohil
 12/18/2018 12:49:38 PM

Quant Time: Dec 18 06:57:10 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF121418.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 14 13:26:17 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.79	152	118778	20.00	ng	0.00
21) Naphthalene-d8	8.07	136	504754	20.00	ng	-0.01
39) Acenaphthene-d10	9.82	164	353451	20.00	ng	-0.01
64) Phenanthrene-d10	11.31	188	562495	20.00	ng	0.00
76) Chrysene-d12	13.94	240	358648	20.00	ng	0.00
87) Perylene-d12	15.39	264	335570	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.42	112	718690	100.69	ng	0.00
7) Phenol-d6	6.44	99	914364	96.30	ng	0.00
23) Nitrobenzene-d5	7.35	82	564473	66.59	ng	-0.01
42) 2,4,6-Tribromophenol	10.62	330	278874	88.08	ng	0.00
45) 2-Fluorobiphenyl	9.15	172	1439892	62.93	ng	-0.01
79) Terphenyl-d14	12.89	244	1002928	65.67	ng	0.00
Target Compounds						
10) Phenol	6.45	94	56322	5.325	ng	91
31) Naphthalene	8.09	128	56079	2.375	ng	# 93
50) Dimethylphthalate	9.54	163	200313	7.869	ng	97
71) Phenanthrene	11.33	178	447987	15.456	ng	94
72) Anthracene	11.38	178	103208	3.482	ng	93
75) Fluoranthene	12.52	202	653244	23.038	ng	98
78) Pyrene	12.74	202	573405	22.677	ng	96
81) Benzo(a)anthracene	13.93	228	288860	14.813	ng	100
83) Chrysene	13.97	228	277548	13.510	ng	99
88) Benzo(b)fluoranthene	14.97	252	343182m	17.541	ng	
89) Benzo(k)fluoranthene	15.00	252	131000m	7.008	ng	
90) Benzo(a)pyrene	15.32	252	246013	13.945	ng	96
91) Dibenzo(a,h)anthracene	16.80	278	32798	2.357	ng	# 91
92) Benzo(a,h,i)perylene	17.23	276	109154	7.990	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF121718\
Data File : BF111556.D
Acq On : 17 Dec 2018 23:12
Operator : JU/SJ
Sample : J6403-29
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
TP-4-B

Manual Integrations
APPROVED

Sohil
12/18/2018 12:49:38 PM

Quant Time: Dec 18 06:57:10 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF121418.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 14 13:26:17 2018
Response via : Initial Calibration

